SGS AXYS

2045 Mills Road West
Sidney, BC, Canada V8L5X2

TEL: (250) 655-5800
TOLL-FREE: 1-888-373-0881

SGS AXYS Client No.: 4114

Client Address: Florida Dept of Environmental Protection
2600 Blair Stone Road
Tallahassee, FL, USA, 32399-2400

The SGS AXYS contact for these data is Megan Preik.

BATCH SUMMARY

Batch ID: WG59329	***Date:*** 17-May-2017
Analysis Type: (AP5) Pharmaceutical & Personal Care Products MLA-075	***Matrix Type:*** Aqueous

BATCH MAKEUP

Contract: 4114

Samples:

L27146-1	QA/ QC BLANK SW FIELD
L27146-2	SP-WITHLAC
L27146-3	SP-OCHLCK
L27146-4	SP-APALACH
L27146-5	SP-ALAFIA

Blank: WG59329-101

Reference or Spike: WG59329-102

Duplicate:

Comments:

1. Data are considered final.
2. The analyte DEET was detected in the Lab Blank (AXYS ID: WG59329-101). Data are not blank corrected. Blank data should be taken into consideration when evaluating sample data.
3. Blank data should be evaluated against specifications using the same blank sample size as the size of the client samples.
4. Percent recovery of several labeled compounds in the Lab Blank and client samples were outside the method nominal limit and flagged with a 'V'. As the isotope dilution method of quantification produces data that are recovery corrected, the slight variances from the method acceptance criteria are deemed not to affect the quantification of these analytes. Percent labeled compound recoveries are used as a general method performance indicator only.

Copyright SGS AXYS Analytical Services Ltd
February 2017

AXYS METHOD MLA-075 Rev 07

Form 3A
PHARMACEUTICAL INITIAL CALIBRATION PERCENT RECOVERIES

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 27-Apr-2017

Instrument ID: LC MS/MS

LC Column ID: C18MS

CS0 Data Filename: QE7Q_077 S: 4
CS1 Data Filename: QE7Q_077 S: 5
CS2 Data Filename: QE7Q_077 S: 6
CS3 Data Filename: QE7Q_077 S: 7
CS4 Data Filename: QE7Q_077 S: 8
CS5 Data Filename: QE7Q_077 S: 9
CS6 Data Filename: QE7Q_077 S: 10
CS7 Data Filename: N/A
CS8 Data Filename: N/A

		PERCENT RECOVERY (%)								
COMPOUND	**LAB FLAG [1]**	**CS0**	**CS1**	**CS2**	**CS3**	**CS4**	**CS5**	**CS6**	**CS7**	**CS8**
Alprazolam		92.6	97.4	104	101	106	100	99.4		
Amitriptyline		99.3	105	104	102	92.4	94.0	103		
Amlodipine		102	90.2	103	110	98.0	95.1	102		
Benzoylecgonine		99.1	90.1	104	94.0	114	100	98.9		
Benztropine		102	86.5	88.6	118	108	97.0			
Betamethasone		71.0	101	129	101	98.0				
Cocaine		112	96.9	90.8	101	97.9	102	99.6		
DEET		82.4	105	96.2	104	106	111	95.9		
Desmethyldiltiazem		85.0	108	97.2	124	94.3	86.9	104		
Diazepam		93.2	95.2	107	99.9	105	101	99.3		
Fluocinonide		71.9	104	120	105	101	99.1			
Fluticasone propionate		109	102	96.3	95.4	96.1	101			
Hydrocortisone		116	121	95.4	86.3	94.7	79.4	108		
10-hydroxy-amitriptyline		138	108	93.1	80.4	86.6	88.4	105		
Meprobamate		120	96.5	105	92.5	94.7	84.8	106		
Methylprednisolone		93.3	90.4	105	104	107	101	99.0		
Metoprolol		103	99.9	95.7	104	101	96.3	101		
Norfluoxetine		111	97.3	94.9	102	93.2	101	100		
Norverapamil		113	97.5	91.9	93.8	104	99.6			
Paroxetine		98.9	95.4	98.4	106	104	96.4	101		
Prednisolone		84.3	91.1	119	110	96.2				
Prednisone		74.9	104	127	94.9					
Promethazine		72.7	104	93.6	123	105	103	97.9		
Propoxyphene		104	102	93.4	97.4	104	99.2	100		
Propranolol		74.9	107	103	105	114	96.1			
Sertraline		123	87.5	91.5	97.1	100	100			
Simvastatin		76.8	108	108	111	96.2				
Theophylline		129	101	92.4	89.9	88.0	96.4	102		
Trenbolone		67.1	100	123	105	107	97.7			
Trenbolone acetate		120	98.1	101	88.4	91.8	99.3	101		
Valsartan		99.6	107	94.3	96.0	109	92.0	102		
Verapamil		90.4	110	96.6	104	97.8	103	99.2		

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

AXYS METHOD MLA-075 Rev 07

Form 3B
PHARMACEUTICAL INITIAL CALIBRATION PERCENT RECOVERIES

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 27-Apr-2017

Instrument ID: LC MS/MS

LC Column ID: C18MS

CS0 Data Filename: QE7Q_077 S: 4
CS1 Data Filename: QE7Q_077 S: 5
CS2 Data Filename: QE7Q_077 S: 6
CS3 Data Filename: QE7Q_077 S: 7
CS4 Data Filename: QE7Q_077 S: 8
CS5 Data Filename: QE7Q_077 S: 9
CS6 Data Filename: QE7Q_077 S: 10
CS7 Data Filename: N/A
CS8 Data Filename: N/A

		PERCENT RECOVERIES (%)								
LABELED COMPOUND	**LAB FLAG [1]**	**CS0**	**CS1**	**CS2**	**CS3**	**CS4**	**CS5**	**CS6**	**CS7**	**CS8**
d5-Alprazolam		109	109	94.4	104	96.8	94.8	92.3		
d6-Amitriptyline		110	101	97.5	100	97.1	96.3	98.4		
d8-Benzoylecgonine		103	102	101	107	84.6	96.5	106		
d3-Benztropine		96.5	103	91.1	90.4	87.0	103	129		
d3-Cocaine		98.1	103	106	100	89.8	98.3	105		
d7-DEET		108	109	102	105	98.1	93.7	83.9		
d5-Diazepam		97.6	101	103	103	93.2	99.8	103		
d4-Hydrocortisone		111	101	88.8	115	103	97.0	83.9		
d3-Methylprednisolone		109	117	107	106	89.4	93.2	77.0		
d7-Metoprolol		108	113	109	103	92.4	94.0	80.7		
d5-Norfluoxetine		101	107	95.4	94.2	104	101	97.1		
d6-Paroxetine		115	107	98.7	103	97.4	94.1	84.6		
d4-Promethazine		110	103	102	85.7	94.3	106	97.8		
d5-Propoxyphene		112	104	103	97.1	90.0	97.0	96.8		
d7-Propranolol		110	107	105	105	90.9	94.1	88.1		
13C1-15N2-Theophylline		114	121	119	102	90.0	75.2	80.0		

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form3B.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: GENERIC-SPECS_PPC_LC_27-Apr-2017_QE7Q__Form3B_GS70011.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 3C
PHARMACEUTICAL INITIAL CALIBRATION RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 27-Apr-2017

Instrument ID: LC MS/MS

LC Column ID: C18MS

CS0 Data Filename: QE7Q_077 S: 4
CS1 Data Filename: QE7Q_077 S: 5
CS2 Data Filename: QE7Q_077 S: 6
CS3 Data Filename: QE7Q_077 S: 7
CS4 Data Filename: QE7Q_077 S: 8
CS5 Data Filename: QE7Q_077 S: 9
CS6 Data Filename: QE7Q_077 S: 10
CS7 Data Filename: N/A
CS8 Data Filename: N/A

		RETENTION TIMES									
COMPOUND	**LAB FLAG [1]**	**CS0**	**CS1**	**CS2**	**CS3**	**CS4**	**CS5**	**CS6**	**CS7**	**CS8**	**MEAN RT**
Alprazolam		23:36	23:42	23:42	23:36	23:42	23:42	23:42			23:40
Amitriptyline		22:30	22:30	22:30	22:30	22:30	22:30	22:30			22:30
Amlodipine		23:54	23:54	23:54	23:48	23:54	23:54	23:54			23:53
Benzoylecgonine		5:40	5:47	5:47	5:30	5:23	5:40	5:33			5:37
Benztropine		22:55	23:03	22:55	22:55	23:03	23:03				22:59
Betamethasone		21:49	21:49	21:49	21:49	21:49					21:49
Cocaine		9:07	9:07	9:04	9:00	9:07	9:07	9:07			9:06
DEET		21:01	21:01	21:01	21:01	21:01	21:01	21:01			21:01
Desmethyldiltiazem		19:00	19:00	19:00	19:00	19:00	19:00	19:00			19:00
Diazepam		29:37	29:41	29:39	29:37	29:41	29:41	29:39			29:39
Fluocinonide		35:21	35:23	35:23	35:21	35:23	35:23				35:22
Fluticasone propionate		37:38	37:40	37:40	37:38	37:40	37:40				37:39
Hydrocortisone		17:55	17:55	17:55	17:53	17:55	17:57	17:53			17:55
10-hydroxy-amitriptyline		12:08	12:08	12:08	12:04	12:08	12:08	12:08			12:07
Meprobamate		11:27	11:23	11:27	11:23	11:23	11:27	11:27			11:25
Methylprednisolone		21:51	21:51	21:51	21:51	21:51	21:51	21:51			21:51
Metoprolol		8:26	8:26	8:20	8:16	8:26	8:26	8:23			8:23
Norfluoxetine		24:49	24:49	24:49	24:49	24:58	24:58	24:49			24:52
Norverapamil		21:08	21:08	21:08	21:08	21:08	21:16				21:09
Paroxetine		20:45	20:45	20:45	20:45	20:45	20:45	20:45			20:45
Prednisolone		17:45	17:50	17:45	17:45	17:45					17:46
Prednisone		17:05	16:59	17:05	16:59						17:02
Promethazine		18:52	18:52	18:52	18:52	18:52	18:52	18:52			18:52
Propoxyphene		21:57	21:57	21:57	21:57	21:57	21:57	21:57			21:57
Propranolol		14:47	14:47	14:47	14:44	14:47	14:47				14:47
Sertraline		26:24	26:24	26:24	26:16	26:24	26:24				26:23
Simvastatin		39:39	39:39	39:39	39:39	39:39					39:39
Theophylline		2:33	2:29	2:33	2:33	2:26	2:36	2:33			2:32
Trenbolone		23:03	23:03	23:03	23:03	23:03	23:03				23:03
Trenbolone acetate		37:26	37:26	37:24	37:24	37:24	37:24	37:24			37:25
Valsartan		32:35	32:35	32:30	32:28	32:32	32:35	32:32			32:32
Verapamil		21:32	21:32	21:32	21:32	21:41	21:41	21:41			21:36

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.
Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form3C.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: GENERIC-SPECS_PPC_LC_27-Apr-2017_QE7Q__Form3C_GS70011.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 3D
PHARMACEUTICAL INITIAL CALIBRATION RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 27-Apr-2017

Instrument ID: LC MS/MS

LC Column ID: C18MS

CS0 Data Filename: QE7Q_077 S: 4
CS1 Data Filename: QE7Q_077 S: 5
CS2 Data Filename: QE7Q_077 S: 6
CS3 Data Filename: QE7Q_077 S: 7
CS4 Data Filename: QE7Q_077 S: 8
CS5 Data Filename: QE7Q_077 S: 9
CS6 Data Filename: QE7Q_077 S: 10
CS7 Data Filename: N/A
CS8 Data Filename: N/A

		RETENTION TIMES								
LABELED COMPOUND	**LAB FLAG[1]**	**CS0**	**CS1**	**CS2**	**CS3**	**CS4**	**CS5**	**CS6**	**CS7**	**CS8**	**MEAN RT**
d5-Alprazolam		23:28	23:28	23:28	23:20	23:28	23:28	23:20			23:26
d6-Amitriptyline		22:30	22:30	22:30	22:22	22:30	22:30	22:30			22:29
d8-Benzoylecgonine		5:23	5:40	5:23	5:23	5:13	5:30	5:23			5:25
d3-Benztropine		22:55	22:55	22:55	22:55	22:55	23:03	22:55			22:56
d3-Cocaine		9:07	9:07	9:04	9:00	9:07	9:11	9:07			9:06
d7-DEET		20:45	20:45	20:45	20:45	20:45	20:45	20:45			20:45
d5-Diazepam		29:23	29:23	29:23	29:21	29:23	29:25	29:23			29:23
d4-Hydrocortisone		17:53	17:53	17:53	17:50	17:53	17:53	17:50			17:52
d3-Methylprednisolone		21:51	21:51	21:51	21:43	21:51	21:51	21:51			21:50
d7-Metoprolol		8:20	8:20	8:13	8:09	8:20	8:23	8:16			8:17
d5-Norfluoxetine		24:49	24:49	24:49	24:49	24:49	24:49	24:49			24:49
d6-Paroxetine		20:37	20:45	20:37	20:37	20:45	20:45	20:37			20:40
d4-Promethazine		18:44	18:44	18:44	18:44	18:44	18:44	18:44			18:44
d5-Propoxyphene		21:49	21:49	21:49	21:49	21:49	21:49	21:49			21:49
d7-Propranolol		14:34	14:34	14:34	14:30	14:34	14:34	14:34			14:33
13C1-15N2-Theophylline		2:29	2:26	2:33	2:29	2:29	2:33	2:33			2:30

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form3D.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: GENERIC-SPECS_PPC_LC_27-Apr-2017_QE7Q__Form3D_GS70011.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 4A
PHARMACEUTICAL CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date:	27-Apr-2017	**VER Data Filename:**	QE7Q_077 S: 13
Instrument ID:	LC MS/MS	**Analysis Date:**	27-Apr-2017
LC Column ID:	C18MS	**Analysis Time:**	23:05:11

COMPOUND	LAB FLAG [1]	RETENTION TIME	EXPECTED CONC. (ng)	CONC. FOUND (ng)	RECOVERY (%)
Alprazolam		23:42	15.0	14.4	95.7
Amitriptyline		22:30	15.0	14.1	94.1
Amlodipine		23:54	75.0	85.6	114
Benzoylecgonine		5:33	15.0	14.9	99.6
Benztropine		23:03	15.2	16.6	109
Betamethasone		21:49	75.0	77.8	104
Cocaine		9:07	7.50	6.99	93.2
DEET		21:01	30.2	28.2	93.5
Desmethyldiltiazem		19:00	7.50	8.03	107
Diazepam		29:39	15.0	15.2	101
Fluocinonide		35:21	300	299	99.7
Fluticasone propionate		37:38	100	88.0	87.9
Hydrocortisone		17:57	3000	3420	114
10-hydroxy-amitriptyline		12:08	7.50	7.39	98.6
Meprobamate		11:27	200	185	92.4
Methylprednisolone		21:51	200	215	108
Metoprolol		8:20	75.0	74.0	98.7
Norfluoxetine		24:58	75.0	74.3	99.1
Norverapamil		21:16	7.50	7.93	106
Paroxetine		20:45	200	224	112
Prednisolone		17:50	300	344	115
Prednisone		17:05	1000	931	93.1
Promethazine		18:52	20.0	20.5	103
Propoxyphene		21:57	15.0	14.5	96.8
Propranolol		14:47	100	113	113
Sertraline		26:24	20.0	19.2	96.5
Simvastatin		39:39	1000	940	94.0
Theophylline		2:33	3000	2820	94.0
Trenbolone		23:03	200	212	106
Trenbolone acetate		37:24	15.0	13.1	87.1
Valsartan		32:32	200	192	95.9
Verapamil		21:41	7.50	6.82	90.9

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form4A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5; Report Filename: GENERIC-SPECS_PPC_LC_QE7Q_077S13__Form4A_SJ2204906.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 4B
PHARMACEUTICAL CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date:	27-Apr-2017	**VER Data Filename:**	QE7Q_077 S: 13
Instrument ID:	LC MS/MS	**Analysis Date:**	27-Apr-2017
LC Column ID:	C18MS	**Analysis Time:**	23:05:11

LABELED COMPOUND	LAB FLAG [1]	RETENTION TIME	EXPECTED CONC. (ng)	CONC. FOUND (ng)	RECOVERY (%)
d5-Alprazolam		23:28	40.0	43.5	109
d6-Amitriptyline		22:30	40.0	40.7	102
d8-Benzoylecgonine		5:33	40.0	40.9	102
d3-Benztropine		23:03	20.0	18.7	93.7
d3-Cocaine		9:07	40.0	39.7	99.1
d7-DEET		20:45	40.0	41.4	103
d5-Diazepam		29:25	40.0	39.7	99.2
d4-Hydrocortisone		17:55	8000	6840	85.4
d3-Methylprednisolone		21:51	2000	2150	108
d7-Metoprolol		8:20	400	440	110
d5-Norfluoxetine		24:49	200	203	101
d6-Paroxetine		20:45	100	111	111
d4-Promethazine		18:44	90.0	93.8	104
d5-Propoxyphene		21:49	60.0	63.8	106
d7-Propranolol		14:34	400	382	95.4
13C1-15N2-Theophylline		2:33	2000	2040	102

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form4B.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5; Report Filename: GENERIC-SPECS_PPC_LC_QE7Q_077S13__Form4B_SJ2204906.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
QA/ QC BLANK SW FIELD
Sample Collection:
17-Apr-2017 10:15

Form 1A

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-1
Matrix:	AQUEOUS	**Sample Size:**	0.978 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 04:30:29	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 20**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng/L		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Alprazolam	U		0.307 (L)	
Amitriptyline	U		0.307 (L)	
Amlodipine	U		1.53 (L)	
Benzoylecgonine	U		0.307 (L)	
Benztropine	U		0.511 (L)	
Betamethasone	U		1.53 (L)	
Cocaine	U		0.153 (L)	
DEET		1.17	0.818 (L)	21:01
Desmethyldiltiazem	U		0.153 (L)	
Diazepam	U		0.307 (L)	
Fluocinonide	U		6.14 (L)	
Fluticasone propionate	U		2.05 (L)	
Hydrocortisone	U		61.4 (L)	
10-hydroxy-amitriptyline	U		0.153 (L)	
Meprobamate	U		4.09 (L)	
Methylprednisolone	U		4.09 (L)	
Metoprolol	U		1.53 (L)	
Norfluoxetine	U		1.53 (L)	
Norverapamil	U		0.153 (L)	
Paroxetine	U		4.09 (L)	
Prednisolone	U		6.14 (L)	
Prednisone	U		20.5 (L)	
Promethazine	U		0.409 (L)	
Propoxyphene	U		0.307 (L)	
Propranolol	U		2.05 (L)	
Sertraline	U		0.409 (L)	
Simvastatin	U		20.5 (L)	
Theophylline [3]	U		61.4 (L)	
Trenbolone	U		4.09 (L)	

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Trenbolone acetate	U		0.307 (L)	
Valsartan	U		4.09 (L)	
Verapamil	U		0.153 (L)	

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form1A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-1_Form1A_QE7Q_077S20_SJ2204915.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
QA/ QC BLANK SW FIELD
Sample Collection:
17-Apr-2017 10:15

Form 2

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-1
Matrix:	AQUEOUS	**Sample Size:**	0.978 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 04:30:29	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 20**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng absolute		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC.	CONC. FOUND	R(%) [2]	RETENTION TIME
d5-Alprazolam		40.0	39.9	99.8	23:28
d6-Amitriptyline		40.0	46.4	116	22:30
d8-Benzoylecgonine		40.0	46.4	116	5:23
d3-Benztropine		20.0	20.8	104	23:03
d3-Cocaine		40.0	43.8	110	9:07
d7-DEET		40.0	32.5	81.3	20:45
d5-Diazepam		40.0	43.7	109	29:25
d4-Hydrocortisone		8000	8890	111	17:53
d3-Methylprednisolone		2000	2140	107	21:51
d7-Metoprolol		400	498	125	8:20
d5-Norfluoxetine		200	195	97.3	24:58
d6-Paroxetine		100	114	114	20:45
d4-Promethazine		100	90.7	90.7	18:44
d5-Propoxyphene		60.0	75.2	125	21:57
d7-Propranolol		400	465	116	14:37
13C1-15N2-Theophylline		2000	2120	106	2:33

(1) Where applicable, custom lab flags have been used on this report.
(2) R(%) = percent recovery.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form2.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-1_Form2_QE7Q_077S20_SJ2204915.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
SP-WITHLAC
Sample Collection:
17-Apr-2017 15:10

Form 1A

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-2
Matrix:	AQUEOUS	**Sample Size:**	0.974 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 05:16:55	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 21**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng/L		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Alprazolam	U		0.308 (L)	
Amitriptyline		1.26	0.421 (S)	22:30
Amlodipine	U		1.54 (L)	
Benzoylecgonine	U		0.308 (L)	
Benztropine	U		0.514 (L)	
Betamethasone	U		1.54 (L)	
Cocaine	U		0.154 (L)	
DEET		6.58	0.822 (L)	21:10
Desmethyldiltiazem	U		0.154 (L)	
Diazepam	U		0.308 (L)	
Fluocinonide	U		6.16 (L)	
Fluticasone propionate	U		2.05 (L)	
Hydrocortisone	U		61.6 (L)	
10-hydroxy-amitriptyline	U		0.188 (S)	
Meprobamate		6.97	4.11 (L)	11:27
Methylprednisolone	U		6.49 (S)	
Metoprolol	U		14.3 (S)	
Norfluoxetine	U		1.54 (L)	
Norverapamil	U		0.154 (L)	
Paroxetine	U		4.11 (L)	
Prednisolone	U		6.16 (L)	
Prednisone	U		34.1 (S)	
Promethazine	U		0.411 (L)	
Propoxyphene	U		0.308 (L)	
Propranolol	U		2.05 (L)	
Sertraline	U		0.411 (L)	
Simvastatin	U		20.5 (L)	
Theophylline [3]	U		61.6 (L)	
Trenbolone	U		4.11 (L)	

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Trenbolone acetate	U		0.308 (L)	
Valsartan		4.42	4.11 (L)	32:28
Verapamil	U		0.154 (L)	

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form1A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-2_Form1A_QE7Q_077S21_SJ2204916.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
SP-WITHLAC
Sample Collection:
17-Apr-2017 15:10

Form 2

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-2
Matrix:	AQUEOUS	**Sample Size:**	0.974 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 05:16:55	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 21**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng absolute		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC.	CONC. FOUND	R(%) [2]	RETENTION TIME
d5-Alprazolam		40.0	35.9	89.7	23:28
d6-Amitriptyline		40.0	49.3	123	22:38
d8-Benzoylecgonine		40.0	39.3	98.3	5:26
d3-Benztropine		20.0	6.22	31.1	23:11
d3-Cocaine		40.0	50.7	127	9:17
d7-DEET		40.0	49.5	124	20:53
d5-Diazepam		40.0	46.2	116	29:27
d4-Hydrocortisone		8000	9010	113	17:53
d3-Methylprednisolone		2000	2170	108	21:51
d7-Metoprolol		400	473	118	8:26
d5-Norfluoxetine		200	246	123	24:58
d6-Paroxetine		100	96.2	96.2	20:53
d4-Promethazine		100	122	122	18:52
d5-Propoxyphene	V	60.0	89.8	150	21:57
d7-Propranolol		400	406	101	14:44
13C1-15N2-Theophylline		2000	1540	77.2	2:33

(1) Where applicable, custom lab flags have been used on this report; V = surrogate recovery is not within method/contract control limits.
(2) R(%) = percent recovery.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form2.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-2_Form2_QE7Q_077S21_SJ2204916.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 1A

PHARMACEUTICALS ANALYSIS REPORT

CLIENT SAMPLE NO.
SP-OCHLCK
Sample Collection:
17-Apr-2017 12:15

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-3
Matrix:	AQUEOUS	**Sample Size:**	0.969 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 06:03:37	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 22**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng/L		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Alprazolam	U		0.310 (L)	
Amitriptyline		2.52	0.602 (S)	22:30
Amlodipine	U		1.55 (L)	
Benzoylecgonine	U		0.310 (L)	
Benztropine	U		0.516 (L)	
Betamethasone	U		1.55 (L)	
Cocaine		0.211	0.155 (L)	9:07
DEET		143	0.826 (L)	21:01
Desmethyldiltiazem	U		0.155 (L)	
Diazepam	U		0.310 (L)	
Fluocinonide	U		6.19 (L)	
Fluticasone propionate	U		2.06 (L)	
Hydrocortisone	U		61.9 (L)	
10-hydroxy-amitriptyline	U		0.212 (S)	
Meprobamate		13.8	4.13 (L)	11:27
Methylprednisolone	U		8.04 (S)	
Metoprolol	U		12.2 (S)	
Norfluoxetine	U		1.55 (L)	
Norverapamil	U		0.155 (L)	
Paroxetine	U		4.13 (L)	
Prednisolone	U		6.19 (L)	
Prednisone	U		53.5 (S)	
Promethazine	U		0.413 (L)	
Propoxyphene	U		0.310 (L)	
Propranolol	U		2.06 (L)	
Sertraline		0.947	0.413 (L)	26:36
Simvastatin	U		20.6 (L)	
Theophylline [3]	U		61.9 (L)	
Trenbolone	U		4.13 (L)	

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Trenbolone acetate	U		0.603 (S)	
Valsartan	U		4.13 (L)	
Verapamil	U		0.155 (L)	

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form1A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-3_Form1A_QE7Q_077S22_SJ2204917.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
SP-OCHLCK
Sample Collection:
17-Apr-2017 12:15

Form 2

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-3
Matrix:	AQUEOUS	**Sample Size:**	0.969 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 06:03:37	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 22**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng absolute		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC.	CONC. FOUND	R(%) [2]	RETENTION TIME
d5-Alprazolam		40.0	18.3	45.8	23:28
d6-Amitriptyline	V	40.0	53.4	134	22:38
d8-Benzoylecgonine		40.0	38.9	97.2	5:33
d3-Benztropine		20.0	10.6	52.9	23:11
d3-Cocaine		40.0	53.6	134	9:17
d7-DEET		40.0	53.6	134	20:45
d5-Diazepam		40.0	47.3	118	29:25
d4-Hydrocortisone		8000	10500	131	17:53
d3-Methylprednisolone		2000	2460	123	21:51
d7-Metoprolol		400	495	124	8:26
d5-Norfluoxetine	V	200	267	134	24:58
d6-Paroxetine		100	99.9	99.9	20:53
d4-Promethazine		100	133	133	18:52
d5-Propoxyphene	V	60.0	95.6	159	21:57
d7-Propranolol		400	462	115	14:44
13C1-15N2-Theophylline		2000	1180	59.0	2:33

(1) Where applicable, custom lab flags have been used on this report; V = surrogate recovery is not within method/contract control limits.
(2) R(%) = percent recovery.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form2.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-3_Form2_QE7Q_077S22_SJ2204917.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 1A

PHARMACEUTICALS ANALYSIS REPORT

CLIENT SAMPLE NO.
SP-APALACH
Sample Collection:
17-Apr-2017 09:55

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-4
Matrix:	AQUEOUS	**Sample Size:**	0.940 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 06:50:03	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 23**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng/L		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Alprazolam	U		0.319 (L)	
Amitriptyline		0.771	0.370 (S)	22:30
Amlodipine	U		1.60 (L)	
Benzoylecgonine	U		0.319 (L)	
Benztropine	U		0.532 (L)	
Betamethasone	U		1.60 (L)	
Cocaine	U		0.160 (L)	
DEET		3.55	0.851 (L)	21:01
Desmethyldiltiazem	U		0.160 (L)	
Diazepam	U		0.319 (L)	
Fluocinonide	U		6.38 (L)	
Fluticasone propionate	U		2.13 (L)	
Hydrocortisone	U		63.8 (L)	
10-hydroxy-amitriptyline	U		0.160 (L)	
Meprobamate	U		4.26 (L)	
Methylprednisolone	U		4.26 (L)	
Metoprolol	U		8.21 (S)	
Norfluoxetine	U		1.60 (L)	
Norverapamil	U		0.160 (L)	
Paroxetine	U		4.26 (L)	
Prednisolone	U		6.38 (L)	
Prednisone	U		21.3 (L)	
Promethazine	U		0.426 (L)	
Propoxyphene	U		0.319 (L)	
Propranolol	U		2.13 (L)	
Sertraline	U		0.426 (L)	
Simvastatin	U		21.3 (L)	
Theophylline [3]	U		63.8 (L)	
Trenbolone	U		4.26 (L)	

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Trenbolone acetate	U		0.319 (L)	
Valsartan	U		4.26 (L)	
Verapamil	U		0.160 (L)	

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form1A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-4_Form1A_QE7Q_077S23_SJ2204918.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
SP-APALACH
Sample Collection:
17-Apr-2017 09:55

Form 2

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-4
Matrix:	AQUEOUS	**Sample Size:**	0.940 L
Sample Receipt Date:	18-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 06:50:03	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 23**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng absolute		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC.	CONC. FOUND	R(%) [2]	RETENTION TIME
d5-Alprazolam		40.0	35.9	89.7	23:28
d6-Amitriptyline	V	40.0	52.6	132	22:38
d8-Benzoylecgonine		40.0	42.7	107	5:23
d3-Benztropine		20.0	11.4	56.9	23:03
d3-Cocaine		40.0	50.9	127	9:11
d7-DEET		40.0	52.2	130	20:45
d5-Diazepam		40.0	45.3	113	29:27
d4-Hydrocortisone		8000	8930	112	17:53
d3-Methylprednisolone		2000	1990	99.3	21:51
d7-Metoprolol		400	494	123	8:26
d5-Norfluoxetine		200	238	119	24:58
d6-Paroxetine		100	107	107	20:45
d4-Promethazine		100	122	122	18:52
d5-Propoxyphene	V	60.0	84.1	140	21:57
d7-Propranolol		400	454	113	14:37
13C1-15N2-Theophylline		2000	1250	62.7	2:29

(1) Where applicable, custom lab flags have been used on this report; V = surrogate recovery is not within method/contract control limits.
(2) R(%) = percent recovery.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form2.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-4_Form2_QE7Q_077S23_SJ2204918.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 1A

PHARMACEUTICALS ANALYSIS REPORT

CLIENT SAMPLE NO.
SP-ALAFIA
Sample Collection:
18-Apr-2017 10:15

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-5
Matrix:	AQUEOUS	**Sample Size:**	1.01 L
Sample Receipt Date:	19-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 07:36:45	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 24**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng/L		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Alprazolam	U		0.298 (L)	
Amitriptyline		3.49	0.373 (S)	22:30
Amlodipine	U		1.49 (L)	
Benzoylecgonine	U		0.298 (L)	
Benztropine	U		0.497 (L)	
Betamethasone	U		1.49 (L)	
Cocaine		0.198	0.149 (L)	8:57
DEET		34.8	0.795 (L)	21:01
Desmethyldiltiazem	U		0.149 (L)	
Diazepam	U		0.298 (L)	
Fluocinonide	U		5.96 (L)	
Fluticasone propionate	U		1.99 (L)	
Hydrocortisone	U		59.6 (L)	
10-hydroxy-amitriptyline	U		0.213 (S)	
Meprobamate	U		3.98 (L)	
Methylprednisolone	U		3.98 (L)	
Metoprolol	U		5.49 (S)	
Norfluoxetine	U		1.49 (L)	
Norverapamil	U		0.149 (L)	
Paroxetine	U		3.98 (L)	
Prednisolone	U		5.96 (L)	
Prednisone	U		19.9 (L)	
Promethazine	U		0.398 (L)	
Propoxyphene	U		0.298 (L)	
Propranolol	U		1.99 (L)	
Sertraline	U		0.398 (L)	
Simvastatin	U		19.9 (L)	
Theophylline [3]	U		59.6 (L)	
Trenbolone	U		3.98 (L)	

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Trenbolone acetate	U		0.298 (L)	
Valsartan	U		3.98 (L)	
Verapamil	U		0.149 (L)	

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form1A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-5_Form1A_QE7Q_077S24_SJ2204919.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
SP-ALAFIA
Sample Collection:
18-Apr-2017 10:15

Form 2

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	FL DEP EMERGING CONTAMINANTS ESOC PILOT		
Contract No.:	4114	**Lab Sample I.D.:**	L27146-5
Matrix:	AQUEOUS	**Sample Size:**	1.01 L
Sample Receipt Date:	19-Apr-2017	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 07:36:45	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 24**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng absolute		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC.	CONC. FOUND	R(%) [2]	RETENTION TIME
d5-Alprazolam		40.0	41.1	103	23:28
d6-Amitriptyline	V	40.0	53.5	134	22:38
d8-Benzoylecgonine		40.0	40.9	102	5:13
d3-Benztropine		20.0	25.2	126	23:03
d3-Cocaine		40.0	51.7	129	9:14
d7-DEET		40.0	51.8	130	20:45
d5-Diazepam		40.0	47.7	119	29:27
d4-Hydrocortisone		8000	8740	109	17:53
d3-Methylprednisolone		2000	1990	99.6	21:51
d7-Metoprolol		400	486	121	8:26
d5-Norfluoxetine		200	221	110	24:58
d6-Paroxetine		100	102	102	20:45
d4-Promethazine		100	127	127	18:52
d5-Propoxyphene	V	60.0	82.2	137	21:57
d7-Propranolol		400	469	117	14:37
13C1-15N2-Theophylline		2000	1400	69.8	2:26

(1) Where applicable, custom lab flags have been used on this report; V = surrogate recovery is not within method/contract control limits.
(2) R(%) = percent recovery.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form2.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_L27146-5_Form2_QE7Q_077S24_SJ2204919.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 1A

PHARMACEUTICALS ANALYSIS REPORT

CLIENT SAMPLE NO.
Lab Blank
Sample Collection:
N/A

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	N/A		
Contract No.:	4114	**Lab Sample I.D.:**	WG59329-101
Matrix:	AQUEOUS	**Sample Size:**	1.00 L
Sample Receipt Date:	N/A	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 03:44:03	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 19**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng/L		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Alprazolam	U		0.300 (L)	
Amitriptyline	U		0.633 (S)	
Amlodipine	U		1.50 (L)	
Benzoylecgonine	U		0.300 (L)	
Benztropine	U		0.500 (L)	
Betamethasone	U		1.50 (L)	
Cocaine	U		0.150 (L)	
DEET		1.84	0.800 (L)	21:01
Desmethyldiltiazem	U		0.150 (L)	
Diazepam	U		0.300 (L)	
Fluocinonide	U		6.00 (L)	
Fluticasone propionate	U		2.00 (L)	
Hydrocortisone	U		60.0 (L)	
10-hydroxy-amitriptyline	U		0.150 (L)	
Meprobamate	U		4.00 (L)	
Methylprednisolone	U		4.00 (L)	
Metoprolol	U		4.81 (S)	
Norfluoxetine	U		1.50 (L)	
Norverapamil	U		0.150 (L)	
Paroxetine	U		4.00 (L)	
Prednisolone	U		6.00 (L)	
Prednisone	U		20.0 (L)	
Promethazine	U		0.400 (L)	
Propoxyphene	U		0.300 (L)	
Propranolol	U		2.00 (L)	
Sertraline	U		0.400 (L)	
Simvastatin	U		20.0 (L)	
Theophylline [3]	U		60.0 (L)	
Trenbolone	U		4.00 (L)	

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG [1]	CONC. FOUND	REPORTING LIMIT (RL)[2]	RETENTION TIME
Trenbolone acetate	U		0.300 (L)	
Valsartan	U		4.00 (L)	
Verapamil	U		0.150 (L)	

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form1A.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_WG59329-101_Form1A_QE7Q_077S19_SJ2204913.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

CLIENT SAMPLE NO.
Lab Blank
Sample Collection:
N/A

Form 2

PHARMACEUTICALS ANALYSIS REPORT

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Project No.	N/A		
Contract No.:	4114	**Lab Sample I.D.:**	WG59329-101
Matrix:	AQUEOUS	**Sample Size:**	1.00 L
Sample Receipt Date:	N/A	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 03:44:03	**Column ID:**	C18MS
Extract Volume (uL):	4000	**Sample Data Filename:**	**QE7Q_077 S: 19**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13
Concentration Units:	ng absolute		

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC.	CONC. FOUND	R(%) [2]	RETENTION TIME
d5-Alprazolam		40.0	47.2	118	23:28
d6-Amitriptyline	V	40.0	57.7	144	22:30
d8-Benzoylecgonine		40.0	50.9	127	5:16
d3-Benztropine		20.0	27.6	138	23:03
d3-Cocaine	V	40.0	56.5	141	9:11
d7-DEET		40.0	52.4	131	20:45
d5-Diazepam		40.0	48.7	122	29:27
d4-Hydrocortisone		8000	11500	144	17:53
d3-Methylprednisolone		2000	2420	121	21:51
d7-Metoprolol	V	400	610	152	8:23
d5-Norfluoxetine	V	200	269	134	24:58
d6-Paroxetine		100	139	139	20:45
d4-Promethazine		100	133	133	18:44
d5-Propoxyphene	V	60.0	96.2	160	21:57
d7-Propranolol	V	400	610	152	14:34
13C1-15N2-Theophylline		2000	2200	110	2:29

(1) Where applicable, custom lab flags have been used on this report; V = surrogate recovery is not within method/contract control limits.
(2) R(%) = percent recovery.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

For Axys Internal Use Only [XSL Template: FC-Form2.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5;
Report Filename: LC_PPC_LC_PPC075_AEPX_WG59329-101_Form2_QE7Q_077S19_SJ2204913.html; Workgroup: WG59329; Design ID: 3092]

AXYS METHOD MLA-075 Rev 07

Form 8A

PHARMACEUTICALS ONGOING PRECISION AND RECOVERY (OPR)

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4114	**Lab Sample I.D.:**	WG59329-102
Matrix:	AQUEOUS	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 01:24:29	**Column ID:**	C18MS
Extract Volume (uL):	4000	**OPR Data Filename:**	**QE7Q_077 S: 16**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13

ALL CONCENTRATIONS REPORTED ON THIS FORM ARE CONCENTRATIONS IN EXTRACT, BASED ON A 1 mL EXTRACT VOLUME.

COMPOUND	LAB FLAG [1]	SPIKE CONC. (ng/mL)	CONC. FOUND (ng/mL)	% RECOVERY	RETENTION TIME
Alprazolam		15.0	14.6	97.0	23:42
Amitriptyline		15.0	15.7	104	22:38
Amlodipine		75.0	75.8	101	24:00
Benzoylecgonine		15.0	15.1	101	5:26
Benztropine		15.0	16.8	112	23:03
Betamethasone		75.0	80.1	107	21:49
Cocaine		7.50	6.59	87.9	9:11
DEET		30.0	29.8	99.5	21:10
Desmethyldiltiazem		7.50	7.87	105	19:00
Diazepam		15.0	15.8	105	29:43
Fluocinonide		300	344	115	35:26
Fluticasone propionate		100	67.4	67.4	37:40
Hydrocortisone		3000	2970	98.9	17:55
10-hydroxy-amitriptyline		7.50	6.10	81.4	12:08
Meprobamate		200	164	81.9	11:27
Methylprednisolone		200	202	101	21:51
Metoprolol		75.0	70.8	94.4	8:26
Norfluoxetine		75.0	73.5	98.0	24:58
Norverapamil		7.50	7.07	94.2	21:16
Paroxetine		200	211	106	20:53
Prednisolone		300	293	97.6	17:50
Prednisone		1000	930	93.0	17:05
Promethazine		20.0	18.8	94.4	18:52
Propoxyphene		15.0	14.1	94.1	22:05
Propranolol		100	107	107	14:47
Sertraline		20.0	17.4	87.0	26:36
Simvastatin		1000	607	60.7	39:39
Theophylline [2]		3000	6380	213	2:26
Trenbolone		200	240	120	23:03
Trenbolone acetate		15.0	12.4	82.7	37:26
Valsartan		200	145	72.4	32:35
Verapamil		7.50	6.98	93.0	21:41

(1) Where applicable, custom lab flags have been used on this report.
(2) Co-elutes with 1,7-Dimethylxanthine, maximum possible concentration is reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

These pages are part of a larger report that may contain information necessary for full data evaluation. Results reported relate only to the sample tested.

AXYS METHOD MLA-075 Rev 07

Form 8B

PHARMACEUTICALS ONGOING PRECISION AND RECOVERY (OPR)

SGS AXYS ANALYTICAL SERVICES
2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4114	**Lab Sample I.D.:**	WG59329-102
Matrix:	AQUEOUS	**Initial Calibration Date:**	27-Apr-2017
Extraction Date:	25-Apr-2017	**Instrument ID:**	LC MS/MS
Analysis Date:	28-Apr-2017 **Time:** 01:24:29	**Column ID:**	C18MS
Extract Volume (uL):	4000	**OPR Data Filename:**	**QE7Q_077 S: 16**
Injection Volume (uL):	15	**Blank Data Filename:**	QE7Q_077 S: 19
Dilution Factor:	N/A	**Cal. Ver. Data Filename:**	QE7Q_077 S: 13

ALL CONCENTRATIONS REPORTED ON THIS FORM ARE CONCENTRATIONS IN EXTRACT, BASED ON A 1 mL EXTRACT VOLUME.

LABELED COMPOUND	LAB FLAG [1]	SPIKE CONC. (ng/mL)	CONC. FOUND (ng/mL)	% RECOVERY	RETENTION TIME
d5-Alprazolam		40.0	37.1	92.7	23:28
d6-Amitriptyline		40.0	44.8	112	22:30
d8-Benzoylecgonine		40.0	43.9	110	5:36
d3-Benztropine		20.0	20.9	105	23:03
d3-Cocaine		40.0	45.8	115	9:07
d7-DEET		40.0	29.5	73.7	20:45
d5-Diazepam		40.0	39.9	99.8	29:25
d4-Hydrocortisone		8000	8400	105	17:53
d3-Methylprednisolone		2000	2280	114	21:51
d7-Metoprolol		400	502	126	8:20
d5-Norfluoxetine		200	202	101	24:58
d6-Paroxetine		100	105	105	20:45
d4-Promethazine		100	85.6	85.6	18:44
d5-Propoxyphene		60.0	69.5	116	21:57
d7-Propranolol		400	441	110	14:37
13C1-15N2-Theophylline		2000	2270	114	2:26

(1) Where applicable, custom lab flags have been used on this report.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Bryan Alonzo_____

These pages are part of a larger report that may contain information necessary for full data evaluation. Results reported relate only to the sample tested.

For Axys Internal Use Only [XSL Template: FC-Form8B.xsl; Created: 18-May-2017 11:03:21; Application: XMLTransformer-1.16.5; Report Filename: LC_PPC_LC_PPC075_AEPX_WG59329-102_Form8B_SJ2204909.html; Workgroup: WG59329; Design ID: 3092]

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
BFR	BTBPE	SGS AXYS MLA-033	MLA-033			Y										Y							Y										
	DBDPE	SGS AXYS MLA-033	MLA-033			Y										Y							Y										
	HBB	SGS AXYS MLA-033	MLA-033			Y										Y							Y										
	PBEB	SGS AXYS MLA-033	MLA-033			Y										Y							Y										
BPA and MPE	4,4'-dihydroxy-2,2-diphenylpropane (Bisphenol A) (BPA)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-(2-ethyl-5-hydroxyhexyl) phthalate (MEHHP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-(2-ethyl-5-oxohexyl) phthalate (MEOHP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-(3-carboxypropyl) phthalate (MCPP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-2-ethylhexyl phthalate (MEHP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-benzyl phthalate (MBzP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-butyl phthalate (MBP) (n + iso)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-cyclohexyl phthalate (MCHP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-ethyl phthalate (MEP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-iso-nonyl phthalate (MiNP)	SGS AXYS MLA-059	MLA-059																			Y											
	Mono-methyl phthalate (MMP)	SGS AXYS MLA-059	MLA-059																			Y											
FTS	4:2 fluorotelomer sulfonate (4:2 FTS)	SGS AXYS MLA-081	MLA-081																													YD	
		SGS AXYS MLA-089	MLA-089											YD																			
	6:2 fluorotelomer sulfonate (6:2 FTS)	SGS AXYS MLA-081	MLA-081																													YD	
		SGS AXYS MLA-089	MLA-089											YD																			
	8:2 fluorotelomer sulfonate (8:2 FTS)	SGS AXYS MLA-081	MLA-081																													YD	
		SGS AXYS MLA-089	MLA-089											YD																			
HBCDD	alpha-hexabromocyclododecane (a-HBCDD)	SGS AXYS MLA-070	MLA-070		Y																												
	beta-hexabromocyclododecane (b-HBCDD)	SGS AXYS MLA-070	MLA-070		Y																												
	gamma-hexabromocyclododecane (g-HBCDD)	SGS AXYS MLA-070	MLA-070		Y																												
OC Pesticides	2,4'-DDD	EPA 625	MLA-007																					Y									
		EPA 8270	MLA-007					Y					Y																					
		EPA 1699	MLA-028					Y						Y										Y								Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y	Y		Y							Y	Y						Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y								Y							Y	Y									
	2,4'-DDE	EPA 625	MLA-007																					Y									
		EPA 8270	MLA-007					Y					Y																					
		EPA 1699	MLA-028					Y						Y										Y								Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y	Y		Y							Y	Y						Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y								Y							Y	Y									
	2,4'-DDT	EPA 625	MLA-007																					Y									
		EPA 8270	MLA-007					Y					Y																					
		EPA 1699	MLA-028					Y						Y										Y								Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y	Y		Y							Y	Y						Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y								Y							Y	Y									
	4,4'-DDD	EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028					Y						Y										Y								Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y	Y		Y							Y	Y						Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y						Y		Y							Y	Y								Y	
	4,4'-DDE	EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028					Y						Y										Y								Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y	Y		Y							Y	Y						Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y						Y		Y							Y	Y								Y	
	4,4'-DDT	EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Aldrin	EPA 625	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Alpha-HCH	EPA 625	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Beta-HCH	EPA 625	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Chlordane, technical	EPA 625	MLA-007																				Y								Y		
		EPA 8270	MLA-007				Y	Y			Y	Y		Y	Y										Y			Y	Y		Y	Y	
		EPA 1699	MLA-028												Y																	Y	
		SGS AXYS MLA-007	MLA-007					Y							Y										Y								Y
	cis-Chlordane (alpha-Chlordane)	EPA 8270	MLA-007					Y			Y		Y	Y	Y										Y			Y			Y	Y	
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	cis-Nonachlor	EPA 8270	MLA-007					Y					Y												Y								
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Delta-HCH	EPA 608	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8081	MLA-007					Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Dieldrin	EPA 608	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Endosulphan I	EPA 608	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y						Y	Y		Y								Y
	Endosulphan II	EPA 608	MLA-007																				Y	Y			Y	Y		Y		Y
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y								Y
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y						Y	Y		Y						Y		Y

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-007	MLA-007		Y	Y		Y						Y	Y							Y		Y							Y	
	Endosulphan sulphate	EPA 608	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Endrin	EPA 608	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Endrin aldehyde	EPA 608	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y								Y		Y							Y	
	Endrin ketone	EPA 8081	MLA-007					Y			Y		Y		Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y			
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Gamma-HCH (Lindane)	EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Heptachlor	EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Heptachlor epoxide	EPA 608	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Hexachlorobenzene	EPA 1625	MLA-007																					Y	Y			Y				Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Methoxychlor	EPA 608	MLA-007																					Y				Y	Y			Y	
		EPA 8081	MLA-007				Y	Y			Y	Y	Y	Y	Y																		
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Mirex	EPA 8270	MLA-007					Y			Y		Y		Y										Y				Y	Y		Y	Y	
		EPA 1699	MLA-028					Y							Y										Y								Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	Oxychlordane	EPA 8270	MLA-007					Y					Y												Y									

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y								Y							Y		Y								
	Toxaphene	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	trans-Chlordane (gamma-Chlordane)	EPA 8270	MLA-007					Y			Y		Y	Y	Y										Y			Y			Y	Y	
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y							Y	Y							Y		Y							Y	
	trans-Nonachlor	EPA 8270	MLA-007					Y					Y												Y								
		EPA 1699	MLA-028					Y							Y										Y							Y	
		SGS AXYS MLA-028	MLA-028		Y	Y		Y					Y		Y	Y							Y		Y					Y		Y	
		SGS AXYS MLA-007	MLA-007		Y	Y		Y								Y									Y								
PAH	1,2,6-Trimethylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1,2-Dimethylnaphthalene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1,4,6,7-Tetramethylnaphthalene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1,7-Dimethylfluorene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1,7-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1,8-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1-Methylchrysene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1-Methylnaphthalene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	1-Methylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2,3,5-Trimethylnaphthalene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2,3,6-Trimethylnaphthalene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2,4-Dimethyldibenzothiophene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2,6-Dimethylnaphthalene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2,6-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2/3-Methyldibenzothiophenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2-Methylanthracene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2-Methylfluorene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	2-methylnaphthalene	EPA 1625	MLA-021																													Y	
		EPA 8270	MLA-021					Y			Y														Y								
		SGS AXYS MLA-021	MLA-021			Y		Y							Y								Y									Y	
	2-Methylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	3,6-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	3-Methylfluoranthene/ Benzo(a)fluorene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	3-Methylphenanthrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	5,9-Dimethylchrysene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	5/6-Methylchrysenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	7-Methylbenzo(a)pyrene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	9/4-Methylphenanthrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	Acenaphthene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y																						
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Acenaphthylene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y																						
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Anthracene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y																						
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Benz[a]anthracene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Benzo[a]pyrene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Benzo[b]fluoranthene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	Y
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Benzo[e]pyrene	SGS AXYS MLA-021	MLA-021			Y										Y							Y									
	Benzo[ghi]perylene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Benzo[j/k]fluoranthenes	SGS AXYS MLA-021	MLA-021																				Y										
	Benzo[k]fluoranthene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	Y
		EPA 8270	MLA-021					Y			Y			Y	Y																		
		SGS AXYS MLA-021	MLA-021					Y							Y	Y							Y		Y							Y	
	Biphenyl	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Acenaphthenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Benz(a)anthracenes/chrysenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Benzofluoranthenes/ Benzopyrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Biphenyls	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Dibenzothiophene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Fluoranthenes/Pyrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Fluorenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Naphthalenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C1-Phenanthrenes/Anthracenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Benz(a)anthracenes/Chrysenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Benzofluoranthenes/ Benzopyrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Biphenyls	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Dibenzothiophene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Fluoranthenes/Pyrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Fluorenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Naphthalenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C2-Phenanthrenes/Anthracenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C3-Benz(a)anthracenes/Chrysenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C3-Dibenzothiophene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C3-Fluoranthenes/Pyrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C3-Fluorenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C3-Naphthalenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C3-Phenanthrenes/Anthracenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C4-Benz(a)anthracenes/Chrysenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C4-Dibenzothiophene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C4-Fluoranthenes/Pyrenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C4-Naphthalenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	C4-Phenanthrenes/Anthracenes	SGS AXYS MLA-021	MLA-021			Y																	Y										
	Chrysene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Dibenz[ah]anthracene	SGS AXYS MLA-021	MLA-021			Y										Y							Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Dibenzo[ah]anthracene	EPA 1625	MLA-021																				Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021					Y							Y										Y							Y	
	Dibenzothiophene	SGS AXYS MLA-021	MLA-021			Y																	Y										
	Fluoranthene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Fluorene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Indeno[1,2,3-cd]pyrene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Naphthalene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Perylene	SGS AXYS MLA-021	MLA-021			Y										Y							Y										
	Phenanthrene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Pyrene	EPA 1625	MLA-021																					Y	Y			Y	Y			Y	
		EPA 8270	MLA-021				Y	Y			Y	Y		Y	Y																		
		SGS AXYS MLA-021	MLA-021			Y		Y							Y	Y							Y		Y							Y	
	Retene	SGS AXYS MLA-021	MLA-021			Y																	Y										
PBDPE	BDE 10 2,6-dibromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 100 2,2',4,4',6-pentabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 105 2,3,3',4,4'-pentabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 11 3,3'-dibromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 116 2,3,4,5,6-pentabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 119 2,3',4,4',6-pentabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 12 3,4-dibromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 126 3,3',4,4',5-pentabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 13 3,4'-dibromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 140 2,2',3,4,4',6'-hexabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 15 4,4'-dibromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 153 2,2',4,4',5,5'-hexabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 154 2,2',4,4',5',6-hexabromodiphenylether	EPA 1614	MLA-033										Y																		Y		
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	BDE 155 2,2',4,4',6,6'-hexabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 166 2,3,4,4',5,6-hexabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 17 2,2',4-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 181 2,2',3,4,4',5,6-heptabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 190 2,3,3',4,4',5,6-heptabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 206 2,2',3,3',4,4',5,5',6-nonabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 207 2,2',3,3',4,4',5,6,6'-nonabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 208 2,2',3,3',4,5,5',6,6'-nonabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 209 Decabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 25 2,3',4-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 28 2,4,4'-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 30 2,4,6-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 35 3,3',4-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 37 3,4,4'-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 47 2,2',4,4'-tetrabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 49 2,2',4,5'-tetrabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 66 2,3',4,4'-tetrabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 7 2,4-dibromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 75 2,4,4',6-tetrabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 77 3,3',4,4'-tetrabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 8 2,4'-dibromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 85 2,2',3,4,4'-pentabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE 99 2,2',4,4',5-pentabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE-183 2,2',3,4,4',5',6-heptabromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										
	BDE-33 2',3,4-tribromodiphenylether	EPA 1614	MLA-033										Y																	Y			
		SGS AXYS MLA-033	MLA-033		Y	Y										Y							Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
PCB Aroclors	PCB Aroclor 1016	EPA 1668	MLA-010					Y						Y	Y									Y						Y	Y	
		EPA 625	MLA-007																				Y	Y			Y	Y			Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y			Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y									Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1016/1242	EPA 8270	MLA-007										Y																				
	PCB Aroclor 1221	EPA 1668	MLA-010					Y						Y	Y										Y						Y	Y	
		EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y		Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y										Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1232	EPA 1668	MLA-010					Y						Y	Y										Y						Y	Y	
		EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y		Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y										Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1242	EPA 1668	MLA-010					Y						Y	Y										Y						Y	Y	
		EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y			Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y										Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1248	EPA 1668	MLA-010					Y						Y	Y										Y						Y	Y	
		EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y		Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y										Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1254	EPA 1668	MLA-010					Y						Y	Y										Y						Y	Y	
		EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y		Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y										Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1260	EPA 1668	MLA-010					Y						Y	Y										Y						Y	Y	
		EPA 625	MLA-007																					Y	Y			Y	Y		Y	Y	
		EPA 8270	MLA-007				Y	Y			Y	Y	Y		Y																		
		SGS AXYS MLA-010	MLA-010					Y							Y										Y							Y	
		SGS AXYS MLA-007	MLA-007			Y		Y							Y	Y							Y		Y							Y	
	PCB Aroclor 1268	SGS AXYS MLA-007	MLA-007			Y										Y							Y										
PCB congeners	PCB 1 2-Chlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 10 2,6-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 100 2,2',4,4',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 101 2,2',4,5,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 101/90/89	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 102 2,2',4,5,6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-010	MLA-010		Y	Y		Y						YD	Y					YD		Y		Y							YD	
	PCB 103 2,2',4,5',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 104 2,2',4,6,6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 105 2,3,3',4,4'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 105/127	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 106 2,3,3',4,5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 107 2,3,3',4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 107/109	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 108 2,3,3',4,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 109 2,3,3',4,6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 11 3,3'-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 110 2,3,3',4',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 111 2,3,3',5,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 111/117	EPA 8270	MLA-007										Y																				
	PCB 112 2,3,3',5,6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 113 2,3,3',5',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 114 2,3,4,4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 115 2,3,4,4',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 116 2,3,4,5,6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 117 2,3,4',5,6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 118 2,3',4,4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 118/106	EPA 8270	MLA-007										Y																				

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 119 2,3',4,4',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 12 3,4-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 12/13	EPA 8270	MLA-007										Y																				
	PCB 120 2,3',4,5,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 121 2,3',4,5',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 122 2,3,3',4',5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 123 2,3',4,4',5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 124 2,3',4',5,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 125 2,3',4',5',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 126 3,3',4,4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 127 3,3',4,5,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 128 2,2',3,3',4,4'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 129 2,2',3,3',4,5-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 13 3,4'-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 130 2,2',3,3',4,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 131 2,2',3,3',4,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 131/142	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 132 2,2',3,3',4,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 132/168	EPA 8270	MLA-007										Y																				

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	PCB 133 2,2',3,3',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 134 2,2',3,3',5,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 134/143	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 135 2,2',3,3',5,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 136 2,2',3,3',6,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 137 2,2',3,4,4',5-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 138 2,2',3,4,4',5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 138/163/164	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 139 2,2',3,4,4',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 14 3,5-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 140 2,2',3,4,4',6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 141 2,2',3,4,5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 142 2,2',3,4,5,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 143 2,2',3,4,5,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 144 2,2',3,4,5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 144/135	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 145 2,2',3,4,6,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 146 2,2',3,4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 147 2,2',3,4',5,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 148 2,2',3,4',5,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 149 2,2',3,4',5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 149/139	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 15 4,4'-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 150 2,2',3,4',6,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 151 2,2',3,5,5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 152 2,2',3,5,6,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 153 2,2',4,4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 154 2,2',4,4',5,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 155 2,2',4,4',6,6'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 156 2,3,3',4,4',5-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 157 2,3,3',4,4',5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 158 2,3,3',4,4',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 158/160	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 159 2,3,3',4,5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y			YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 16 2,2',3-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 16/32	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 160 2,3,3',4,5,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 161 2,3,3',4,5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 162 2,3,3',4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 163 2,3,3',4',5,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 164 2,3,3',4',5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 165 2,3,3',5,5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 166 2,3,4,4',5,6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 167 2,3',4,4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 168 2,3',4,4',5',6-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 169 3,3',4,4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 17 2,2',4-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 170 2,2',3,3',4,4',5-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 170/190	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 171 2,2',3,3',4,4',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 172 2,2',3,3',4,5,5'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 172/192	EPA 8270	MLA-007										Y																				

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 173 2,2',3,3',4,5,6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 174 2,2',3,3',4,5,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 174/181	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 175 2,2',3,3',4,5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 176 2,2',3,3',4,6,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 177 2,2',3,3',4,5',6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 178 2,2',3,3',5,5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 179 2,2',3,3',5,6,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 18 2,2',5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 180 2,2',3,4,4',5,5'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 181 2,2',3,4,4',5,6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 182 2,2',3,4,4',5,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 183 2,2',3,4,4',5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 184 2,2',3,4,4',6,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 185 2,2',3,4,5,5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 186 2,2',3,4,5,6,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 187 2,2',3,4',5,5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-901	MLA-901		Y																													
	PCB 187/182	EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 188 2,2',3,4',5,6,6'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 189 2,3,3',4,4',5,5'-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 19 2,2',6-Trichlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 190 2,3,3',4,4',5,6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 191 2,3,3',4,4',5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 192 2,3,3',4,5,5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 193 2,3,3',4',5,5',6-Heptachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 194 2,2',3,3',4,4',5,5'-Octachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
		SGS AXYS MLA-901	MLA-901		Y																													
	PCB 195 2,2',3,3',4,4',5,6-Octachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 196 2,2',3,3',4,4',5,6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 196/203	EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 197 2,2',3,3',4,4',6,6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	YD						YD				Y			Y	Y	Y	Y	YD	Y
		EPA 8270	MLA-007											Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	PCB 198 2,2',3,3',4,5,5',6-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 199 2,2',3,3',4,5,5',6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 2 3-Chlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 20 2,3,3'-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 200 2,2',3,3',4,5,6,6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 201 2,2',3,3',4,5',6,6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 202 2,2',3,3',5,5',6,6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 203 2,2',3,4,4',5,5',6-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 204 2,2',3,4,4',5,6,6'-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 205 2,3,3',4,4',5,5',6-Octachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 206 2,2',3,3',4,4',5,5',6-Nonachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 207 2,2',3,3',4,4',5,6,6'-Nonachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 208 2,2',3,3',4,5,5',6,6'-Nonachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 209 Decachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 21 2,3,4-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	PCB 22 2,3,4'-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 23 2,3,5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 23/34	EPA 8270	MLA-007										Y																				
	PCB 24 2,3,6-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 24/27	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 25 2,3',4-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 26 2,3',5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 27 2,3',6-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 28 2,4,4'-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 29 2,4,5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 3 4-Chlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 30 2,4,6-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 31 2,4',5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 32 2,4',6-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 33 2,3',4'-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 33/20/21	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 34 2,3',5'-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 35 3,3',4-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 36 3,3',5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-010	MLA-010		Y	Y		Y						YD	Y					YD		Y		Y							YD	
	PCB 37 3,4,4'-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 38 3,4,5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 39 3,4',5-Trichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 4 2,2'-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 4/10	EPA 8270	MLA-007										Y																				
	PCB 40 2,2',3,3'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 41 2,2',3,4-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 41/71/64/68	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 42 2,2',3,4'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 42/59	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 43 2,2',3,5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 44 2,2',3,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 45 2,2',3,6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 46 2,2',3,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 47 2,2',4,4'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 47/48/75	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 48 2,2',4,5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 49 2,2',4,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 49/43	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 5 2,3-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-010	MLA-010		Y	Y		Y						YD	Y					YD		Y		Y							YD	
	PCB 50 2,2',4,6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 51 2,2',4,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 52 2,2',5,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 52/73	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 53 2,2',5,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 54 2,2',6,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 55 2,3,3',4-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 56 2,3,3',4'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 56/60	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 57 2,3,3',5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 58 2,3,3',5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 59 2,3,3',6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 6 2,3'-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 60 2,3,4,4'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 61 2,3,4,5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 62 2,3,4,6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 62/65	EPA 8270	MLA-007										Y																				
	PCB 63 2,3,4',5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 64 2,3,4',6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 65 2,3,5,6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	PCB 66 2,3',4,4'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 66/80	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 67 2,3',4,5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 68 2,3',4,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 69 2,3',4,6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 7 2,4-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 7/9	EPA 8270	MLA-007										Y																				
	PCB 70 2,3',4',5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 70/76	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 71 2,3',4',6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 72 2,3',5,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 73 2,3',5',6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 74 2,4,4',5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-901	MLA-901		Y																												
	PCB 74/61	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 75 2,4,4',6-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 76 2,3',4',5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 77 3,3',4,4'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 78 3,3',4,5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 79 3,3',4,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 8 2,4'-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 8/5	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 80 3,3',5,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-010	MLA-010		Y	Y		Y						YD	Y					YD		Y		Y							YD	
	PCB 81 3,4,4',5-Tetrachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 82 2,2',3,3',4-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 83 2,2',3,3',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 83/108	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 84 2,2',3,3',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 85 2,2',3,4,4'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 85/120	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 86 2,2',3,4,5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 87 2,2',3,4,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 87/115/116	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 88 2,2',3,4,6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 88/121	EPA 8270	MLA-007										Y																				
	PCB 89 2,2',3,4,6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 9 2,5-Dichlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y		YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 90 2,2',3,4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 91 2,2',3,4',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 92 2,2',3,5,5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 93 2,2',3,5,6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 94 2,2',3,5,6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 95 2,2',3,5',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 95/93	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	PCB 96 2,2',3,6,6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 97 2,2',3,4',5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 97/86	EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	PCB 98 2,2',3,4',6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
	PCB 98/102	EPA 8270	MLA-007										Y																				
	PCB 99 2,2',4,4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y		Y	Y	Y	Y	Y	YD						YD				Y		Y	Y	Y	Y		YD	Y
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y		Y							YD	Y					YD		Y		Y							YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
		SGS AXYS MLA-901	MLA-901		Y																													
	PCB congeners, total	EPA 1668	MLA-010								Y																		Y					
	Sum - Dichlorobiphenyls (BZ-12-+ BZ-13)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Heptachlorobiphenyls (BZ-171 + BZ-173)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Heptachlorobiphenyls (BZ-180 + BZ-193)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Heptachlorobiphenyls (BZ-183 + BZ-185)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-128 + BZ-166)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-129 + BZ-138 + BZ-160 + BZ-163)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-134 + BZ-143)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-135 + BZ-151 + BZ-154)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-139 + BZ-140)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-147 + BZ-149)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-153 + BZ-168)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Hexachlorobiphenyls (BZ-156 + BZ-157)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Pentachlorobiphenyls (BZ-107 + BZ-124)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Pentachlorobiphenyls (BZ-108 + BZ-124)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Pentachlorobiphenyls (BZ-110 + BZ-115)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Pentachlorobiphenyls (BZ-83 + BZ-99)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	
	Sum - Pentachlorobiphenyls (BZ-85 + BZ-116 + BZ-117)	EPA 1668	MLA-010											YD						YD												YD	
		SGS AXYS MLA-010	MLA-010											YD						YD												YD	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Sum - Pentachlorobiphenyls (BZ-86 + BZ-87 + BZ 97 + BZ-109 + BZ-119 + BZ-125)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Pentachlorobiphenyls (BZ-86 + BZ-87 + BZ-97 + BZ-108 + BZ-119 +BZ-125)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Pentachlorobiphenyls (BZ-88 + BZ-91)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Pentachlorobiphenyls (BZ-90 + BZ-101 + BZ-113)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Pentachlorobiphenyls (BZ-93 + BZ-95 + BZ-98 + BZ-100 + BZ-102)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-40 + BZ-41 + BZ-71)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-44 + BZ-47 + BZ-65)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-45 + BZ-51)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-49 + BZ-69)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-50 + BZ-53)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-59 + BZ-62 + BZ-75)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Tetrachlorobiphenyls (BZ-61 + BZ-70 + BZ-74 + BZ-76)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Trichlorobiphenyls (BZ-18 + BZ-30)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Trichlorobiphenyls (BZ-20 + BZ-28)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Trichlorobiphenyls (BZ-21 + BZ-33)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Sum - Trichlorobiphenyls (BZ-26 + BZ-29)	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010												YD						YD											YD	
	Total Dichlorobiphenyls	EPA 1668	MLA-010												YD						YD											YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y									YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	Total Heptachlorobiphenyls	EPA 1668	MLA-010												YD						YD											YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y									YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	Total Hexachlorobiphenyls	EPA 1668	MLA-010												YD						YD											YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y									YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										
	Total Monochlorobiphenyls	EPA 1668	MLA-010												YD						YD											YD	
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y									YD	
	Total Nonachlorobiphenyls	EPA 1668	MLA-010												YD						YD											YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y									YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Total Octachlorobiphenyls	EPA 1668	MLA-010												YD					YD												YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y										YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y											
	Total PCBs	EPA 1668	MLA-010												YD					YD												YD	
		SGS AXYS MLA-010	MLA-010												YD					YD												YD	
	Total Pentachlorobiphenyls	EPA 1668	MLA-010												YD					YD												YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y										YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y											
	Total Polychlorinated biphenyls	SGS AXYS MLA-007	MLA-007			Y										Y							Y											
	Total Tetrachlorobiphenyls	EPA 1668	MLA-010												YD					YD												YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y										YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y											
	Total Trichlorobiphenyls	EPA 1668	MLA-010												YD					YD												YD	
		EPA 8270	MLA-007										Y																				
		SGS AXYS MLA-010	MLA-010		Y	Y									YD	Y					YD		Y										YD	
		SGS AXYS MLA-007	MLA-007			Y										Y							Y											
PCDDF	1,2,3,4,6,7,8-HpCDD	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,4,6,7,8-HpCDF	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,4,7,8,9-HpCDF	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,4,7,8-HxCDD	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,4,7,8-HxCDF	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,6,7,8-HxCDD	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,6,7,8-HxCDF	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,7,8,9-HxCDD	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,7,8,9-HxCDF	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,7,8-PeCDD	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	1,2,3,7,8-PeCDF	EPA 1613	MLA-017										Y		YD					YD				Y	Y		Y	Y	Y	Y	Y	YD	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	2,3,4,6,7,8-HxCDF	EPA 1613	MLA-017										Y		YD						YD			Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	2,3,4,7,8-PeCDF	EPA 1613	MLA-017										Y		YD						YD			Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	2,3,7,8-TCDD	EPA 1613	MLA-017										Y		YD						YD			Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	2,3,7,8-TCDF	EPA 1613	MLA-017										Y		YD						YD			Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	OCDD	EPA 1613	MLA-017										Y		YD						YD			Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	OCDF	EPA 1613	MLA-017										Y		YD						YD			Y	Y		Y	Y	Y	Y	Y	YD	
		EPA 8290	MLA-017				Y	Y		Y	Y	Y		Y	YD		Y		Y	Y	YD				Y			Y	Y			YD	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y		Y							YD	Y	Y				YD		Y		Y							YD	
	Total HpCDD	EPA 1613	MLA-017												YD						YD			Y	Y						Y	YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	
	Total HpCDF	EPA 1613	MLA-017												YD						YD			Y	Y						Y	YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	
	Total HxCDD	EPA 1613	MLA-017												YD						YD			Y	Y						Y	YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	
	Total HxCDF	EPA 1613	MLA-017												YD						YD			Y	Y						Y	YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	
	Total PCDD	EPA 1613	MLA-017												YD						YD												YD	
		EPA 8290	MLA-017												YD						YD												YD	
		SGS AXYS MLA-017	MLA-017												YD						YD												YD	
	Total PCDD+PCDF	EPA 1613	MLA-017												YD						YD												YD	
		EPA 8290	MLA-017												YD						YD												YD	
		SGS AXYS MLA-017	MLA-017												YD						YD												YD	
	Total PCDF	EPA 1613	MLA-017												YD						YD												YD	
		EPA 8290	MLA-017												YD						YD												YD	
		SGS AXYS MLA-017	MLA-017												YD						YD												YD	
	Total PeCDD	EPA 1613	MLA-017												YD						YD												YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	
	Total PeCDF	EPA 1613	MLA-017												YD						YD												YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	
	Total TCDD	EPA 1613	MLA-017												YD						YD												YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y						Y		YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y								YD	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids									Tissue						Urine	Water	Water, Non-Potable								
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Total TCDF	EPA 1613	MLA-017												YD					YD			Y	Y						Y	YD	
		EPA 8290	MLA-017					Y		Y		Y		Y	YD		Y		Y	Y	YD				Y				Y			YD	
		SGS AXYS MLA-017	MLA-017					Y							YD		Y				YD				Y							YD	
PFC	Perfluorobutanesulfonate (PFBS)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorobutanoate (PFBA)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorodecanoate (PFDA)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorododecanoate (PFDoA)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluoroheptanoate (PFHpA)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorohexanesulfonate (PFHxS)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorohexanoate (PFHxA)	EPA 537 modified	MLA-041												YD																	
			MLA-043																	YD												
			MLA-060																													YD	
		SGS AXYS MLA-060	MLA-060																			Y			Y	Y	Y					YD	
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD																		

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorononanoate (PFNA)	EPA 537 modified	MLA-041												YD																		
			MLA-043																		YD												
			MLA-060																														YD
		SGS AXYS MLA-060	MLA-060																				Y										YD
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD										Y	Y	Y						
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorooctane sulfonamide (PFOSA)	EPA 537 modified	MLA-041												YD																		
			MLA-043																		YD												
			MLA-060																														YD
		SGS AXYS MLA-060	MLA-060																				Y										YD
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD										Y	Y	Y						
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorooctanesulfonate (PFOS)	EPA 537 modified	MLA-041												YD																		
			MLA-043																		YD												
			MLA-060																														YD
		SGS AXYS MLA-060	MLA-060																				Y										YD
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD										Y	Y	Y						
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluorooctanoate (PFOA)	EPA 537 modified	MLA-041												YD																		
			MLA-043																		YD												
			MLA-060																														YD
		SGS AXYS MLA-060	MLA-060																				Y										YD
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD										Y	Y	Y						
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluoropentanoate (PFPeA)	EPA 537 modified	MLA-041												YD																		
			MLA-043																		YD												
			MLA-060																														YD
		SGS AXYS MLA-060	MLA-060																				Y										YD
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD										Y	Y	Y						
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
	Perfluoroundecanoate (PFUnA)	EPA 537 modified	MLA-041												YD																		
			MLA-043																		YD												
			MLA-060																														YD
		SGS AXYS MLA-060	MLA-060																				Y										YD
		SGS AXYS MLA-041	MLA-041			Y		Y	Y	Y					YD										Y	Y	Y						
		SGS AXYS MLA-043	MLA-043													Y	Y	Y	Y		YD												
		SGS AXYS MLA-042	MLA-042		Y																												
PPCP	1,7-Dimethylxanthine	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	10-hydroxy-amitriptyline	SGS AXYS MLA-075	MLA-075			Y																	Y										
	2-hydroxy-ibuprofen	SGS AXYS MLA-075	MLA-075			Y																	Y										
	4-Epianhydrochlortetracycline (EACTC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable									
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	4-Epianhydrotetracycline (EATC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	4-Epichlortetracycline (ECTC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	4-Epioxytetracycline (EOTC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	4-Epitetracycline (ETC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Acetaminophen	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Albuterol	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Alprazolam	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Amitriptyline	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Amlodipine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Amphetamine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Anhydrochlortetracycline (ACTC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Anhydrotetracycline (ATC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Atenolol	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Atorvastatin	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Azithromycin	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Benzoylecgonine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Benztropine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Betamethasone	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Bisphenol A	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Caffeine	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Carbadox	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Carbamazepine	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Cefotaxime	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Chlortetracycline (CTC)	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Cimetidine	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Ciprofloxacin	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Clarithromycin	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Clinafloxacin	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Clonidine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Cloxacillin	EPA 1694	MLA-075											Y																	Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Cocaine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Codeine	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Cotinine	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	DEET (N,N-diethyl-m-toluamide)	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Dehydronifedipine	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Demeclocycline	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Desmethyldiltiazem	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Diazepam	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Digoxigenin	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Digoxin	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Diltiazem	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Diphenhydramine	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Doxycycline	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Enalapril	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Enrofloxacin	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Erythromycin	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Erythromycin anydrate	EPA 1694	MLA-075											Y																		Y		
	Flumequine	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Fluocinonide	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Fluoxetine	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Fluticasone propionate	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Furosemide	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Gemfibrozil	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Glipizide	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Glyburide	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Hydrochlorothiazide	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Hydrocodone	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Hydrocortisone	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Ibuprofen	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Isochlortetracycline (ICTC)	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Lincomycin	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Lomefloxacin	EPA 1694	MLA-075											Y																		Y		
		SGS AXYS MLA-075	MLA-075			Y																	Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Meprobamate	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Metformin	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Methylprednisolone	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Metoprolol	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Miconazole	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Minocycline	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Naproxen	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Norfloxacin	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Norfluoxetine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Norgestimate	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Norverapamil	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Ofloxacin	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Ormetoprim	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Oxacillin	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Oxolinic acid	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Oxycodone	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Oxytetracycline (OTC)	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Paroxetine	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Penicillin G	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Penicillin V	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Prednisolone	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Prednisone	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Promethazine	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Propoxyphene	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Propranolol	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Ranitidine	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Roxithromycin	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sarafloxacin	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sertraline	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Simvastatin	SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sulfachloropyridazine	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sulfadiazine	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sulfadimethoxine	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sulfamerazine	EPA 1694	MLA-075											Y																		Y	
		SGS AXYS MLA-075	MLA-075			Y																	Y										
	Sulfamethazine	EPA 1694	MLA-075											Y																		Y	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids									Tissue						Urine	Water	Water, Non-Potable								
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Sulfamethizole	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Sulfamethoxazole	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Sulfanilamide	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Sulfathiazole	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Tetracycline (TC)	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Theophylline	SGS AXYS MLA-075	MLA-075			Y																Y										
	Thiabendazole	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Trenbolone	SGS AXYS MLA-075	MLA-075			Y																Y										
	Trenbolone acetate	SGS AXYS MLA-075	MLA-075			Y																Y										
	Triamterene	SGS AXYS MLA-075	MLA-075			Y																Y										
	Triclocarban	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Triclosan	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Trimethoprim	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Tylosin	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Valsartan	SGS AXYS MLA-075	MLA-075			Y																Y										
	Verapamil	SGS AXYS MLA-075	MLA-075			Y																Y										
	Virginiamycin	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
	Warfarin	EPA 1694	MLA-075											Y																	Y	
		SGS AXYS MLA-075	MLA-075			Y																Y										
Targeted Metabolites	11, 14, 17-eicosatrienoic acid (eicosatrienoic acid)	SGS AXYS MLM-001	MLM-001												Y																	
	11, 14-eicosadienoic acid	SGS AXYS MLM-001	MLM-001												Y																	
	3-hydroxytyrosine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Acetylcarnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Acetylornithine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Alanine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	alpha-Aminoadipic acid	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Arginine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Asparagine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Aspartate	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Asymmetric dimethylarginine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Butenylcarnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Butyrylcarnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	C22:5 ISOMER 1 (tentatively all-cis-4, 8, 12, 15, 19-docosapentaenoic acid)	SGS AXYS MLM-001	MLM-001												Y																	
	C22:5 ISOMER 2 (all-cis-7,10,13,16,19-docosapentaenoic acid (DPA)	SGS AXYS MLM-001	MLM-001												Y																	
	C22:5 ISOMER 3 (tentatively all-cis-4, 7, 10, 13, 16-docosapentaenoic acid)	SGS AXYS MLM-001	MLM-001												Y																	
	Carnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Carnosine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	chenodeoxycholic acid	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	cholic acid	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Citrulline	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Creatinine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Decadienylcarnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	decanoic acid (capric acid)	SGS AXYS MLM-001	MLM-001												Y																	
	Decanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	Decenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y										Y						Y											
	deoxycholic acid	SGS AXYS MLM-001	MLM-001		Y										Y						Y											

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	docosahexaenoic acid (DHA)	SGS AXYS MLM-001	MLM-001													Y																
	docosatetraenoic acid (adrenic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Dodecanedioylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Dodecanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Dodecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Dopamine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	eicosapentaenoic acid (EPA)	SGS AXYS MLM-001	MLM-001													Y																
	Eicosatetraenoic acid (arachidonic acid)	SGS AXYS MLM-001	MLM-001													Y																
	eicosatrienoic acid (dihomo-γ-linolenic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Glutaconylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Glutamate	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Glutamine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Glutarylcarnitine (Hydroxyhexanoylcarnitine)	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Glycine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	glycochenodeoxycholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	glycocholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	glycodeoxycholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hexadecadienylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	hexadecanoic acid (palmitic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Hexadecanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	hexadecenoic acid (palmitoleic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Hexadecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hexanoylcarnitine (Fumarylcarnitine)	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hexenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hexose (sum isomers)	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Histamine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Histidine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxyhexadecadienylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxyhexadecanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxyhexadecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxylbutyrylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxyoctadecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxyproline	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxypropionylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxysphingomyeline C14:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxysphingomyeline C16:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxysphingomyeline C22:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxysphingomyeline C22:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxysphingomyeline C24:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxytetradecadienylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxytetradecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Hydroxyvalerylcarnitine (Methylmalonylcarnitine)	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Isoleucine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Kynurenine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Leucine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lithocholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Lysine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C14:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C16:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C16:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C17:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C18:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C18:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C18:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C20:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C20:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C24:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C26:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	lysoPhosphatidylcholine acyl C28:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	lysoPhosphatidylcholine acyl C28:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Methionine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Methioninesulfoxide	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Methylglutarylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Nitrotyrosine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Nonaylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	octadecadienoic acid (linoleic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Octadecadienylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	octadecanoic acid (stearic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Octadecanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	octadecatrienoic acid (γ-linolenic acid)	SGS AXYS MLM-001	MLM-001													Y																
	Octadecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Octanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Ornithine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phenylalanine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phenylethylamine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C30:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C30:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C30:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C32:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C32:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C34:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C34:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C34:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C34:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C36:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C36:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C36:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C36:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C36:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C36:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C38:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C38:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C38:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C38:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C38:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C38:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C40:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C40:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C40:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C40:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C40:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C40:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C42:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C42:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C42:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C42:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C42:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C42:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C44:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C44:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C44:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine acyl-alkyl C44:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C24:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C26:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C28:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C30:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
				CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	Phosphatidylcholine diacyl C30:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C32:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C32:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C32:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C32:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C34:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C34:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C34:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C34:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C36:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C38:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C38:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C38:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C38:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C38:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C38:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C40:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C40:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C40:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C40:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C40:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C40:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C42:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C42:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C42:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C42:4	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C42:5	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Phosphatidylcholine diacyl C42:6	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Pimelylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Proline	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Propenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Propionylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Putrescine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sarcosine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Serine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Serotonin	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Spermidine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Spermine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C16:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C16:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C18:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C18:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C20:2	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C22:3	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C24:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C24:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C26:0	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Sphingomyeline C26:1	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Symmetric dimethylarginine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	Taurine	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	taurochenodeoxycholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y										
	taurocholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y										

Accreditation Scope

SGS AXYS Analytical Services Ltd.
(formerly AXYS Analytical Services Ltd.)
file ref.: ACC-101 Rev. 33

				Pulp	Serum	Solids										Tissue						Urine	Water	Water, Non-Potable								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE*	Maine DOH	ANAB	Pennsylvania DEP
	taurodeoxycholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	taurolithocholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	tauroursodexoycholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Tetradecadienylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	tetradecanoic acid (myristic acid)	SGS AXYS MLM-001	MLM-001													Y																	
	Tetradecanoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Tetradecenoylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Threonine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Tiglylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Total dimethylarginine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Tryptophan	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Tyrosine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	ursodexoycholic acid	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Valerylcarnitine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
	Valine	SGS AXYS MLM-001	MLM-001		Y											Y						Y											
TBBPA	Tetrabromobisphenol A	SGS AXYS MLA-079	MLA-079		Y																												

Note*
Analysis of pesticides and PCBs in non-potable water samples by AXYS method MLA-007, with the exception of NPDES or State permitted discharges and Stormwater applications, may fall within the scope of Washington State Department of Ecology solids matrix accreditation, subject to approval of the Ecology Project Manager.

Legend

Y	Accreditation scope
YD	Accreditation scope, including US DOD scope
BFR	Brominated flame retardants (non-PBDPE)
BPA and mPE	Bisphenol A and mono-Phthalate Esters
FTS	Fluorotelomer sulfonates
HBCDD	Hexabromocyclododecane
OC Pesticides	Organochlorine Pesticides
PAH	Polycyclic Aromatic Hydrocarbons
PBDPE	Polybrominated diphenylethers
PCB	Polychlorinated Biphenyls
PCDDF	Polychlorinated dibenzodioxins/furans
PFC	Perfluorinated Compounds
PPCP	Pharmaceutical and Personal Care Products
TBBPA	Tetrabromobisphenol A
California DPH	California Department of Public Health, Lab ID 2911 (target analytes shown are those approved 2014)
Florida DOH	Florida Department of Health, Lab ID E871007, (NELAC Standard)
Pennsylvania DEP	Pennsylvania Department of Environmental Protection
Minnesota DOH	Minnesota Department of Health, Lab ID 232-999-430, (NELAC Standard)
New Jersey DEP	New Jersey Department of Environmental Protection, Lab ID CANA005, (NELAC Standard)
New York DOH	New York Department of Health, Lab ID 11674, (NELAC Standard)
Washington DE	Washington Department of Ecology, Lab ID C404
Virginia DGS	Virginia Department of General Services, Division of Consolidated Laboratory Services, Lab ID 460224, (NELAC Standard)
Maine DOH	Maine Center for Disease Control and Prevention, Department of Health and Human Services, Lab ID CN00003

CALA	Canadian Association for Laboratory Accreditation Inc., Lab ID A2637, (ISO/IEC 17025:2005 Standard)
ANAB	ANSI-ASQ National Accreditation Board, certificate ADE-1861, (ISO/IEC 17025:2005 and US DOD Standards)