

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00976-Q+00970

1-6-06 version

Date 1/5/10 ^{1/5/11} Calibration Time (Pre Sample): 12:20 Post Sample Time: 11:00 on 1/6/11

Circle/Fill In: Lab Calibration or Calibration on Site yes

Temperature (Monthly)

Date of Last Temperature Verification _____

Pass or Fail
P F

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 11.00 mg/L @ Temp. 11.96 °C % DO 101.9 (P) F

Calibrated Reading: 10.74 mg/L @ Temp. 12.04 °C % DO 99.6 (P) F

Pre Sample

Initial Calib. Verification: 10.55 mg/L @ Temp. 12.30 °C % DO 98.5 (P) F

Post Sample

Continuing Calib. Verification 9.40 mg/L @ Temp. 17.86 °C % DO 99.0 (P) F

Specific Conductance

In Calibrate Mode

Calibration: Exp. Date 06/11 Lot # 2912477 Standard 1000 Initial Reading 974 Calibrated Reading 999 (P) F
μmhos/cm μmhos/cm μmhos/cm

Pre Sample

Initial Calib. Verification: 10/11 1004434 100 102 (P) F
μmhos/cm μmhos/cm μmhos/cm

Post Sample

Continuing Calib. Verification 06/11 2912477 1000 1008 (P) F
μmhos/cm μmhos/cm μmhos/cm

pH

Standard Exp. Date Lot # Initial Reading Temperature Calibrated Reading Temperature
In Calibrate Mode In Calibrate Mode

7.00 SU 11/11 096106 7.11 SU 11.11 °C 6.99 SU 11.12 °C (P) F

Initial Calibration Verification:

In Read Mode

Pre Sample 10/06 SU 09/11 095143 10.07 SU 11.01 °C (P) F

Continuing Calibration Verification:

Post Sample 7.00 SU 11/11 096106 6.98 SU 12.60 °C (P) F

Maintenance

Specific Conductance Probe Cleaned? Y (N)

Notes:

Dissolved Oxygen Membrane Changed? Y (N)

-Car buoy cleaned with Iquinox
on 1/4/11

Site ID: Hornsby

Project: SPR61101

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ±5%) (pH ±0.2SU) (Temperature ±0.2°C) Warm bath Appx. 40°C

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00976-Q+00970

1-6-06 version

Date ~~4/5/10~~ ^{1/5/11} Calibration Time (Pre Sample): 12:20 Post Sample Time: 11:00 on 1/6/11

Circle/Fill In:
Temperature (Monthly)

Lab Calibration or Calibration on Site yes

Date of Last Temperature Verification

Pass or Fail
P F

Ice Bath: Probe °C NIST °C Warm Bath: Probe °C NIST °C

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 11.00 mg/L @ Temp. 11.96 °C % DO 101.9 P F

Calibrated Reading: 10.74 mg/L @ Temp. 12.04 °C % DO 99.6 P F

Pre Sample

Initial Calib. Verification: 10.55 mg/L @ Temp. 12.30 °C % DO 98.5 P F

Post Sample

Continuing Calib. Verification 9.40 mg/L @ Temp. 17.86 °C % DO 99.0 P F

Specific Conductance

In Calibrate Mode

Calibration: Exp. Date 06/11 Lot # 2912477/1000 Standard Initial Reading 974 Calibrated Reading 999 P F

Pre Sample

Initial Calib. Verification: 10/11 1004434 100 Initial Reading 102 Calibrated Reading P F

Post Sample

Continuing Calib. Verification 06/11 2912477 1000 Initial Reading 1008 Calibrated Reading P F

pH

Standard Exp. Date Lot # Initial Reading Temperature Calibrated Reading Temperature

Initial Calibration: 7.00 SU 11/11 096106 7.11 SU 11.11 °C 6.99 SU 11.12 °C P F

Initial Calibration Verification: In Read Mode

Pre Sample 10.00 SU 09/11 095143 10.07 SU 11.01 °C P F

Continuing Calibration Verification:

Post Sample 7.00 SU 11/11 096106 6.98 SU 12.60 °C P F

Maintenance

Notes:

Specific Conductance Probe Cleaned? Y N

Dissolved Oxygen Membrane Changed? Y N

-Car buoy cleaned with Iquinox
on 1/4/11

Site ID: Hornsby

Project: SPR61101

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.2 °C) Warm bath Appx. 40 °C