

Figure 23. Calibration Log

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: Q003629

September 2008 version

Boldly "X" this
box if there is
qualified data on
this page.

Date 7/1/2011 Calibration Time: 8:00 Project Name: wakulla springs 1/4 1/4 ↓

All Times Are: ETZ or CTZ (circle one)

Circle/Fill In: Lab Calibration or Calibration on Site _____

Temperature (Monthly) Date of Last Temperature Verification _____

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 8.54 mg/L @ Temp. 23.11 °C % DO 100 (P) F

Calibrated Reading: 8.55 mg/L @ Temp. 23.11 °C % DO 100 (P) F

In Run Mode

Initial Calib. Verification: 8.56 mg/L @ Temp. 23.12 °C % DO 100 (P) F

Specific Conductance

In Calibrate Mode

Calibration:

Exp. Date	Lot #	Standard	Initial Reading	Calibrated Reading	
<u>8/2011</u>	<u>2002463</u>	<u>100</u>	<u>96</u>	<u>100</u>	(P) F
<u>8/2011</u>	<u>2002464</u>	<u>1000</u>	<u>1008</u>	<u>1001</u>	(P) F

In Run Mode

Initial Calib. Verification: 8/2011 2002463 100 98 100 (P) F
8/2011 2002464 1000 1004 1001 (P) F

pH

Standard Exp. Date

Initial Calibration:

Lot #	Initial Reading	Temperature	Calibrated Reading	Temperature	
<u>108043</u>	<u>7.04</u> SU	<u>22.97</u> °C	<u>7.01</u> SU	<u>22.98</u> °C	(P) F

4.0 SU _____ SU _____ °C _____ SU _____ °C P F

10.0 SU 11/2012 107444 10.03 SU 22.89 °C 10.00 SU 22.90 °C (P) F

Initial Calibration Verification:

In Run Mode

7.0 SU 12/2012 108043 7.01 SU 23.00 °C (P) F

Maintenance: Weekly PH Slope: _____

Specific Conductance Probe Cleaned? Y N

Notes:

Dissolved Oxygen Membrane Changed? Y N

Site ID: wakulla
Project: springs 1/4 1/4

Edgar Wade
Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.5 °C) Warm bath Appx. 40 °C
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places

CONTINUING CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)Meter ID: QD03629

September 2008 version

Boldly "X" this box if there is qualified data on this page.

Date 7/1/2011 Mid-Day Time: End-of-Day Time: 16:00

All Times Are: (ETZ) or CTZ (circle one)

Temperature (Monthly)Date of Last Temperature Verification

Ice Bath: Probe °C NIST °C Warm Bath: Probe °C NIST °C

Pass or Fail

P F

Dissolved Oxygen *In Run Mode*

Mid-Day mg/L @ Temp. °C % DO P F

End-of-Day 8.39 mg/L @ Temp. 24.21 °C % DO 100 (P) F

Specific Conductance *In Run Mode*

Exp. Date Lot # Standard Initial Reading

Mid-Day P F

 μ hos/cm μ hos/cm

End-of-Day 8/20/11 2002463 100 101 (P) F

8/20/11 2002464 1000 1002 (P) F

pH *In Run Mode*

Bracketing Exp. Date Lot # Initial Reading Temperature
Standard

Mid-Day SU P F

End-of-Day 7.0 SU 12/2012 108043 7.01 SU 24.23 (P) F

10.0 11/2012 107444 10.00 24.27 (P) F

Notes:

Site ID:
Project:

wakulla
springs 1/4

Maintenance: Specific Conductance Probe Cleaned? Y (N)
Dissolved Oxygen Membrane Changed: Y (N)

Edgar Wade
Print name (of person performing calibrations.)

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.5 °C) Warm bath Appx. 40 °C
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places