

Chemical Analysis Report

WTRSHD-MGT-2012-02-01-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SPRG1201**
Request ID: **RQ-2012-01-30-23**
Customer: **WTRSHD-MGT**
Project ID: **SPG_QTR_WQ**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd.
Tallahassee, FL 32399-2400
Attn: Kathryn Holland

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Kathryn Brackett
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 17-FEB-2012 15:05



Abbreviations and data remark codes

CERT # - NELAP (National Environmental Laboratory Accreditation Program) Certification Number of the laboratory that performed the analysis.

LCS – Laboratory Control Sample; in the QC Failures column, this notation indicates a batch recovery failure.

MS – Matrix Spike; in the QC Failures column, this notation indicates a batch recovery failure.

RPD – Relative Percent Difference; in the QC Failures column, this notation indicates a batch failure for precision.

CCV – Continuing Calibration Verification; in the QC Failures column, this notation indicates a failure of the calibration verification check sample.

RSD – Relative Standard Deviation expressed as a percentage.

SMP – Sample.

** - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.

A - Value reported is the mean of two or more determinations

B - Results based on colony counts outside the acceptable range.

I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.

J - Estimated value

K - Actual value is known to be less than value given

L - Actual value is known to be greater than value given

N - Presumptive evidence of presence of material.

O - Sampled, but analysis lost or not performed.

Q - Sample held beyond normal holding time.

T - Value reported is less than the criterion of detection.

U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.

V - Analyte was detected in both sample and method blank.

X - Too few individuals to calculate SCI value.

Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate

Z - Colonies were too numerous to count (TNTC).

Precision is reported as relative percent difference unless otherwise noted.

Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published test performance acceptance criteria.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

All sample collection performed by Bureau of Laboratories staff followed the field Standard Operating Procedures adopted by reference in Rule 62-160 FAC unless otherwise noted.

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Results for NELAP accredited tests contained in this report meet the requirements specified by the National Environmental Laboratory Accreditation Conference (NELAC). All samples received were in acceptable condition and met NELAC requirements unless otherwise noted. Results generated in this report pertain to the samples collected and submitted for analysis.

Job: TLH-2012-02-01-41

Job: TLH-2012-02-01-42

Job: TLH-2012-02-01-43

Group: Nutrients

Group: Metals

Group: Pesticides

Sample Location: BIG BLUE SPRING (WACISSA)**Collection Date/Time: 02/01/2012 11:50 AM****Field ID: BIG BLUE SPRING (WACISSA)****Matrix: W-OTHER**

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Failures
1406003	EPA 180.1 Rev. 2.0	Turbidity	0.25		NTU	P235645	
	EPA 300.0 Rev. 2.1	Chloride	5.7		mg Cl/L	P235981	
		Sulfate	6.4		mg SO4/L	P235972	
	SM 2120 B	Color (true)	2.5	U	PCU	P235628	
	SM 2320 B-97	Alkalinity	169		mg CaCO3/L	P235866	
	SM 2540 C-97	TDS	190		mg/L	P235960	
1406005	EPA 350.1 Rev. 2.0	Ammonia-N	0.010	U	mg N/L	P235626	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.11	I	mg N/L	P235750	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.13		mg N/L	P235695	
	EPA 365.1 Rev. 2.0	Total-P	0.050		mg P/L	P235708	
	SM 5310 B-00	Organic Carbon	1.0	U	mg C/L	P235650	
1406007	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.045		mg P/L	P235665	
1406009	EPA 200.7 Rev. 4.4	Boron	16	I	ug/L	P236031	
		Calcium	56.1		mg/L	P236031	
		Magnesium	8.30		mg/L	P236031	
		Potassium	0.46	I	mg/L	P236031	
		Sodium	3.1		mg/L	P236031	
1406011	DEP SOP: LC-001-2 (based on EPA 8321B)	Sucralose**	0.0095	U	ug/L	P235749	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: One of the matrix spike recoveries is unavailable due to high analyte concentration in the QC sample.

SM 2120 B: The precision data is unavailable due to the low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0: Total-P: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Sample Location: WACISSA SPRING # 2**Collection Date/Time: 02/01/2012 10:40 AM****Field ID: WACISSA SPRING # 2****Matrix: W-OTHER**

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Failures
1406002	EPA 180.1 Rev. 2.0	Turbidity	0.20		NTU	P235645	
	EPA 300.0 Rev. 2.1	Chloride	5.1		mg Cl/L	P235981	
		Sulfate	4.8		mg SO4/L	P235972	
	SM 2120 B	Color (true)	2.5	U	PCU	P235628	
	SM 2320 B-97	Alkalinity	141		mg CaCO3/L	P235866	
	SM 2540 C-97	TDS	161		mg/L	P235960	
1406004	EPA 350.1 Rev. 2.0	Ammonia-N	0.010	U	mg N/L	P235626	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P235750	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.46		mg N/L	P235695	
	EPA 365.1 Rev. 2.0	Total-P	0.11		mg P/L	P235708	
	SM 5310 B-00	Organic Carbon	1.0	U	mg C/L	P235650	
1406006	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.032		mg P/L	P235665	
1406008	EPA 200.7 Rev. 4.4	Boron	15	U	ug/L	P236031	
		Calcium	45.8	A	mg/L	P236031	
		Magnesium	8.57	A	mg/L	P236031	
		Potassium	0.46	I	mg/L	P236031	
		Sodium	2.9	A	mg/L	P236031	

Field ID: WACISSA SPRING # 2

Matrix: W-OTHER

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Failures
1406010	DEP SOP: LC-001-2 (based on EPA 8321B)	Sucralose**	0.0095	U	ug/L	P235749	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: One of the matrix spike recoveries is unavailable due to high analyte concentration in the QC sample.

SM 2120 B: The precision data is unavailable due to the low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0: Total-P: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: DEP SOP: LC-001-2 (based on EPA 8321B)

Batch ID: P235749

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P235645

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P236031

Component	Result	Code	Units
Boron	15	U	ug/L
Calcium	0.075	U	mg/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235972

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235981

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P235626

Component	Result	Code	Units
Ammonia-N	0.010	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P235750

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P235695

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P235708

Component	Result	Code	Units
Total-P	0.005	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P235665

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B

Batch ID: P235628

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97

Batch ID: P235866

Component	Result	Code	Units
Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97

Batch ID: P235960

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 5310 B-00

Batch ID: P235650

Component	Result	Code	Units
Organic Carbon	1.0	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: DEP SOP: LC-001-2 (based on EPA 8321B)

Batch ID: P235749

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	80.1	87.5	P/P	40 - 130

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P236031

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P236031

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Boron	96.8		P	85 - 115
Calcium	103		P	85 - 115
Magnesium	102		P	85 - 115
Potassium	104		P	85 - 115
Sodium	98.1		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235972

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235981

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P235626

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P235750

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	91.9		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P235695

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P235708

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	107		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P235665

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.4		P	90 - 110

Reference Method: SM 2320 B-97

Batch ID: P235866

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-97

Batch ID: P235866

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Alkalinity	100		P	85 - 115

Reference Method: SM 5310 B-00

Batch ID: P235650

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: DEP SOP: LC-001-2 (based on EPA 8321B)

Batch ID: P235749

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406010	Sucralose	85.4	77.2	P/P	40 - 130

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P236031

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406008	Boron	103	105	P/P	80 - 120
1406008	Calcium	97.4		P	80 - 120
1406008	Magnesium	99.3		P	80 - 120
1406008	Potassium	104	103	P/P	80 - 120
1406008	Sodium	106	105	P/P	80 - 120
1406400	Boron	103		P	80 - 120
1406400	Calcium	98.2		P	80 - 120
1406400	Magnesium	97.4		P	80 - 120
1406400	Potassium	103		P	80 - 120
1406400	Sodium	104		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235972

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406026	Sulfate	94.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235981

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406026	Chloride	97.0		P	90 - 110
1406148	Chloride	96.7	96.6	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P235626

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406004	Ammonia-N	97.8	99.8	P/P	90 - 110

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P235750

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406079	Kjeldahl Nitrogen	105	106	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P235695

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1405972	NO2NO3-N	96.0	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P235665

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1406007	O-Phosphate-P	96.2	102	P/P	90 - 110

Reference Method: SM 5310 B-00

Batch ID: P235650

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1404395	Organic Carbon	98.1	98.7	P/P	85 - 115

Quality Assurance Report

Precision

Reference Method: DEP SOP: LC-001-2 (based on EPA 8321B)

Batch ID: P235749

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406010	Sucralose	10.1	Spike	P	0 - 30
LFB	Sucralose	8.83	LCS	P	0 - 30

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P235645

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406072	Turbidity	0.653	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P236031

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406008	Boron	1.86	Spike	P	0 - 20
1406008	Calcium	2.67	Sample	P	0 - 20
1406008	Magnesium	1.30	Sample	P	0 - 20
1406008	Potassium	1.47	Spike	P	0 - 20
1406008	Sodium	3.60	Sample	P	0 - 20

Quality Assurance Report

Precision

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235972

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406148	Sulfate	0.731	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P235981

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406148	Chloride	0.0110	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P235626

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406004	Ammonia-N	2.07	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P235750

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406079	Kjeldahl Nitrogen	0.389	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P235695

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1405972	NO2NO3-N	2.65	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P235708

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406004	Total-P	2.43	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P235665

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406007	O-Phosphate-P	3.69	Spike	P	0 - 20

Reference Method: SM 2320 B-97

Batch ID: P235866

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406015	Alkalinity	0.930	Sample	P	0 - 20
1406026	Alkalinity	0.852	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2540 C-97

Batch ID: P235960

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1406406	TDS	4.44	Sample	P	0 - 10

Reference Method: SM 5310 B-00

Batch ID: P235650

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1404395	Organic Carbon	0.610	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 1406010

Field Sample ID: WACISSA SPRING # 2

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
DEP SOP: LC-001-2 (based on EPA 8321B)	Sucralose-d6	48.5	P	40 - 150

Lab Sample ID: 1406011

Field Sample ID: BIG BLUE SPRING (WACISSA)

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
DEP SOP: LC-001-2 (based on EPA 8321B)	Sucralose-d6	47.4	P	40 - 150

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A43355

Included Lab Sample IDs: 1406004, 1406005

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.8	98.6	P/P	90 - 110
Ammonia-N	99.3	97.6	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A43356

Included Lab Sample IDs: 1406002, 1406003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	95.0	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A43360

Included Lab Sample IDs: 1406002, 1406003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A43360

Included Lab Sample IDs: 1406002, 1406003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	102	P/P	95 - 105

Reference Method: SM 5310 B-00

Run ID: A43361

Included Lab Sample IDs: 1406004, 1406005

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	99.0	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A43367

Included Lab Sample IDs: 1406006, 1406007

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.0	98.2	P/P	90 - 110
O-Phosphate-P	98.8	98.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A43381

Included Lab Sample IDs: 1406004, 1406005

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A43390

Included Lab Sample IDs: 1406004, 1406005

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	105	103	P/P	90 - 110

Reference Method: DEP SOP: LC-001-2 (based on EPA 8321B)

Run ID: A43410

Included Lab Sample IDs: 1406010, 1406011

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	100	103	P/P	70 - 140
Sucralose	108	100	P/P	70 - 140

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A43411

Included Lab Sample IDs: 1406004, 1406005

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.0	93.6	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A43464

Included Lab Sample IDs: 1406002, 1406003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report

Calibration Verification

Reference Method: SM 2320 B-97

Run ID: A43464

Included Lab Sample IDs: 1406002, 1406003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Alkalinity	99.7	98.0	P/P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A43516

Included Lab Sample IDs: 1406002, 1406003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	98.3	97.4	P/P	90 - 110
Sulfate	104	94.7	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A43543

Included Lab Sample IDs: 1406008, 1406009

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Boron	98.6	99.0	P/P	90 - 110
Boron	99.0	97.5	P/P	90 - 110
Calcium	102	104	P/P	90 - 110
Calcium	104	102	P/P	90 - 110
Magnesium	100	99.3	P/P	90 - 110
Magnesium	102	100	P/P	90 - 110
Potassium	101	102	P/P	90 - 110
Potassium	102	104	P/P	90 - 110
Sodium	101	102	P/P	90 - 110
Sodium	102	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report

Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			LCS	Precision SMP	MS
DEP SOP: LC-001-2 (based on EPA 8321B)	Sucralose	80.1	87.5	85.4	77.2		8.83		10.1
EPA 180.1 Rev. 2.0	Turbidity							0.653	
EPA 200.7 Rev. 4.4	Boron	96.8		103	105	103			1.86
	Calcium	103		97.4	98.2			2.67	
	Magnesium	102		99.3	97.4			1.30	
	Potassium	104		104	103	103			1.47
	Sodium	98.1		106	105	104		3.60	
EPA 300.0 Rev. 2.1	Chloride	97.3		96.7	96.6	97.0			0.0110
	Sulfate	97.4		94.0					0.731
EPA 350.1 Rev. 2.0	Ammonia-N	97.9		97.8	99.8				2.07
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	91.9		105	106				0.389
EPA 353.2 Rev. 2.0	NO2NO3-N	99.5		96.0	99.0				2.65
EPA 365.1 Rev. 2.0	Total-P	107							2.43
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.4		96.2	102				3.69

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	Precision		
				LCS	SMP	MS
SM 2320 B-97	Alkalinity	100			0.852 0.930	
SM 2540 C-97	TDS				4.44	
SM 5310 B-00	Organic Carbon	97.2	98.1 98.7			0.610

Reference Method Descriptions

Method / NELAC Cert. #	Description
DEP SOP: LC-001-2 (based on EPA 8321B) / E31780	Analysis of sucralose in water matrices by HPLC/MS/MS
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
SM 2120 B / E31780	True Color measured at 450 nm or Apparent Color
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: LC-001-2 (based on EPA 8321B)	02/01/2012	02/02/2012	Hoor Shaik	02/02/2012	S. M. Reddy	1406010, 1406011
EPA 180.1 Rev. 2.0	02/01/2012			02/02/2012 16:31	Bryan A. Werth	1406002, 1406003
EPA 200.7 Rev. 4.4	02/01/2012	02/08/2012	Elliott D. Healy	02/14/2012	Elisa J Lutz	1406008, 1406009
EPA 300.0 Rev. 2.1	02/01/2012			02/07/2012	Gary Dearman	1406002, 1406003
EPA 350.1 Rev. 2.0	02/01/2012			02/02/2012	Diana M. Dunbar	1406004, 1406005
EPA 351.2 Rev. 2.0	02/01/2012	02/03/2012	Bryan A. Werth	02/06/2012	Pallavi R. Gadwal	1406004, 1406005
EPA 353.2 Rev. 2.0	02/01/2012			02/03/2012	Sima Feizi	1406004, 1406005
EPA 365.1 Rev. 2.0	02/01/2012	02/03/2012	Pallavi R. Gadwal	02/06/2012	Pallavi R. Gadwal	1406004, 1406005
EPA 365.1 Rev. 2.0 dissolved	02/01/2012			02/02/2012 14:27	Pallavi R. Gadwal	1406007
	02/01/2012			02/02/2012 14:40	Pallavi R. Gadwal	1406006
SM 2120 B	02/01/2012			02/02/2012 12:37	Rochelle Y Jackson	1406002
	02/01/2012			02/02/2012 12:39	Rochelle Y Jackson	1406003
SM 2320 B-97	02/01/2012			02/09/2012	Dale L. Simmons	1406002, 1406003
SM 2540 C-97	02/01/2012			02/06/2012	Yijie Li	1406002, 1406003
SM 5310 B-00	02/01/2012			02/03/2012	Dale L. Simmons	1406004, 1406005