

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG

(FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00977-Q103272

1-6-06 version

Date 1/11/12 Calibration Time (Pre Sample): 10:30 Post Sample Time: 10:30 on 1/12/12

Circle/Fill In: Lab Calibration or Calibration on Site Yes

Temperature (Monthly)

Date of Last Temperature Verification

Pass or Fail
P F

Ice Bath: Probe °C NIST °C Warm Bath: Probe °C NIST °C

Dissolved Oxygen

In Calibrate Mode

Initial Reading:

8.86 mg/L @ Temp. 21.02 °C % DO 99.5 P F

Calibrated Reading:

8.76 mg/L @ Temp. 21.54 °C % DO 99.5 P F

Pre Sample

Initial Calib. Verification:

8.62 mg/L @ Temp. 21.57 °C % DO 98.0 P F

Post Sample

Continuing Calib. Verification

9.77 mg/L @ Temp. 17.30 °C % DO 101.7 P F

Specific Conductance

In Calibrate Mode
Calibration:

Exp. Date Lot # Standard
02/12 2008438 1000
μmhos/cm

Initial Reading

983
μmhos/cm

Calibrated Reading

1005 P F
μmhos/cm

Pre Sample

Initial Calib. Verification:

09/12 2103755 100
μmhos/cm

104
μmhos/cm

P F

Post Sample

Continuing Calib. Verification

02/12 2008438 1000
μmhos/cm

1002
μmhos/cm

P F

pH

Standard Exp. Date
Initial Calibration:

Lot #
In Calibrate Mode

Initial Reading Temperature

Calibrated Reading Temperature

7.00 SU 12/12 108043 7.16 SU 17.80 °C 7.00 SU 17.85 °C P F

Initial Calibration Verification:

In Read Mode

Pre Sample

10.00 SU

9.91 SU 17.87 °C

P F

Continuing Calibration Verification:

Post Sample

7.00 SU

12/12 108043 6.94 SU 15.62 °C

P F

Maintenance

Notes:

Specific Conductance Probe Cleaned? Y N

Dissolved Oxygen Membrane Changed? Y N

Site ID: Mill Pond Spring

Project: SPR61201

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.2 °C) Warm bath Appx. 40 °C