

Daily Turbidity Verification

ICV Use One Secondary Gel Standard

Meter ID	Date of Last Calibration	Print Samplers Names	Project Name
#2	7-8-13	Blair/Eckles	MSF Nutrient

Date/ Time	Lot #	Exp. Date	Standard	Reading	Pass?
11:00 8-20-13	A4279	11/13 11/13	541	541	Y / N
8-20-13 16:20	" "	11/13	6.10	5.85	Y / N

List Site ID's

MSFMW 3,2,20,6

List Site ID's

10:00 8-21-13	A4279	11/13	51.5	50.3	Y / N
8-21-13 17:40	A4279	11/13	6.10	6.30	Y / N

List Site ID's

MSFMW AAK7015

List Site ID's

10:05 8-22-13	A4279	11/13	541	550	Y / N
16:00 8-22-13	A4279	11/13	6.10	6.19	Y / N

List Site ID's

MSFMW 15,7

List Site ID's

Comments:
