

Daily Turbidity Verification

ICV Use One Secondary Gel Standard

Meter ID	Date of Last Calibration	Print Samplers Names	Project Name
#1	7/13	Blair/Eckles	MSFNS

Date/ Time	Lot #	Exp. Date	Standard	Reading	Pass?
8-26-13 13:56	A4279	11/13	6.10	6.77	Y / N
8-26-13	A4279	11/13	51.5	52.0	Y / N

List Site ID's

MSFMW 11, 1, 4, 13

List Site ID's

8-27-13 10:21	A4279	11/13	541	547	Y / N
8-27-13 17:30	A4279	11/13	6.10	5.89	Y / N

List Site ID's

MSFMW 5, 10, 9, 12, 8

List Site ID's

Poe Spring

8-28-13 10:36	A4279	11/13	541	537	Y / N
8-28-13 17:05	A4279	11/13	51.5	51.3	Y / N

List Site ID's

Blue Spring, Devil's Eye, Ginnie, Twin Spring

List Site ID's

Comments:
