

CALIBRATION and Continuing Calibration LOG

(FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD02228 - Q102027

1-6-06 version

Boldly "X" this box if there is qualified data on this page.

Date 7/20/09 Calibration Time (Pre Sample): 13:25 Post Sample Time: 11:24 on 7/21/09

Circle/Fill In:

Lab Calibration

or

Calibration on Site

yes

Temperature (Monthly)

Date of Last Temperature Verification _____

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading:

7.80 mg/L @ Temp. 29.52 °C % DO 102.2 P F

Calibrated Reading:

7.60 mg/L @ Temp. 29.68 °C % DO 98.9 P F

Pre Sample

Initial Calib. Verification:

7.43 mg/L @ Temp. 29.88 °C % DO 97.9 P F

Post Sample

Continuing Calib. Verification

7.95 mg/L @ Temp. 25.54 °C % DO 97.2 P F

Specific Conductance

4808213

In Calibrate Mode

Calibration:

Exp. Date

Lot #

Standard

Initial Reading

Calibrated Reading

08/09

4808213

1000

1001

1000

P F

Pre Sample

Initial Calib. Verification:

07/09

2807533

100

102

P F

Post Sample

Continuing Calib. Verification

08/09

4808213

1000

989

P F

pH

Standard Exp. Date

Lot #

Initial Reading

Temperature

Calibrated Reading

Temperature

Initial Calibration:

In Calibrate Mode

7.00 SU

01/10

082161

6.96

SU

29.90 °C

7.00

SU

29.93 °C

P F

Initial Calibration Verification:

In Read Mode

10.00 SU

05/10

082707

10.13

SU

30.24 °C

P F

Continuing Calibration Verification:

7.00 SU

01/10

082161

7.12

SU

27.44 °C

P F

Maintenance

Specific Conductance Probe Cleaned? Y N

Notes:

Dissolved Oxygen Membrane Changed? Y N

-Car buoy cleaned w/ liqinox on
7/20/09

Site ID: Falmouth

Project: SPR60907

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.2 °C) Warm bath Appx. 40°C