

Figure 23. Calibration Log (Example)

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD 02228 H4000 LAB QUANTA September 2008 version

Boldly "X" this box if there is qualified data on this page.

Date 07 MAR 2010 Calibration Time: 1635 Project Name: ROYAL CANADIAN MOUNTED POLICE

All Times Are: ETZ or CTZ (circle one)

Circle/Fill In: Lab Calibration or Calibration on Site

Temperature (Monthly) Date of Last Temperature Verification

Ice Bath: Probe °C NIST °C Warm Bath: Probe °C NIST °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 8.79 mg/L @ Temp. 20.35 °C %DO 98.7 (P) F 9.0

Calibrated Reading: 9.02 mg/L @ Temp. 20.37 °C %DO 100.0 (P) F

In Run Mode

Initial Calib. Verification: _____ mg/L @ Temp: _____ °C %DO _____ P F

Specific Conductance

In Calibrate Mode

Calibration: Aug 2011 2002464 1000 1029 1000 (P) F
µmhos/cm µmhos/cm µmhos/cm

In Run Mode

Initial Calib. Verification: Aug 2011 2002463 100 100 (P) F
µmhos/cm µmhos/cm

pH

Standard Exp. Date Lot # Initial Reading Temperature Calibrated Reading Temperature

Initial Calibration: In Calibrate Mode

7.0 SU 05/2011 092513 7.07 SU 20.21 °C 7.00 SU 20.21 °C (P) F

4.0 SU _____ SU _____ °C _____ SU _____ °C P F

10.0 SU 05/2011 092627 9.95 SU 20.32 °C 10.00 SU 20.32 °C (P) F

Initial Calibration Verification:

In Run Mode

_____ SU _____ SU _____ °C _____ P F

Maintenance: Weekly PH Slope: _____

Specific Conductance Probe Cleaned? (Y) N

Notes:

Dissolved Oxygen Membrane Changed? Y (N)

PASS

Site ID:

Project:

JAMES MCCLENN

Print name of person performing calibrations.

Figure 23. Calibration Log (Example)

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD 00976 4400 LAB QUANTA September 2008 version

Boldly "X" this box if there is qualified data on this page.

Date 07 MAR 2010 Calibration Time: 1900 Project Name: Royal & Co. Inc. SPRINGSAF0

All Times Are: (ETZ) or CTZ (circle one)

Circle/Fill In: (Lab Calibration) or Calibration on Site

Temperature (Monthly) Date of Last Temperature Verification

Ice Bath: Probe °C NIST °C Warm Bath: Probe °C NIST °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 8.18 mg/L @ Temp. 22.0 °C % DO 93.4

P (F) 8.74 TARGET

Calibrated Reading: 8.76 mg/L @ Temp. 21.8 °C % DO 99.6

(P) F 8.77

In Run Mode

Initial Calib. Verification: _____ mg/L @ Temp. _____ °C % DO _____

P F

Specific Conductance

In Calibrate Mode

Calibration:

Exp. Date	Lot #	Standard	Initial Reading	Calibrated Reading
<u>Aug 2011</u>	<u>2002464</u>	<u>1000</u>	<u>1025</u>	<u>1000</u>
		µmhos/cm	µmhos/cm	µmhos/cm

In Run Mode

Initial Calib. Verification: Aug 2011 2002464 100 97 (F) F

pH

Standard	Exp. Date	Lot #	Initial Reading	Temperature	Calibrated Reading	Temperature
----------	-----------	-------	-----------------	-------------	--------------------	-------------

Initial Calibration:

In Calibrate Mode

7.0 SU 05/2011 092513 6.75 SU 20.34 °C 7.00 SU 20.37 °C (P) F

4.0 SU _____ SU _____ °C _____ SU _____ °C P F

10.0 SU 05/2011 092627 10.15 SU 20.40 °C 10.00 SU 20.43 °C (P) F

Initial Calibration Verification:

In Run Mode

_____ SU _____ SU _____ °C P F

Maintenance: Weekly PH Slope: _____ Specific Conductance Probe Cleaned? Y(N)
Notes: Dissolved Oxygen Membrane Changed? Y(N)

PASS CALIBRATION

Site ID:
Project:

JAMES M'CLEAN

Print name of person performing calibrations.