

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG

(FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00977

1-6-06 version

Date 7/15/10 Calibration Time (Pre Sample): 12:32 Post Sample Time: 16:50

Circle/Fill In:

Lab Calibration

or

Calibration on Site

yes

Temperature (Monthly)

Date of Last Temperature Verification

Pass or Fail
P F

Ice Bath: Probe °C NIST °C Warm Bath: Probe °C NIST °C

Dissolved Oxygen

In Calibrate Mode

Initial Reading:

7.56 mg/L @ Temp. 30.73 °C % DO 101.2 P F

Calibrated Reading:

7.44 mg/L @ Temp. 30.75 °C % DO 99.8 P F

Pre Sample

Initial Calib. Verification:

7.40 mg/L @ Temp. 30.74 °C % DO 99.0 P F

Post Sample

Continuing Calib. Verification

7.19 mg/L @ Temp. 32.79 °C % DO 98.5 P F

7.10

Specific Conductance

In Calibrate Mode

Calibration:

Exp. Date 06/11 Lot # 2912477 Standard 1000
μmhos/cm

Initial Reading

991
μmhos/cm

Calibrated Reading

999 P F
μmhos/cm

Pre Sample

Initial Calib. Verification:

06/11 2912475 100
μmhos/cm

103
μmhos/cm

P F

Post Sample

Continuing Calib. Verification

06/11 2912475 100
μmhos/cm

103
μmhos/cm

P F

pH

Standard Exp. Date

Lot #

Initial Reading

Temperature

Calibrated Reading

Temperature

Initial Calibration:

In Calibrate Mode

7.00 SU 11/11 09096106 7.02 SU 32.18 °C 6.99 SU 32.16 °C P F

Initial Calibration Verification:

In Read Mode

Pre Sample

10.00 SU 09/11 095143 9.87 SU 32.63 °C

P F

Continuing Calibration Verification:

Post Sample

10.00 SU 09/11 095143 10.00 SU 34.54 °C

P F

Maintenance

Specific Conductance Probe Cleaned? Y N

Notes:

Dissolved Oxygen Membrane Changed? Y N

- buoy cleaned with HCL on 7/15/10

Site ID: Wacissa

Project: SPRG1007

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.2 °C) Warm bath Appx. 40 °C