

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG

(FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00976-Q+00970

1-6-06 version

Date 9/14/10 Calibration Time (Pre Sample): 9:30 Post Sample Time: 10:45 on 9/15/10

Circle/Fill In:

Lab Calibration

or

Calibration on Site yes

Temperature (Monthly)

Date of Last Temperature Verification _____

Pass or Fail
P F

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

Dissolved Oxygen

In Calibrate Mode

Initial Reading:

7.73 mg/L @ Temp. 27.09 °C % DO 97.7 (P) F

Calibrated Reading:

7.92 mg/L @ Temp. 27.14 °C % DO 99.8 (P) F

Pre Sample

Initial Calib. Verification:

7.94 mg/L @ Temp. 27.23 °C % DO 100.5 (P) F

Post Sample

Continuing Calib. Verification

8.18 mg/L @ Temp. 25.02 °C % DO 99.7 (P) F

Specific Conductance

In Calibrate Mode

Calibration:

Exp. Date

Lot #

Standard

Initial Reading

Calibrated Reading

06/11

2912477

1000

992

1001 (P) F

µmhos/cm

µmhos/cm

µmhos/cm

Pre Sample

Initial Calib. Verification:

10/11

1009434

100

100

(P) F

µmhos/cm

µmhos/cm

Post Sample

Continuing Calib. Verification

06/11

2912477

1000

1005

(P) F

µmhos/cm

µmhos/cm

pH

Standard Exp. Date

Lot #

Initial Reading

Temperature

Calibrated Reading

Temperature

In Calibrate Mode

In Calibrate Mode

7.00 SU

11/11

096106

7.04

SU

27.95 °C

6.99

SU 27.96 °C

(P) F

Initial Calibration Verification:

In Read Mode

Pre Sample

10.00 SU

09/11

095143

10.01

SU

27.82 °C

(P) F

Continuing Calibration Verification:

Post Sample

7.00 SU

11/11

096106

7.15

SU

26.33 °C

(P) F

Maintenance

Notes:

Specific Conductance Probe Cleaned? Y N

Dissolved Oxygen Membrane Changed? Y N

Site ID: Little Springs

Project:

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.2 °C) Warm bath Appx. 40 °C