

Boldly "X" this
box if there is
qualified data on
this page.

CALIBRATION and Continuing Calibration LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD02228-2+02027

1-6-06 version

Date: 10/12/10 Calibration Time (Pre Sample): 6:30 Post Sample Time: _____

Circle/Fill In: Lab Calibration or Calibration on Site _____

Temperature (Monthly)

Date of Last Temperature Verification _____

Ice Bath: Probe 7.39 °C NIST 7.50 °C Warm Bath: Probe 42.12 °C NIST 42.3 °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading: _____ mg/L @ Temp. _____ °C % DO _____ P F

Calibrated Reading: _____ mg/L @ Temp. _____ °C % DO _____ P F

Pre Sample

Initial Calib. Verification: _____ mg/L @ Temp. _____ °C % DO _____ P F

Post Sample

Continuing Calib. Verification _____ mg/L @ Temp. _____ °C % DO _____ P F

Specific Conductance

	Exp. Date	Lot #	Standard	Initial Reading	Calibrated Reading	
In Calibrate Mode				<u>992</u>		P F
Calibration:			µmhos/cm	µmhos/cm	µmhos/cm	

Pre Sample

Initial Calib. Verification: _____ µmhos/cm _____ µmhos/cm _____ P F

Post Sample

Continuing Calib. Verification _____ µmhos/cm _____ µmhos/cm _____ P F

pH

Standard	Exp. Date	Lot #	Initial Reading	Temperature	Calibrated Reading	Temperature	
7.00 SU							P F

Initial Calibration Verification:

In Read Mode

Pre Sample

SU _____ SU _____ °C _____ P F

Continuing Calibration Verification:

Post Sample

SU _____ SU _____ °C _____ P F

Maintenance

Specific Conductance Probe Cleaned? Y N

Notes:

Dissolved Oxygen Membrane Changed? Y N

- DO membrane changed on
9/8/10

- would not calibrate, given
to Mike for maintenance

Site ID:

Project:

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ± 5%) (pH ± 0.2SU) (Temperature ±0.2°C) Warm bath Appx. 40°C