

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG

(FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00977

1-6-06 version

Date: 10/27/10 Calibration Time (Pre Sample): 8:00 Post Sample Time: 7:15 on 10/28/10

Circle/Fill In: Lab Calibration or Calibration on Site Yes

Temperature (Monthly)

Date of Last Temperature Verification

Pass or Fail
P F

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

Dissolved Oxygen

In Calibrate Mode

Initial Reading:

8.35 mg/L @ Temp. 24.58 °C % DO 100.1 P F

Calibrated Reading:

8.30 mg/L @ Temp. 24.57 °C % DO 99.7 P F

Pre Sample

Initial Calib. Verification:

8.28 mg/L @ Temp. 24.59 °C % DO 99.5 P F

Post Sample

Continuing Calib. Verification

8.28 mg/L @ Temp. 24.78 °C % DO 99.8 P F

Specific Conductance

In Calibrate Mode

Calibration:

Exp. Date 06/11 Lot # 2912477 Standard 1000 Initial Reading 1031 Calibrated Reading 1002 P F
μmhos/cm μmhos/cm μmhos/cm

Pre Sample

Initial Calib. Verification:

10/11 1009434 100 100 P F
μmhos/cm μmhos/cm μmhos/cm

Post Sample

Continuing Calib. Verification

06/11 2912477 1000 980 P F
μmhos/cm μmhos/cm μmhos/cm

pH

Standard Exp. Date Lot # Initial Reading Temperature Calibrated Reading Temperature

Initial Calibration: In Calibrate Mode
7.00 SU 11/11 096106 7.03 SU 25.19 °C 6.99 SU 25.19 °C P F

Initial Calibration Verification:

In Read Mode

Pre Sample
10.00 SU 09/11 095143 9.89 SU 25.24 °C P F

Continuing Calibration Verification:

Post Sample
7.00 SU 11/11 096106 7.03 SU 25.94 °C P F

Maintenance

Notes:

Specific Conductance Probe Cleaned? Y N
Dissolved Oxygen Membrane Changed? Y N

Site ID: Double Keyhole Spring
Project: SPR61010

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.2 °C) Warm bath Appx. 40 °C