

Boldly "X" this box if there is qualified data on this page.

CALIBRATION and Continuing Calibration LOG

(FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: QD00977

1-6-06 version

Date 10/28/10 Calibration Time (Pre Sample): 7:15 Post Sample Time: 8:15 on 11/3/10

Circle/Fill In: Lab Calibration or Calibration on Site yes

Temperature (Monthly)

Date of Last Temperature Verification _____

Pass or Fail

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading:

8.28 mg/L @ Temp. 24.78 °C % DO 99.8 P F

Calibrated Reading:

8.26 mg/L @ Temp. 24.75 °C % DO 99.5 P F

Pre Sample

Initial Calib. Verification:

8.22 mg/L @ Temp. 24.49 °C % DO 98.5 P F

Post Sample

Continuing Calib. Verification

9.25 mg/L @ Temp. 19.57 °C % DO 100.6 P F

Specific Conductance

In Calibrate Mode

Calibration:

Exp. Date 06/11 Lot # 2912477 Standard 1000 Initial Reading 980 Calibrated Reading 1000 P F
μmhos/cm μmhos/cm μmhos/cm

Pre Sample

Initial Calib. Verification:

10/11 1009434 100 101 P F
μmhos/cm μmhos/cm

Post Sample

Continuing Calib. Verification

06/11 2912477 1000 1019 P F
μmhos/cm μmhos/cm

pH

Standard Exp. Date

Lot #

Initial Reading

Temperature

Calibrated Reading

Temperature

In Calibrate Mode

In Calibrate Mode

7.00 SU 11/11 096106 7.03 SU 25.94 °C 7.00 SU 25.89 °C P F

Initial Calibration Verification:

In Read mode

Pre Sample

10.00 SU 09/11 095143 9.95 SU 25.61 °C P F

Continuing Calibration Verification:

Post Sample

7.00 SU 11/11 096106 7.17 SU 19.92 °C P F

Maintenance

Notes:

Specific Conductance Probe Cleaned? Y N

Dissolved Oxygen Membrane Changed? Y N

Site ID: Gar Spring

Project: SPR61010

Laura Hester

Print name of person performing calibrations.

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ± 5%) (pH ± 0.2SU) (Temperature ±0.2°C) Warm bath Appx. 40°C