

Daily Turbidity Verification

ICV Use One Secondary Gel Standard

Meter ID	Date of Last Calibration	Print Samplers Names	Project Name
# 1	2/29/13	Chris Green Edgar Wade	MSFRNT Santa Fe Nutrient Study

Date/Time	Lot #	Exp. Date	Standard	Reading	Pass?
5/7/13 12:51	24641-01	5/29/13	3.91	3.98	(Y) / N
5/7/13 17:53	24641-02	5/29/13	51.0	51.1	(Y) / N

List Site ID's

~~MSFMW19 / MSFMW18 / MSFMW14 / MSFMW16~~

List Site ID's

5-8-13 6:49	24641-01	5-29-13	3.91 3.92	3.92	(Y) / N
5-8-13 19:16	24641-02	5-29-13	51.0	51.2	(Y) / N

List Site ID's

~~MSFMW17 / MSFMW7 / MSFMW15 / MSFMW12~~

List Site ID's

5-9-13 9:36	24641-02	5-29-13	51.0	50.8	(Y) / N
5-9-13 13:21	24641-01	5-29-13	3.91	3.89	(Y) / N

List Site ID's

~~MSFMW11 / MSFMW1 / MSFMW4~~

List Site ID's

Comments:

Edgar Wade