

Daily Turbidity Verification

ICV Use One Secondary Gel Standard

Meter ID	Date of Last Calibration	Print Samplers Names	Project Name
# 1	2-28-13	Chris Green Edgar Wade	Santa Fe Nutrient Study

Date/ Time	Lot #	Exp. Date	Standard	Reading	Pass?
5-14-13 12:34	24641-01	5-28-13	3.91	3.93	(Y) / N
5-14-13	24641-02	5-28-13	51.0	50.7	(Y) / N

List Site ID's

MSFMW13 / MSFMW9 / MSFMW8 / MSFMW10

List Site ID's

5-15-13 9:48	24641-01	5-28-13	3.91	3.88	(Y) / N
5-15-13	24641-02	5-28-13	51.0	51.3	(Y) / N

List Site ID's

MSFMW5 / MSFMW6 / MSFMW3 / MSFMW20 / MSFMW2

List Site ID's

					Y / N
					Y / N

List Site ID's

List Site ID's

Comments:

Edgar Wade