



MHP-4 MERRITTS MILL POND

SPRS1303 MHP-4




MERRITTS MILL POND #3; NW

SPRS1303 MHP-3



MERRITTS MILL POND #2; B

SPRS1303 MHP-2



MERRITTS MILL POND #1

SPRS1303 MHP-1

RQ-2013-02-11-44

SPRS1303

CALIBRATION and Continuing Calibration LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID:

1-6-06 version

Date 3/13/13 Calibration Time (Pre Sample): 7:00 Post Sample Time: 22:15

Circle/Fill In: Lab Calibration or Calibration on Site
Temperature (Monthly)

Date of Last Temperature Verification _____

Ice Bath: Probe _____ °C NIST _____ °C Warm Bath: Probe _____ °C NIST _____ °C

Pass or Fail
P F**Dissolved Oxygen**

In Calibrate Mode

Initial Reading: 9.28 mg/L @ Temp. 19.02 °C % DO 100.1 (P) FCalibrated Reading: 9.27 mg/L @ Temp. 19.01 °C % DO 100.0 (P) F

Pre Sample

Initial Calib. Verification: 9.27 mg/L @ Temp. 19.00 °C % DO 100.0 (P) F

Post Sample

Continuing Calib. Verification 9.41 mg/L @ Temp. 18.46 °C % DO 100.3 (P) F**Specific Conductance**

In Calibrate Mode

Calibration: 4/13 2111199 100 99 100 (P) F
μmhos/cm μmhos/cm μmhos/cm

Pre Sample

Initial Calib. Verification: 4/13 2111199 100 100 (P) F
μmhos/cm μmhos/cm μmhos/cm

Post Sample

Continuing Calib. Verification 4/13 2111199 100 100 (P) F
μmhos/cm μmhos/cm μmhos/cm**pH**

Standard Exp. Date Lot # Initial Reading Temperature Calibrated Reading Temperature

Initial Calibration:

In Calibrate Mode

7.00 SU 10/13 115471 6.98 SU 19.79 °C 7.00 SU 19.75 °C (P) F

Initial Calibration Verification:

In Read Mode

Pre Sample

7.00 SU 10/13 115471 7.00 SU 19.65 °C (P) F

Continuing Calibration Verification:

Post Sample

7.00 SU 10/13 115471 6.98 SU 20.14 °C (P) F**Maintenance**

Notes:

Specific Conductance Probe Cleaned? Y NDissolved Oxygen Membrane Changed? Y N

Site ID:

Project: SPRS1303Edgar Wade

Print name of person performing calibrations.

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ± 5%) (pH ± 0.2SU) (Temperature ±0.2°C) Warm bath Appx. 40°C

Br...

q1

SHANGRI-LA SPRINGS SIP

SPRS1303 SHANGRI-LA SPRINGS

GATOR HOLE SPRING

SPRS1303 GATOR HOLE SPRING (JACKSON)

HEIDI HOLE SPRING

SPRS1303 HEIDI HOLE SPRING

HOLE IN THE WALL SPRING

SPRS1303 HOLE-IN-THE-WALL SPRING

TWIN CAVES SPRING

SPRS1303 TWIN CAVES SPRING

RQ-2013-02-11-43

SPRS1303