



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 4/30/13 TIME 7:30

POST-SAMPLE CALIBRATION (circle one) ☐ IN-LAB ☒ ON-SITE ON-SITE LOCATION post-Silver Glen

DATE 4/30/13 TIME 19:10

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.55	22.72	99.2	(P)
CALIBRATION	8.63	22.67	100.0	(P)
PRE-VERIFICATION	8.63	22.64	100.0	(P)
POST-VERIFICATION	8.70	22.12	99.7	(P)

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	103	(P)
CALIBRATION	100	2301231	6/14	100	(P)
PRE-VERIFICATION	100	2301231	6/14	100	(P)
POST-VERIFICATION	100	2301231	6/14	101	(P)

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.0	115471	10/13	7.04	22.61	(P)
CALIBRATION	7.0	115471	10/13	7.00	22.59	(P)
PRE-VERIFICATION	7.0	115471	10/13	7.00	22.58	(P)
POST-VERIFICATION	7.0	115471	10/13	7.02	22.09	(P)

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDs

RQ-2013-04-29-35

SPRG1304



SILVER GLEN SPRINGS

SPRG1304 SILVER GLEN SPRINGS



QAQC BLANK 1

SPRG1304

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Paul Blair

Edgar Wade

Paul Blair