



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ☐ ON-SITE ☒ ON-SITE LOCATION pre - Silver Spring Area

DATE 5/1/13 TIME 7.45

POST-SAMPLE CALIBRATION (circle one) IN-LAB ☐ ON-SITE ☒ ON-SITE LOCATION post - Silver Spring

DATE _____ TIME _____

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	9.02	20.74	100.7	(P)
CALIBRATION	8.96	20.73	100.0	(P)
PRE-VERIFICATION	8.96	20.73	100.0	(P)
POST-VERIFICATION	9.03	20.52	100.1	(P)

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	99	(P)
CALIBRATION	100	2301231	6/14	100	(P)
PRE-VERIFICATION	100	2301231	6/14	100	(P)
POST-VERIFICATION	100	2301231	6/14	101	(P)

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.0	115471	10/13	7.09	21.31	(P)
CALIBRATION	7.0	115471	10/13	7.00	21.28	(P)
PRE-VERIFICATION	7.0	115471	10/13	7.00	21.28	(P)
POST-VERIFICATION	7.0	115471	10/13	7.03	20.68	(P)

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☐ NO ☒

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☐ NO ☒

RQ & SITE IDS

RQ-2013-04-29-37

SR-SW-Yearling House

SR-SW-1

RQ-2013-04-29-35

SPRG1304



RECEPTION HALL SPRING

SPRG1304 RECEPTION HALL SPRING



SILVER SPRING MAIN

SPRG1304 SILVER SPRING MAIN



SALT SPRINGS (MARION)

SPRG1304 SALT SPRINGS (MARION)

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Edgar Wade

Paul Blair

Paul Blair