



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one)

IN-LAB

ON-SITE

ON-SITE LOCATION

post Rock Spring

DATE 5/21/13

TIME 7:00

POST-SAMPLE CALIBRATION (circle one)

IN-LAB

ON-SITE

ON-SITE LOCATION

post Rock Spring (orange)

DATE 5/21/13

TIME 21:00

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.66	22.46	99.9	Pass
CALIBRATION	8.67	22.44	100.0	Pass
PRE-VERIFICATION	8.67	22.44	100.0	Pass
POST-VERIFICATION	8.60	22.71	99.7	Pass

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	101	Pass
CALIBRATION	100	2301231	6/14	100	Pass
PRE-VERIFICATION	100	2301231	6/14	100	Pass
POST-VERIFICATION	100	2301231	6/14	101	Pass

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	115471	10/13	6.97	22.55	Pass
CALIBRATION	7.00	115471	10/13	7.00	22.44	Pass
PRE-VERIFICATION	7.00	115471	10/13	7.00	22.43	Pass
POST-VERIFICATION	7.00	115471	10/13	6.98	22.74	Pass

Acceptance Criteria:

DO ± 0.3 mg/L

Sp. Cond. ± 5%

pH ± 0.2 SU

Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one)

YES

NO

Was the Dissolved Oxygen Membrane changed? (circle one)

YES

NO

RQ & SITE IDS

RQ-2013-05-20-34

SPRG1304



APOPKA SPRING

SPRG1304 APOPKA SPRING



WEKIWA SPRING (ORANGE)

SPRG1304 WEKIWA SPRING (ORANGE)



ROCK SPRINGS (ORANGE)

SPRG1304 ROCK SPRINGS (ORANGE)

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Edgar Wade

Kirstin Eller

Kirstin Eller