



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 4/12/13 TIME 4:45

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 4/13/13 TIME 8:15

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.41	22.70	103.2	P
CALIBRATION	8.42	22.70	100	P
PRE-VERIFICATION	8.62	22.70	100	P
POST-VERIFICATION	8.59	22.69	99.5	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	4/14	130	F
CALIBRATION	100	2301231	4/14	100	P
PRE-VERIFICATION	100	2301231	4/14	100	P
POST-VERIFICATION	100	2301231	4/14	102	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	115471	10/13	7.02	22.19	P
CALIBRATION	7	115471	10/13	7.00	22.19	P
PRE-VERIFICATION	7	115471	10/13	6.95	22.17	P
POST-VERIFICATION	7	115471	10/13	7.02	22.22	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ~~NO~~ ~~FE~~

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

RQ-2013-06-10-27

SPRG1304



MORRISON SPRING

SPRG1304 MORRISON SPRING



PONCE DE LEON SPRINGS (HOL)

SPRG1304 PONCE DE LEON SPRINGS (HOLMES)



GAINER SPRING #1C

SPRG1304 GAINER SPRING #1C



CYPRESS SPRING

SPRG1304 CYPRESS SPRING

SAMPLER NAMES & SIGNATURES (All team members present)