



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID #

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 8/15/13 TIME 8:30

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 8/15/13 TIME 14:30

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.38	23.66	98.9	P
CALIBRATION	8.47	23.66	100.0	P
PRE-VERIFICATION	8.46	23.66	100.0	P
POST-VERIFICATION	7.95	27.13	100.0	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	102	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	101	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	115471	10/13	7	23.52	P
CALIBRATION	7	115471	10/13	7	23.52	P
PRE-VERIFICATION	7	115471	10/13	7	23.52	P
POST-VERIFICATION	7	115471	10/13	6.96	23.31	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDS

RQ-2013-08-12-21

SPRG1307



BIG BLUE SPRING (JEFFERSON)

SPRG1307 BIG BLUE SPRING (JEFFERSON)



WACISSA SPRING #2

SPRG1307 WACISSA SPRING #2

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller