



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 9/26/13 TIME 7:30

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 9/26/13 TIME 17:20

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.67	22.29	99.8	Pass
CALIBRATION	8.69	22.29	100.0	pass
PRE-VERIFICATION	8.69	22.29	100.0	pass
POST-VERIFICATION	8.72	22.31	100.1	pass

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	101	pass
CALIBRATION	100	2301231	6/14	100	pass
PRE-VERIFICATION	100	2301231	6/14	100	pass
POST-VERIFICATION	100	2301231	6/14	101	pass

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	115471	10/13	6.96	22.35	pass
CALIBRATION	7.00	115471	10/13	7.00	22.34	pass
PRE-VERIFICATION	7.00	115471	10/13	7.00	22.34	pass
POST-VERIFICATION	7.00	115471	10/13	7.02	22.23	pass

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDS

RQ-2013-09-02-37

SPRG1307



MORRISON SPRING

SPRG1307 MORRISON SPRING



PONCE DE LEON SPRINGS (HC)

SPRG1307 PONCE DE LEON SPRINGS (HOLMES)



QAQC BLANK 1

SPRG1307

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wady
Edgar Wady