



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 11/6/13 TIME 8:15

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 11/6/13 TIME 16:00

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.77	21.86	100.2	P
CALIBRATION	8.77	21.84	100.0	P
PRE-VERIFICATION	8.77	21.83	100.0	P
POST-VERIFICATION	8.40	24.05	99.9	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	4/14	100	P
CALIBRATION	100	2301231	4/14	100	P
PRE-VERIFICATION	100	2301231	4/14	100	P
POST-VERIFICATION	100	2301231	4/14	100	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	121302	3/14	6.99	21.74	P
CALIBRATION	7	121302	3/14	7.00	21.74	P
PRE-VERIFICATION	7	121302	3/14	7.00	21.74	P
POST-VERIFICATION	7	121302	3/14	6.97	21.97	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. ± 5% pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

RQ-2013-11-04-92

SPRS1313



BLUE SPRING (MADISON)

SPRS1313 MADISON BLUE SPRING- STATUS NETWORK

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller