



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 11/19/13 TIME 7:30

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 11/20/13 TIME _____

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	9.42	23.52	111.2	F
CALIBRATION	8.47	23.50	100.0	P
PRE-VERIFICATION	8.47	23.46	99.9	P
POST-VERIFICATION	9.31	18.11	98.6	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	100	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	20	F

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	121302	3/14	7.33	24.51	F
CALIBRATION	7	121302	3/14	7.00	24.41	P
PRE-VERIFICATION	7	121302	3/14	7.01	24.32	P
POST-VERIFICATION	7	121302	3/14	7.10	22.37	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDS

RQ-2013-11-18-26

SPRG1310



ICHETUCKNEE HEAD SPRING

SPRG1310 ICHETUCKNEE HEAD SPRING



MILL POND SPRINGS (COLUMBIA)

SPRG1310 MILL POND SPRINGS (COLUMBIA)



MISSION SPRINGS

SPRG1310 MISSION SPRINGS




BLUE HOLE SPRING (COLUMBIA)

SPRG1310 BLUE HOLE SPRING (COLUMBIA)



DEVILS EAR SPRING (GILCHRIST)

SPRG1310 DEVILS EAR SPRING (GILCHRIST)



DEVIL'S EYE SPRING (GILCHRIST)

SPRG1310 DEVIL'S EYE SPRING (GILCHRIST)

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller
[Signature]