



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 12/3/13 TIME 8:00

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 12/4/13 TIME 7:45

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.80	20.41	98.3	P
CALIBRATION	9.02	20.39	100.0	P
PRE-VERIFICATION	9.04	20.22	100.0	P
POST-VERIFICATION	8.81	22.43	101.9	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	126	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	101	P
POST-VERIFICATION	100	2301231	6/14	100	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	121302	3/14	6.82	20.15	P
CALIBRATION	7	121302	3/14	7.00	20.15	P
PRE-VERIFICATION	7	121302	3/14	7.00	20.15	P
POST-VERIFICATION	7	121302	3/14	7.02	21.57	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDs

RQ-2013-12-02-83

SPRG1310



CYPRESS SPRING

SPRG1310 CYPRESS SPRING



GAINER SPRING #1C

SPRG1310 GAINER SPRING #1C

SAMPLER NAMES & SIGNATURES (All team members present)

K. Elmer