



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 10901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 12/11/13 TIME 7:45

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 12/12/13 TIME 8:30

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	9.19	79.02	100.4	P
CALIBRATION	9.16	79.01	100.0	P
PRE-VERIFICATION	9.17	79.39	100.0	P
POST-VERIFICATION	9.32	79.23	101.0	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	78	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	93	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	132908	4/15	6.97	18.76	P
CALIBRATION	7	132908	4/15	7.00	18.73	P
PRE-VERIFICATION	7	132908	4/15	7.00	18.73	P
POST-VERIFICATION	7	132908	4/15	7.13	19.86	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDS

RQ-2013-11-11-57	Big Blue Jefferson	WR-1
SPRS1314	Cassidy Spring	WR-2
RQ-2013-11-11-58	Wacissa #2	WR-3
	Log Spring	WR-4
	Thomas Spring	

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eiler
H. Eiler