



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

SPRINGS SAMPLING FIELD SHEET

STATION INFORMATION

STATION ID WAKULLA K TUNNEL DATE 2/5/14 TIME(EST) 11:00 (24HR)

SPRING NAME WAKULLA WEATHER: CLEAR PARTLY-CLOUDY CLOUDY RAIN

WATER QUALITY MEASUREMENTS

Analytes collected for this site are as listed per the container inventory, including any QA/QC samples ☒ YES ☐ NO

CHEMICAL STABILITY MONITORING

Time	Temperature (°C)	Conductivity (µs/cm)	DO (% sat)	DO (mg/l)	pH (su)
10:49	19.66	265	21.3	1.94	6.96
10:54	19.68	265	21.2	1.94	6.92
10:59	19.65	265	21.4	1.96	6.85

Stability Criteria: Temp $\pm 0.2^{\circ}\text{C}$ Sp Cond $\pm 5\%$ of reading pH ± 0.2 SU DO $\leq 20\%$ or 0.2 mg/l Turbidity $\leq 20\%$ or 5 NTUs

FINAL FIELD READING

Meter ID	Temperature (°C)	Conductivity (µs/cm)	D.O. (% sat)	D.O. (mg/l)	pH (su)
101901	19.64	265	21.5	1.97	6.85

SAMPLE PRESERVATION

Were all samples collected using the appropriate 0.45-micron filter (when applicable) & Preserved with the appropriate type and amount of acid (sulfuric for nutrients & nitric for metals)? ☒ YES ☐ NO

Nitric Acid Lot # 1112080 Sulfuric Acid Lot # 120591 Filter Lot # 7305004

Were all collected samples properly preserved and stored on wet ice at 4°C within 15 minutes? ☒ YES ☐ NO

RQ LABEL

RQ-2014-02-03-20
SPRG1401

SITE LABEL


WAKULLA K TUNNEL
SPRG1401 WAKULLA K TUNNEL

QA/QC Taken? YES ☒ NO

TIME(EST) _____ (24HR)

BLANK TYPE EQUIPMENT FIELD

PERISTALTIC PUMP ID # 2

QA/QC LABEL (if applicable)  ID)

QAQC BLANK 2
SPRG1401

COMMENTS

slight yellow color - no odor

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Gary Maddox