

Boldly "X" this box if there is qualified data on this page.

CONTINUING CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: YSI 600 XCM #11 Datalogger ID: YSI 600 #8 Sept 2008 version

Date 3/19/14 Pre Sample Time: 8:30 Post Sample Time: 15:25
All Times Are: ETZ or CTZ (circle one)

Temperature (Monthly) Please Use "Monthly Temperature Verification Log"

Date of Last Temperature Verification 3/13/14

Pass or Fail

☒ P ☐ F

Dissolved Oxygen

In Run Mode Pre Sample 9.95 mg/L @ Temp. 15.7 °C % DO 100.1 ☒ P ☐ F
nearest 100th nearest 10th nearest 10th

Post Sample 8.70 mg/L @ Temp. 22.0 °C % DO 99.5 ☒ P ☐ F
nearest 100th nearest 10th nearest 10th

Specific Conductance

Exp. Date Lot # Standard Initial Reading
In Run Mode Pre Sample 10/1 2304066 100 100 ☒ P ☐ F
µmhos/cm µmhos/cm

Post Sample 1/15 2307012 1000 1001 ☒ P ☐ F
µmhos/cm µmhos/cm

pH

Standard Exp. Date Lot # Initial Reading Temperature
In Run Mode Pre Sample 10 SU 2/15 130972 10.0 SU 15.7 °C ☒ P ☐ F
nearest 10th nearest 10th

Post Sample 4 SU 3/15 131596 4.0 SU 20.5 °C ☒ P ☐ F
nearest 10th nearest 10th

Notes:

Site ID:
Project:

MSF MW # 18 / #14 / #16 / #15 / #13 / #10
AAK 7015

Maintenance: Specific Conductance Probe Cleaned? ☒ Y ☐ N
Dissolved Oxygen Membrane Changed: ☒ Y ☐ N

M. A. Eickes
Print name of person performing calibrations.

Acceptance Criteria: (D.O. $\pm$ 0.3 mg/l of theoretical value) (Specific Conductance $\pm$ 5%) (pH $\pm$ 0.2 SU) (Temperature $\pm$ 0.5 °C)