



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 3/26/14 TIME 8:00

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 3/27/14 TIME 10:00

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	20.41	20.41	100.7	P
CALIBRATION	20.41	20.41	99.9	P
PRE-VERIFICATION	20.41	20.41	99.8	P
POST-VERIFICATION	20.41	21.48	91.3	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	111	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	94	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	135752	9/15	6.81	21.52	P
CALIBRATION	7.00	135752	9/15	7.00	21.51	P
PRE-VERIFICATION	7.00	135752	9/15	7.00	21.50	P
POST-VERIFICATION	7.00	135752	9/15	7.12	20.98	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO ☐

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO ☐

RQ & SITE IDs

RQ-2014-03-10-25

SPRS1403



WACISSA SPRING #2

SPRG1401 WACISSA SPRING #2

QAQC Blank 1



BIG BLUE SPRING (JEFFERSON)

SPRG1401 BIG BLUE SPRING (JEFFERSON)

CASSIDY SPRING

LOG SPRING

THOMAS SPRING

RQ-2014-03-10-26

WR-SW-1

WR-SW-2

WR-SW-3

WR-SW-4

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Kirstin Eller

Kristina Bridger