



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 4/17/14 TIME 7:30

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 4/17/14 TIME 18:15

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	9.06	19.55	98.8	P
CALIBRATION	9.17	19.57	100.0	P
PRE-VERIFICATION	9.17	19.57	100.0	P
POST-VERIFICATION	8.96	20.69	99.9	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	102	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	101	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	135752	9/15	7.06	21.64	P
CALIBRATION	7.00	135752	9/15	7.00	21.61	P
PRE-VERIFICATION	7.00	135752	9/15	7.00	21.61	P
POST-VERIFICATION	7.00	135752	9/15	7.04	20.75	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDS

RQ-2014-03-10-29

SHALLOW 1	DEEP 1	WELL #1
SHALLOW 2	DEEP 2	
SHALLOW 3	DEEP 4	TRAP SAMPLE
SHALLOW 4	DEEP 5	
SHALLOW 6	DEEP 6	

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Gary Maddy