

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 12 YSI 600 XLM

September 2008 GENERAL version

Date 6-17-14 Calibration Time: 7:07 Project Name: Santa Fe Study
22LR1405

All Times Are: ETZ or CTZ (circle one)

Circle/Fill In: Lab Calibration or Calibration on Site

Temperature (Monthly)

Date of Last Temperature Verification 5-20-14

Ice Bath: Probe 12.6 °C NIST 12.7 °C Warm Bath: Probe 42.0 °C NIST 42.1 °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 8.16 mg/L @ Temp. 23.0 °C % DO 95.0 P F

Calibrated Reading: 8.58 mg/L @ Temp. 23.0 °C % DO 100.0 P F

In Run Mode

Initial Calib. Verification: 8.66 mg/L @ Temp. 22.9 °C % DO 101.3 P F

Specific Conductance

In Calibrate Mode

Calibration:	Exp. Date	Lot #	Standard	Initial Reading	Calibrated Reading
<u>3/15</u>	<u>2310714</u>	<u>1000</u>	<u>1002</u>	<u>1000</u>	<u>P</u> F
		µmhos/cm	µmhos/cm	µmhos/cm	

In Run Mode

Initial Calib. Verification:	Exp. Date	Lot #	Standard	Initial Reading	Calibrated Reading
<u>8/14</u>	<u>2304388</u>	<u>100</u>	<u>100</u>	<u>100</u>	<u>P</u> F
		µmhos/cm	µmhos/cm	µmhos/cm	

pH

Standard	Exp. Date	Lot #	Initial Reading	Temperature	Calibrated Reading	Temperature
----------	-----------	-------	-----------------	-------------	--------------------	-------------

In Calibrate Mode

7.0 SU	<u>1/16</u>	<u>140018</u>	<u>7.0</u>	SU <u>22.2</u> °C	<u>7.0</u>	SU <u>22.2</u> °C <u>P</u> F
--------	-------------	---------------	------------	-------------------	------------	------------------------------

4.0 SU	<u>3/15</u>	<u>131596</u>	<u>4.0</u>	SU <u>22.3</u> °C	<u>4.0</u>	SU <u>22.3</u> °C <u>P</u> F
--------	-------------	---------------	------------	-------------------	------------	------------------------------

10.0 SU	<u>12/15</u>	<u>138049</u>	<u>10.0</u> <u>10.1</u>	SU <u>22.2</u> °C	<u>10.0</u>	SU <u>22.2</u> °C <u>P</u> F
---------	--------------	---------------	----------------------------	-------------------	-------------	------------------------------

In Run Mode

Initial Calib. Verification: 7 SU 1/16 140018 7.0 SU 22.2 °C P F

Maintenance: Weekly PH Slope: _____ Specific Conductance Probe Cleaned? Y N

Notes: Dissolved Oxygen Membrane Changed? Y N

22-LR-8021

Site ID:



QA/QC BLANK SW 3

22LR1405

Chris Green

Print name of person performing calibrations.

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ± 5%) (pH ± 0.2SU) (Temperature ±0.5°C) Warm bath Appx. 40°C
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places

Boldly "X" this box if there is qualified data on this page.

CONTINUING CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: 12 YSI 600 XLm

September 2008 GENERAL version

Date 6-17-14 Mid-Day Time: _____ End-of-Day Time: 18:39

All Times Are: ETZ or CTZ (circle one)

Temperature (Monthly)

Date of Last Temperature Verification 5-20-14

Ice Bath: Probe 12.6 °C NIST 12.7 °C Warm Bath: Probe 42.0 °C NIST 42.1 °C

Pass or Fail

P F

Dissolved Oxygen In Run Mode

Mid-Day _____ mg/L @ Temp. _____ °C % DO _____

P F

End-of-Day 7.96 mg/L @ Temp. 98.9 °C % DO 26.1

P F

Specific Conductance In Run Mode

Exp. Date Lot # Standard Reading

Mid-Day _____ _____ _____ _____
µmhos/cm µmhos/cm

P F

End-of-Day 11/14 1306509 500 501
µmhos/cm µmhos/cm

P F

pH In Run Mode

Bracketing Standard Exp. Date Lot # Reading Temperature

Mid-Day _____ SU _____ _____ SU _____ °C

P F

End-of-Day 4 SU 3/15 131596 4.0 SU 24.6 °C
10 2/15 130972 10.0 24.5

P F

P

Notes:

22-LR-8021

Site ID:



QA/QC BLANK SW 3

22LR1405

Maintenance: Specific Conductance Probe Cleaned? Y N
Dissolved Oxygen Membrane Changed? Y N

Chris Green

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.5 °C) Warm bath Appx. 40 °C
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places