

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Datalogger ID: UCI 658 #8

Acceptance Criteria: ($DO \pm 0.3 \text{ mg/l}$ of theoretical value) (Specific Conductance $\pm 5\%$) ($pH \pm 0.2 \text{ SU}$) (Temperature $\pm 0.5^\circ \text{C}$)