

CONTINUING CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 124SI600XLM

September 2008 GENERAL version

Date 10-28-14 Mid-Day Time: _____ End-of-Day Time: 19:08

All Times Are: (ETZ) or CTZ (circle one)

Temperature (Monthly)

Date of Last Temperature Verification 10-01-14

Ice Bath: Probe 13.3 °C NIST 13.4 °C Warm Bath: Probe 37.6 °C NIST 37.7 °C

Pass or Fail
(P) F

Dissolved Oxygen In Run Mode

Mid-Day _____ mg/L @ Temp. _____ °C % DO _____ P F

End-of-Day 9.04 mg/L @ Temp. 20.58 °C % DO 100.5 (P) F
9.04

Specific Conductance In Run Mode

Exp. Date Lot # Standard Reading

Mid-Day _____ _____ _____ μmhos/cm _____ μmhos/cm P F

End-of-Day 4/15 140402A 500 501 (P) F
μmhos/cm μmhos/cm

pH In Run Mode

Bracketing Standard Exp. Date Lot # Reading Temperature

Mid-Day _____ SU _____ _____ _____ SU _____ °C P F

End-of-Day 10 SU 12/15 138049 10.01 SU 20.88 °C (P) F

Notes:

Site ID:
Project:

Maintenance:

Specific Conductance Probe Cleaned?

Y (N)

Dissolved Oxygen Membrane Changed:

Y (N)

Chris Green

Print name of person performing calibrations.

Acceptance Criteria: (DO ± 0.3 mg/l) (Specific Conductance $\pm 5\%$) (pH ± 0.2 SU) (Temperature ± 0.5 °C) Warm bath Appx. 40 °C
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 12 YSI 600 XLIM

September 2008 GENERAL version

Date 10-28-14 Calibration Time: 7:44 Project Name: Middle Santa Fe Restoration Project

All Times Are: ETZ or CTZ (circle one)

Circle/Fill In: Lab Calibration or Calibration on Site _____

Temperature (Monthly) Date of Last Temperature Verification 10-1-14

Ice Bath: Probe 13.3 °C NIST 13.4 °C Warm Bath: Probe 37.6 °C NIST 37.7 °C

Pass or Fail
P F

Dissolved Oxygen

In Calibrate Mode

Initial Reading: 9.83 mg/L @ Temp. 20.49 °C % DO 109.1 P F

Calibrated Reading: 9.02 mg/L @ Temp. 20.41 °C % DO 100.0 P F

In Run Mode

Initial Calib. Verification: 9.26 mg/L @ Temp. 19.07 °C % DO 100.0 P F

Specific Conductance

In Calibrate Mode

Calibration:	Exp. Date	Lot #	Standard	Initial Reading	Calibrated Reading
	<u>9/15</u>	<u>2402978</u>	<u>1000</u> µmhos/cm	<u>1001</u> µmhos/cm	<u>1000</u> µmhos/cm

In Run Mode

Initial Calib. Verification:	<u>12/14</u>	<u>2307729</u>	<u>100</u> µmhos/cm	<u>100</u> µmhos/cm	
------------------------------	--------------	----------------	------------------------	------------------------	--

pH

Standard	Exp. Date	Lot #	Initial Reading	Temperature	Calibrated Reading	Temperature
----------	-----------	-------	-----------------	-------------	--------------------	-------------

In Calibrate Mode

7.0 SU	<u>1/16</u>	<u>140018</u>	<u>7.00</u> SU	<u>20.57</u> °C	<u>7.00</u> SU	<u>20.50</u> °C
--------	-------------	---------------	----------------	-----------------	----------------	-----------------

4.0 SU	<u>3/15</u>	<u>131596</u>	<u>3.99</u> SU	<u>20.27</u> °C	<u>4.00</u> SU	<u>20.26</u> °C
--------	-------------	---------------	----------------	-----------------	----------------	-----------------

10.0 SU	<u>12/15</u>	<u>138049</u>	<u>10.04</u> SU	<u>20.35</u> °C	<u>10.01</u> SU	<u>20.26</u> °C
---------	--------------	---------------	-----------------	-----------------	-----------------	-----------------

In Run Mode

Initial Calib. Verification: 7 SU 1/16 140018 7.00 SU 20.47 °C P F

Maintenance: Weekly PH Slope: _____ Specific Conductance Probe Cleaned? Y N

Notes: Dissolved Oxygen Membrane Changed? Y N

7 - -25.0 pHmV
4 - 145.6 pHmV
10 - 798.3 pHmV

Site ID:
Project:

Chris Green

Print name of person performing calibrations.

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ± 5%) (pH ± 0.2SU) (Temperature ±0.5°C) Warm bath Appx. 40°C
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places