



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 148287

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 11/13/14 TIME 8:00

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 11/14/14 TIME 8:00

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.72	22.716	101.1	P
CALIBRATION	8.62	22.661	99.9	P
PRE-VERIFICATION	8.62	22.659	99.9	P
POST-VERIFICATION	8.94	21.003	100.2	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	1000	2405350	10/15	999.7	P
CALIBRATION	1000	2405350	10/15	999.9	P
PRE-VERIFICATION	1000	2405350	10/15	999.9	P
POST-VERIFICATION	100	2307731	10/15	100.4	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	10.00	135751	9/15	10.03	22.843	P
CALIBRATION	10.00			10.00	22.830	P
PRE-VERIFICATION	10.00			10.00	22.826	P
POST-VERIFICATION	4.00	135444	10/15	4.04	21.358	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDS



GAINER SPRING #1C

SPRG1410 GAINER SPRING #1C

RQ-2014-11-10-24

SPRG1410



PONCE DE LEON SPRINGS (H0)

SPRG1410 PONCE DE LEON SPRINGS (H0)

RQ-2014-11-10-25

SPRG1410

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller
K. Eller

Edgar Wadde
Edgar Wadde