

Andrew Tintle
State of Florida DEP - Innov Park Lab
2051 East Paul Dirac Dr./Collins Bldg
Tallahassee, FL 32310

SOIL-67



Final Report

Soil/Hazardous Waste Proficiency Testing

Soil Study

Open Date: 07/20/09

Close Date: 09/03/09

Report Issued Date: 09/24/09

September 24, 2009

Andrew Tintle
State of Florida DEP - Innov Park Lab
2051 East Paul Dirac Dr./Collins Bldg
Tallahassee, FL 32310

Enclosed is your final report for ERA's SOIL-67 Proficiency Testing (PT) study. Your final report includes an evaluation of all results submitted by your laboratory to ERA.

Data Evaluation Protocols: All analytes in ERA's SOIL-67 Proficiency Testing (PT) study have been evaluated using the following tiered approach. If the analyte is listed in the most current National Environmental Laboratory Accreditation Conference (NELAC) PT Field of Testing tables, the evaluation was completed by comparing the reported result to the acceptance limits generated using the criteria contained in the NELAC FoPT tables. If the analyte is not included in the NELAC FoPT tables, the reported result has been evaluated using the procedures outlined in ERA's Standard Operating Procedure for the Generation of Performance Acceptance Limits (SOP 0260).


Corrective Action Help: As part of your accreditation(s), you may be required to identify the root cause of any "Not Acceptable" results, implement the necessary corrective actions, and then satisfy your PT requirements by participating in a Supplemental (Quik™ Response) or future ERA PT study. ERA's technical staff is available to help your laboratory resolve any technical issues that may be impairing your PT performance and possibly affecting your routine data quality. Our laboratory and technical staff have well over three hundred years of collective experience in performing the full range of environmental analyses. As part of our technical support, ERA offers QC samples that can be helpful in helping you work through your technical issues.

Thank you for your participation in ERA's SOIL-67 Proficiency Testing study. If you have any questions, please contact Shawn Kassner, Proficiency Testing Manager, or Curtis Wood, Director of Regulatory Affairs and Business Development, at 1-800-372-0122.

Sincerely,



Shawn Kassner
Proficiency Testing Manager



Jay R. McBurney
Quality Program Manager

attachments
smk

Report Recipient	Contact/Phone Number	Reporting Type
Florida	Steve Arms / 904-791-1502	All Analytes

SOIL-67 Definitions & Study Discussion

Study Dates: 07/20/09 - 09/03/09

Report Issued: 09/24/09

SOIL Study Definitions

The Reported Value is the value that the laboratory reported to ERA.

The ERA assigned value for the Organic Proficiency Testing Standards is equal to 100% of the parameter present in the standard as determined by gravimetric and/or volumetric measurements made during standard preparation as applicable. The ERA assigned value for the Inorganic Proficiency Testing Standards, with the exception of the TCLP Metals in Soil, is equal to the maximum amount of the parameter available in the standard by applicable EPA methodologies. The ERA assigned value for the TCLP metals is equal to the mean of ERA's internal analytical analyses. All NELAC parameters not added to a standard are given an assigned Value of "0", per the guidance issued by the NELAC Board of Directors, on December 14, 2000. Non-NELAC parameters not added to a standard may be given an assigned value of less than a minimum verified concentration as determined in the background soil for applicable EPA methodologies.

The Acceptance Limits are established per the NELAC PT program criteria or ERA's SOP for the Generation of Performance Acceptance Limits™ as applicable.

The Performance Evaluation:

Acceptable	= Reported Value falls within the Acceptance Limits.
Not Acceptable	= Reported Value falls outside the Acceptance Limits.
No Evaluation	= Reported Value cannot be evaluated.
Not Reported	= No Value reported.

The Method Description is the method the laboratory reported to ERA.

SOIL Study Discussion

ERA's SOIL-67 Proficiency Testing (PT) study has been reviewed by ERA senior management and certified compliant with the requirements of the National Environmental Laboratory Accreditation Conference (NELAC), Proficiency Testing Program Standards, Chapter 2, July 2003.

Per the requirements of the NELAC Proficiency Testing Program, a full review of all homogeneity, stability, and accuracy verification data was completed. All analytical verification data for all analytes in the Soil study standards met the acceptance criteria contained in the NELAC Proficiency Testing Program Standards, Chapter 2, July 2003. If the analyte is included in the NELAC Fields of Testing list the acceptance limits were calculated based on the NELAC Proficiency Testing Program Standards, Chapter 2, July 2003. If the analyte is not included in the NELAC Fields of Testing list, the acceptance limits were calculated using the procedures outlined in ERA's Standard Operating Procedure for the Generation of Performance Acceptance Limits (SOP 0260, Rev. 2.0).

The data submitted by participating laboratories was also examined for study anomalies. There were no anomalies observed during the statistical review of the data.

ERA's SOIL-67 Proficiency Testing study reports shall not be reproduced except in its entirety and not without the permission of the participating laboratory. The report must not be used by the participating laboratories to claim product endorsement any agency of the U. S. government.

The data contained herein are confidential and intended for your use only.

If you have any questions or concerns regarding your assessment in ERA's SOIL Proficiency Testing program, please contact Shawn Kassner, Proficiency Testing Manager, or Curtis Wood, Director of Regulatory Affairs and Business Development, at 1-800-372-0122.



Study: **SOIL-67**

ERA Customer Number: **S744810**

Laboratory Name: **State of Florida DEP -
Innov Park Lab**

Organic Results



SOIL-67 Final Complete Report

Andrew Tintle
Quality Assurance Officer
State of Florida DEP - Innov Park Lab
2051 East Paul Dirac Dr./Collins Bldg
Tallahassee, FL 32310
850-245-8075

EPA ID: FL01182
ERA Customer Number: S744810
Report Issued: 09/24/09
Study Dates: 07/20/09 - 09/03/09

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

SOIL Volatiles in Soil (cat# 623)

4315	Acetone	µg/kg	93.0	309	40.0 - 617	Acceptable	EPA 8260C
4320	Acetonitrile	µg/kg		796	222 - 1070	Not Reported	
4325	Acrolein	µg/kg		0.00		Not Reported	
4375	Benzene	µg/kg	75.1	75.5	44.4 - 104	Acceptable	EPA 8260C
4385	Bromobenzene	µg/kg		0.00		Not Reported	
4395	Bromodichloromethane	µg/kg	154	158	104 - 217	Acceptable	EPA 8260C
4400	Bromoform	µg/kg	84.0	95.7	48.2 - 142	Acceptable	EPA 8260C
4950	Bromomethane	µg/kg	< 10.0	0.00		Acceptable	EPA 8260C
4410	2-Butanone (MEK)	µg/kg	< 20.0	0.00		Acceptable	EPA 8260C
5000	tert-Butyl methyl ether (MTBE)	µg/kg	213	193	81.7 - 281	Acceptable	EPA 8260C
4450	Carbon disulfide	µg/kg		0.00		Not Reported	
4455	Carbon tetrachloride	µg/kg	74.9	84.4	41.2 - 125	Acceptable	EPA 8260C
4475	Chlorobenzene	µg/kg	30.2	30.2	15.8 - 41.3	Acceptable	EPA 8260C
4575	Chlorodibromomethane	µg/kg	< 5.00	0.00		Acceptable	EPA 8260C
4485	Chloroethane	µg/kg	< 10.0	0.00		Acceptable	EPA 8260C
4500	2-Chloroethylvinylether	µg/kg		0.00		Not Reported	
4505	Chloroform	µg/kg	131	133	80.7 - 185	Acceptable	EPA 8260C
4960	Chloromethane	µg/kg	< 10.0	0.00		Acceptable	EPA 8260C
4570	1,2-Dibromo-3-chloropropane (DBCP)	µg/kg		83.9	41.0 - 136	Not Reported	
4585	1,2-Dibromoethane (EDB)	µg/kg		159	109 - 222	Not Reported	
4595	Dibromomethane	µg/kg		0.00		Not Reported	
4610	1,2-Dichlorobenzene	µg/kg	< 5.00	0.00		Acceptable	EPA 8260C
4615	1,3-Dichlorobenzene	µg/kg	< 5.00	0.00		Acceptable	EPA 8260C
4620	1,4-Dichlorobenzene	µg/kg	< 5.00	0.00		Acceptable	EPA 8260C
4625	Dichlorodifluoromethane (Freon 12)	µg/kg	< 10.0	0.00		Acceptable	EPA 8260C
4630	1,1-Dichloroethane	µg/kg	172	168	95.7 - 242	Acceptable	EPA 8260C
4635	1,2-Dichloroethane	µg/kg	173	182	105 - 251	Acceptable	EPA 8260C
4640	1,1-Dichloroethylene	µg/kg	194	196	112 - 320	Acceptable	EPA 8260C
4645	cis-1,2-Dichloroethylene	µg/kg	63.2	57.4	36.0 - 82.3	Acceptable	EPA 8260C
4700	trans-1,2-Dichloroethylene	µg/kg	144	133	94.4 - 185	Acceptable	EPA 8260C
4655	1,2-Dichloropropane	µg/kg	167	163	95.2 - 218	Acceptable	EPA 8260C
4680	cis-1,3-Dichloropropylene	µg/kg	95.2	102	67.7 - 144	Acceptable	EPA 8260C



SOIL-67 Final Complete Report

Andrew Tintle
Quality Assurance Officer
State of Florida DEP - Innov Park Lab
2051 East Paul Dirac Dr./Collins Bldg
Tallahassee, FL 32310
850-245-8075

EPA ID: FL01182
ERA Customer Number: S744810
Report Issued: 09/24/09
Study Dates: 07/20/09 - 09/03/09

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
SOIL Volatiles in Soil (cat# 623) (Continued)							
4685	trans-1,3-Dichloropropylene	µg/kg	177	193	136 - 263	Acceptable	EPA 8260C
4765	Ethylbenzene	µg/kg	144	146	81.7 - 208	Acceptable	EPA 8260C
4860	2-Hexanone	µg/kg	< 10.0	0.00		Acceptable	EPA 8260C
4900	Isopropylbenzene	µg/kg		152	91.8 - 234	Not Reported	
4975	Methylene chloride	µg/kg	62.2	61.4	28.4 - 91.2	Acceptable	EPA 8260C
4995	4-Methyl-2-pentanone (MIBK)	µg/kg	61.9	103	42.6 - 154	Acceptable	EPA 8260C
5005	Naphthalene	µg/kg		82.2	28.1 - 134	Not Reported	
5100	Styrene	µg/kg	106	102	69.5 - 143	Acceptable	EPA 8260C
5105	1,1,1,2-Tetrachloroethane	µg/kg		0.00		Not Reported	
5110	1,1,2,2-Tetrachloroethane	µg/kg	81.4	95.7	49.2 - 140	Acceptable	EPA 8260C
5115	Tetrachloroethylene	µg/kg	< 5.00	0.00		Acceptable	EPA 8260C
5140	Toluene	µg/kg	126	121	69.6 - 170	Acceptable	EPA 8260C
5155	1,2,4-Trichlorobenzene	µg/kg		0.00		Not Reported	
5160	1,1,1-Trichloroethane	µg/kg	156	169	97.0 - 242	Acceptable	EPA 8260C
5165	1,1,2-Trichloroethane	µg/kg	121	119	69.8 - 164	Acceptable	EPA 8260C
5170	Trichloroethylene	µg/kg	192	185	95.4 - 264	Acceptable	EPA 8260C
5175	Trichlorofluoromethane	µg/kg	178	162	78.5 - 278	Acceptable	EPA 8260C
5180	1,2,3-Trichloropropane (TCP)	µg/kg		0.00		Not Reported	
5225	Vinyl acetate	µg/kg		0.00		Not Reported	
5235	Vinyl chloride	µg/kg	< 10.0	0.00		Acceptable	EPA 8260C
5260	Xylenes, total	µg/kg	234	257	127 - 377	Acceptable	EPA 8260C

