

Chemical Analysis Report

WQAP-2018-07-06-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Algal Bloom Response - As needed**
Request ID: **RQ-2018-07-02-45**
Customer: **WQAP**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
2295 Victoria AVE
Suite 364
Fort Myers, FL 33901
Attn: Kirby Wolfe

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 09-JUL-2018 13:29

A handwritten signature in black ink, appearing to read "Colin Wright", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Manuel Dr at Caloosahatchee

Collection Date/Time: 07/05/2018 11:40

Field ID: SD070518-001

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2004997	EPA 8321B	Microcystin LR**	3.5		ug/L	P348010	
		Microcystin LA**	0.25	U	ug/L	P348010	
		Microcystin RR**	0.25	U	ug/L	P348010	
		Microcystin YR**	0.25	U	ug/L	P348010	
		Anatoxin-a**	0.25	U	ug/L	P348010	
		Cylindrospermopsin**	0.25	U	ug/L	P348010	
		Microcystin LF**	0.25	U	ug/L	P348010	
		Microcystin LY**	0.25	U	ug/L	P348010	

Sample Location: SE 20th PL Cape Coral

Collection Date/Time: 07/05/2018 13:10

Field ID: SD070518-002

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2004998	EPA 8321B	Microcystin LR**	82		ug/L	P348010	
		Microcystin LA**	0.25	U	ug/L	P348010	
		Microcystin RR**	0.25	U	ug/L	P348010	
		Microcystin YR**	0.25	U	ug/L	P348010	
		Anatoxin-a**	0.25	U	ug/L	P348010	
		Cylindrospermopsin**	0.25	U	ug/L	P348010	
		Microcystin LF**	0.25	U	ug/L	P348010	
		Microcystin LY**	0.25	U	ug/L	P348010	

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B
Batch ID: P348010

Component	Result	Code	Units
Anatoxin-a	0.25	U	ug/L
Cylindrospermopsin	0.25	U	ug/L
Microcystin LA	0.25	U	ug/L
Microcystin LF	0.25	U	ug/L
Microcystin LR	0.25	U	ug/L
Microcystin LY	0.25	U	ug/L
Microcystin RR	0.25	U	ug/L
Microcystin YR	0.25	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P348010

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Anatoxin-a	109		P	40 - 150
Cylindrospermopsin	101		P	50 - 160
Microcystin LA	105		P	40 - 160
Microcystin LF	108		P	40 - 150
Microcystin LR	109		P	40 - 160
Microcystin LY	98.0		P	40 - 150
Microcystin RR	101		P	40 - 160
Microcystin YR	110		P	50 - 160

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P348010

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2004953	Anatoxin-a	88.0	95.5	P/P	40 - 150
2004953	Cylindrospermopsin	90.9	103	P/P	50 - 160
2004953	Microcystin LA	101	115	P/P	40 - 160
2004953	Microcystin LF	94.5	103	P/P	40 - 150
2004953	Microcystin LR	100	111	P/P	40 - 160
2004953	Microcystin LY	102	109	P/P	40 - 150
2004953	Microcystin RR	99.7	107	P/P	40 - 160
2004953	Microcystin YR	106	114	P/P	50 - 160

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P348010

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2004953	Anatoxin-a	8.21	Spike	P	0 - 30
2004953	Cylindrospermopsin	12.9	Spike	P	0 - 30
2004953	Microcystin LA	13.1	Spike	P	0 - 30
2004953	Microcystin LF	8.88	Spike	P	0 - 30
2004953	Microcystin LR	8.62	Spike	P	0 - 30
2004953	Microcystin LY	6.98	Spike	P	0 - 30
2004953	Microcystin RR	6.82	Spike	P	0 - 30
2004953	Microcystin YR	7.55	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 2004997
Field Sample ID: SD070518-001

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbamazepine-d10	99.3	P	30 - 160

Lab Sample ID: 2004998
Field Sample ID: SD070518-002

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbamazepine-d10	92.8	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B

Run ID: A85071

Included Lab Sample IDs: 2004997, 2004998

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Anatoxin-a	104	94.1	P/P	50 - 160
Anatoxin-a	94.1	88.0	P/P	50 - 160
Cylindrospermopsin	89.3	88.1	P/P	60 - 160
Cylindrospermopsin	92.7	89.3	P/P	60 - 160
Microcystin LA	104	104	P/P	50 - 160
Microcystin LA	95.5	104	P/P	50 - 160
Microcystin LF	91.8	92.3	P/P	60 - 160
Microcystin LF	94.6	91.8	P/P	60 - 160
Microcystin LR	97.4	98.9	P/P	60 - 160
Microcystin LR	98.9	99.3	P/P	60 - 160
Microcystin LY	96.0	99.9	P/P	60 - 160
Microcystin LY	98.1	96.0	P/P	60 - 160
Microcystin RR	89.4	90.9	P/P	60 - 160
Microcystin RR	92.8	89.4	P/P	60 - 160
Microcystin YR	92.6	99.7	P/P	60 - 160
Microcystin YR	97.7	92.6	P/P	60 - 160

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision SMP	MS
			LCS			
EPA 8321B	Anatoxin-a	109	88.0	95.5		8.21
	Cylindrospermopsin	101	90.9	103		12.9
	Microcystin LA	105	101	115		13.1
	Microcystin LF	108	94.5	103		8.88
	Microcystin LR	109	100	111		8.62
	Microcystin LY	98.0	102	109		6.98
	Microcystin RR	101	99.7	107		6.82
	Microcystin YR	110	106	114		7.55

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 8321B / E31780	Microcystins in water matrices by HPLC/MS/MS	2004997, 2004998

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 8321B	07/06/2018	07/06/2018 12:40	Juyoung Kim	07/06/2018 14:26	Juyoung Kim	2004998
	07/06/2018	07/06/2018 12:40	Juyoung Kim	07/06/2018 17:13	Juyoung Kim	2004997
	07/06/2018	07/06/2018 12:40	Juyoung Kim	07/06/2018 20:16	Juyoung Kim	2004998