

CYANOBACTERIA BLOOM RESPONSE FIELD DATA SHEET

(7-6-06 V2)

WATERBODY NAME: HICKEYS CREEK		DATE (M/D/Y): 9/9/19	TIME: (AM) (PM) 1010	Case ID#: SD09092019-001
LOCATION: HICKEYS CREEK @ 80		COUNTY: LEE	NEAREST TOWN/CITY: AWA	
Visit Type: Initial ☒ Recon ☒ Sampling ☒ Follow-up ☐	Samples collected: ☒ yes ☐ no	Samples submitted to: FDEP LAB		Photos Taken: ☒ yes ☐ no
Sample collection method: Single Grab: ☒ Grab Composite: ☐ Depth integrated composite: ☐				
Sample & Analysis types: Aqueous: ☒ Surface Scum: ☐ Benthic mat: ☐ Tissue: ☐				
Algal ID & enumeration: ☒ Nutrients: ☒ Chl-a: ☒ AGP: ☒ Turbidity: ☒ Toxin: ☒ <u>Microcystin</u> Specify				
Estimated size of bloom: N/A	Impounded: ☒ yes ☐ no	Stream/River/Canal 18 m wide		
Lake/Estuary Typical depth 1 m deep	Inflows: ☐ yes ☐ no	Typical Transect 1 m deep 2 m deep 1 m deep		
Outflows: ☐ yes ☐ no	Water velocity 0.2 m/s 0.2 m/s 0.2 m/s Left bank Middle Right bank			

Canopy Cover: Open: ☐ Lightly Shaded (11-45%): ☒ Moderately Shaded (46-80%): ☐ Heavily Shaded: ☐

Description of Bloom/ Latitude **26° 42' 53.202"** Longitude **-81° 40' 21.203"**
Degrees Minutes Seconds Degrees Minutes Seconds

Present Weather conditions

NO APPARENT BLOOM AT TIME OF VISIT.

Last rain event

Print a map from GIS tool or Internet mapping service (e.g., Mapquest, Google Map)

Rainfall amount

Sketch bloom area on the attached map. Indicate wind direction.

WATER QUALITY	Depth (m):	Temp. (°C):	pH (SU):	D.O. (mg/l):	Cond. (µmho/cm) or Salinity (ppt):	D.O. %	TD(m)	Secchi (m)
Top	0.3	30.5	7.52	5.34	635	71.3	1.3m	1.3 "S"
Mid-depth								
Bottom								
System Type: Stream: ☒ (1st - 2nd order 3rd - 4th order 5th - 6th order 7th order or greater) Lake: ☐ Wetland: ☐ Estuary: ☐ Other: ☐								
Water Odors (check box): Normal: ☒ Sewage: ☐ Petroleum: ☐ Chemical: ☐ Other: ☐								
Water Surface (check box): Clear: ☒ Scum: ☐ Globbs: ☐ Slick: ☐								
Water Clarity (check box): Clear: ☐ Slightly: ☒ Turbid: ☐ Opaque: ☐								
Bloom Color (check box): Green: ☐ Red: ☐ Brown: ☐ Other: ☐ N/A								
Notes					Abundance:			
					Absent	Rare	Common	Abundant
					☒	☐	☐	☐
					☐	☐	☒	☐
					☐	☐	☒	☐
					☒	☐	☐	☐

VISIT TEAM:

JOHN BARRINGTON CHRIST COWLEY FDEP SOURCE	AGENCY:
--	-------------

SAL (ppt)
0.3

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Pro move

RQ: Algae / L-2202

Project: Trond 5065900

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/18/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Chase Conley	9/9/19	0740	25.5	757.4	8.2	8.17	99.7	—	1.03	P/F	L/F
ICV	Chase Conley	9/9/19	0742	25.4	757.3	8.2	8.18	99.6	—	—	P/F	L/F
CCV	BARRINGTON	9/9/19	1225	25.8	758	8.1	7.87	96.8	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Chase Conley	9/9/19	0745	2907F48	7/21	500	502	P/F	L/F
ICV	Chase Conley	9/9/19	0747	2904D82	4/21	100	99	P/F	L/F
CCV	BARRINGTON	9/9/19	1227	2907D40	7/21	1000	1001	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Chase Conley	9/9/19	0750	191304	3/21	7.0	26.7	7.0	-64	P/F	L/F
Calibr.	Chase Conley	9/9/19	0752	191303	3/21	4.01	25.9	4.0	111.5	P/F	L/F
Calibr.	Chase Conley	9/9/19	0755	191735	4/21	10.03	25.9	9.99	245	P/F	L/F
ICV	Chase Conley	9/9/19	0759	191303	3/21	4.01	25.9	4.0	—	P/F	L/F
CCV	BARRINGTON	9/9/19	1230	191735	4/21	10.02	24.1	10.14	—	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 181, slope from 4 to 7 175.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **NO** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Deep Blue; Date of last verification: 6/18/19

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: