

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Dosi

RQ: 2021-10-11-60

Project: Algae Bloom Response

2021-09-20-184

MacLassie SLI+

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 11/10/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	12/28/21	804	17.5	762.1	9.6	—	100.3	—	1.062	P/F	L/F
ICV	↓	↓	806	17.5	762.2	9.6	9.60	100.4	—	1.062	P/F	L/F
CCV	↓	↓	1619	21.8	760.1	8.8	8.74	99.5	—	1.062	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	12/28/21	808	210917B	9/2022	1000	1000	P/F	L/F
ICV	↓	↓	810	210426A	5/2022	100	103.0	P/F	L/F
CCV	↓	↓	1621	211122B	12/2022	1413	1417	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	12/28/21	813	2109M25	9/2023	7.03	18.0	7.03	-28.7	P/F	L/F
Calibr.	↓	↓	815	2108FS1	8/2023	4.00	18.0	4.00	152.2	P/F	L/F
Calibr.	↓	↓	816	2108H37	2/2023	10.07	18.1	10.07	-195.6	P/F	L/F
ICV	↓	↓	819	2109M25	9/2023	7.03	18.0	7.06	-24.7	P/F	L/F
CCV	↓	↓	1622	2108FS1	8/2023	4.00	19.1	4.03	150.0	P/F	L/F
CCV	↓	↓	1623	2108H37	2/2023	10.06	18.9	10.03	-195.8	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 166.9, slope from 4 to 7 180.9 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 11/10/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	12/28/21	820	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Cyanobacteria Bloom Response Field Data Sheet

RQ-2021- 10-11-60

Collected By: J Pansh / B. Davis / V. Appleby

Customer: NEROC ZIFA

Field Report

Prepared By: J Pansh

Project ID:

BLOOM_RESP

Sampling Agency:

FDEP

NEROC

WIN ID/Field ID:



NEHAB0007

St. Johns River at Beechers Point

Comments:

Rubin & Rocky Mcintyre collected with
Riverbed condos @ bell south west 200 meters
of win site

Collection

Equipment:

Pole, Bottle

Matrix:

SW Fresh

Grab

Coordinates:

29.466300, -81.685230

Tide:

☐ Rising

☒ Falling

☐ Slack

☐ N/A

High Tide Time:

Low Tide Time:

Water Quality

Date	Time ☒ EST ☐ CEN	D.O. (mg/L)	D.O. (%)	Temp °C	pH (SU)	Sample Depth (m)	Cond. (umho/cm)	Salinity (PPTH)	Secchi Disk (m)	Total Depth (m)
<u>12/28/21</u>	<u>1035</u>	<u>8.87</u>	<u>95.9</u>	<u>18.9</u>	<u>7.89</u>	<u>0.3</u>	<u>911</u>	<u>0.45</u>	<u>0.6</u>	<u>1.2</u>

Bloom Observations

Algal Bloom Observed?

☐ Yes

☒ No

Water Surface

☒ Clear

☐ Scum

☐ Globs

☐ Slick

Water Clarity

☐ Clear

☒ Slightly Turbid

☐ Turbid

☐ Opaque

Bloom Color

☐ Green

☐ Red

☐ Brown

☒ Other: None

Algae Distribution

☐ Suspended in water column

☐ On surface

☒ Other: None seen

Water Odors

☒ Normal

☐ Sewage

☐ Petroleum

☐ Chemical

Other:

Parameter Suite

Parameter Methods

Preservation

pH
Check

#Bottles
Sent to
Lab

Bottle
Group

Bloom ID (PT-50ML)

☒ BLOOM-ID

☒ Ice

1

A

Chlorophyll (BP-1L)

☒ CHLSUITE-W

☒ Ice

1

B

Toxins (BG-WD250)

☒ W-MCYST-AA

☒ Ice

1

B

Preserved Nutrients
(500mL HDPE)

☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP

☒ Ice

☒ H2SO4

☒ <2

1

B

Lot

SA 11 05060

Exp:

4/30/22

☒ <2

1

Filtered Nutrients
(P125mL)

☒ W-PO4-F

☒ Field Filtered 0.45um Filter

☒ Ice

1

B

Quantitative Algae

☐ DTY-QN-C / PKD-QN-C

☐ Ice

☐ Lugols

Relinquished By:

Date/Time

Number of Coolers Shipped

Received By:

Date/Time

J Pansh

12/28/21

1