

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Bruce

RQ: 2022-05-09-53

Project: Algae Response

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/2/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Bruce Davis	6/1/22	757	22.2	761.4	8.7	✓	100.2	✓	1.02	P/F	L/F
ICV	↓	↓	800	22.1	761.4	8.8	8.76	100.3	✓	1.02	P/F	L/F
CCV	↓	↓	1643	23.9	759.4	8.4	8.40	99.5	✓	1.02	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Bruce Davis	6/1/22	727	220225A	3/23	1000	1000	P/F	L/F
ICV	↓	↓	729	210917A	9/22	100	102.3	P/F	L/F
CCV	↓	↓	1646	220225A	3/23	1000	997	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Bruce Davis	6/1/22	743	4203B87	2/24	7.01	22.4	7.01	-12.2	P/F	L/F
Calibr.			747	2201P88	1/24	4.00	22.7	4.00	166.3	P/F	L/F
Calibr.			751	2202A63	7/23	10.02	22.7	10.02	-184.0	P/F	L/F
ICV	↓	↓	754	4203B87	2/24	7.01	22.4	7.08	-13.9	P/F	L/F
CCV	↓	↓	1648	4203B87	2/24	7.01	22.8	7.13	-168	P/F	L/F
CCV	↓	↓	1650	2202A63	7/23	10.03	22.6	10.03	-185.9	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 171.8, slope from 4 to 7 178.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/3/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Bruce Davis	6/1/22	801	0.00	0.005	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-05-20 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-09-53				Collected By:		Brian Davis, Katie Halbert					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Newnan's Lake - Near Windsor Park				Comments:							
Field ID: NEHAB0064				Coordinates:		29.63666, -82.20232					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0064				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
06/01/2022	10:10	57.2	4.54	27.2	6.44	0.15	0.3	0.3	71.8	0.03	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☐ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☐ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☐ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☐ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae ID (PT-50mL)		☐ ALGAE-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (BG-WD250)		☒ W-MCYST-AA / W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP Acid Lot: SA1308030 Exp: 12/01/22				☒ Ice ☒ H2SO4		☒ <2	1	B	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
				Filter Lot # R1EB01329							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Brian Davis		06/01/2022 16:00		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-06-06 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-09-53				Collected By:		Brian Davis, Katie Halbert					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Lochloosa Lake entrance to Cross Creek				Comments: Very windy and white capping							
Field ID: NEHAB0079				Coordinates:		29.50741, -82.14958					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0079				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
06/01/2022	13:20	114.1	8.71	29.3	7.97	0.3	2.3	0.5	92.8	0.04	
	13:21	80.4	6.24	28.8	0.07	1.8				0.04	
	13:22	46	3.56	28.5	6.73	2				0.04	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae ID (PT-50mL)		☐ ALGAE-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (BG-WD250)		☒ W-MCYST-AA / W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B	
		Acid Lot:		SA1308030		Exp:		12/01/22			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
		Filter Lot #		R1EB01239							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Brian Davis		06/01/2022 16:00		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-05-20 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-09-53				Collected By:		Brian Davis, Katie Halbert					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Orange Lake-McIntosh Bay				Comments:							
Field ID: NEHAB0065				Coordinates:		29.43785, -82.18916					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0065				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ■ EST □ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
06/01/2022	11:40	111.8	8.74	28.1	9.16	0.3	2.4	0.8	85	0.04	
	11:44	4.7	0.37	25.3	6.15	1.9			106.4	0.05	
	11:47	3.1	0.26	24.7	6.09	2			109.8	0.05	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae ID (PT-50mL)		☐ ALGAE-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (BG-WD250)		☒ W-MCYST-AA / W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP SA1308030 Acid Lot: Exp:				☒ Ice ☒ H2SO4		<2	1	B	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
				Filter Lot # R1EB01329							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped				Received By:		Date/Time	
Brian Davis		06/01/2022 16:00		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-06-06 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-09-53				Collected By:		Brian Davis, Katie Halbert				
Customer:		WQAP		Field Report Prepared By:		Brian Davis				
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC				
Location: 183rd ave Canal off Cross Creek				Comments:						
Field ID: NEHAB0076				Coordinates:		29.48508, -82.16343				
WIN Secondary Resource Type: Canal				Collection Equip:		WaterBottle				
WIN ID: NEHAB0076				High Tide Time:		Low Tide Time:				
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A		
Water Quality										
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)
06/01/2022	12:25	14.4	1.16	26.1	6.46	0.3	0.6	0.6	170.2	0.08
Bloom Observations										
Algal Bloom Observed? ☒ Yes ☐ No										
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque		
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:		
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:				
Algae Accumulation		☐ Mat		☒ Specks		☒ Scum		☐ Glob		
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended ☐ Benthic ☐ Epiphytic		
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group
Algae ID (PT-50mL)		☐ ALGAE-ID				☐ Ice				
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B
Toxins (BG-WD250)		☒ W-MCYST-AA / W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B
		Acid Lot:		SA1308030		Exp:		12/1/22		
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B
				Filter Lot # R1EB01239						
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols				
Relinquished By:		Date/Time		Number of Coolers Shipped			Received By:		Date/Time	
Brian Davis		06/01/2022 16:00		2						