

RQ-2022-05-23-47

Login Checklist

Login Station ID: #2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 6/24/2022 9:10

*Cooler Temperature	Samples Frozen? YES	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
		HNO ₃	Formalin		Present?		*Intact?		
					Yes	No	Yes	No	
2.4				15		✓			5225 5547 0500

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 °C (above 10.0 °C for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☐ No ☒

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☐ No ☐Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☐ No ☒ Init: IM

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☐ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ☒ No ☐ NA ☐ Init: IM

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes ☐ No ☐ NA ☒ Init: IM

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: IM

Date: 6/24/2022

Event contains NCRs (Y/N)? ☒

Event ID:

Algae Bloom Response-as needed
WQAP-2022-06-24-01 (BLOOM-RESP)
Date Received: 24-JUN-2022 09:10

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #2 exp:10/19/22
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	210620 exp: 4/15/23
Chlorine Test Strips	0070 exp: 02/2023

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

RQ-2022-05-23-47

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: YSI 0051RQ: 2022-06-13-19Project: LSR main

Boldly "F" this
box if there are
qualified data on
this page.

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/2/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Plamsh</u>	<u>6/23/22</u>	<u>730</u>	<u>22.5</u>	<u>757</u>	<u>8.6</u>	<u>8.6</u>	<u>99.7</u>	<u>—</u>	<u>1.025</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>2</u>	<u>6/23/22</u>	<u>731</u>	<u>22.5</u>	<u>757</u>	<u>8.6</u>	<u>8.6</u>	<u>99.6</u>	<u>—</u>	<u>1.025</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>Plamsh</u>	<u>6/23/22</u>	<u>1006</u>	<u>24.5</u>	<u>755</u>	<u>8.3</u>	<u>8.35</u>	<u>100.1</u>	<u>—</u>	<u>—</u>	<u>(P) F</u>	<u>(L) F</u>
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Plamsh</u>	<u>6/23/22</u>	<u>733</u>	<u>226225A</u>	<u>3/23</u>	<u>1000</u>	<u>1000</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>2</u>	<u>6/23/22</u>	<u>734</u>	<u>210417A</u>	<u>9/22</u>	<u>100</u>	<u>97.3</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>Plamsh</u>	<u>6/23/22</u>	<u>812</u>	<u>210417C</u>	<u>9/22</u>	<u>50000</u>	<u>49677</u>	<u>(P) F</u>	<u>(L) F</u>
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Plamsh</u>	<u>6/23/22</u>	<u>742</u>	<u>4203857</u>	<u>2/24</u>	<u>7.</u>	<u>22.6</u>	<u>7.01</u>	<u>-12.7</u>	<u>(P) F</u>	<u>(L) F</u>
Calibr.	<u>2</u>	<u>6/23/22</u>	<u>742</u>	<u>2205F96</u>	<u>5/24</u>	<u>4.</u>	<u>22.3</u>	<u>4.00</u>	<u>164</u>	<u>(P) F</u>	<u>(L) F</u>
Calibr.	<u>2</u>	<u>6/23/22</u>	<u>743</u>	<u>2202A63</u>	<u>7/23</u>	<u>10.</u>	<u>22.7</u>	<u>10.03</u>	<u>184</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>2</u>	<u>6/23/22</u>	<u>745</u>	<u>2205F96</u>	<u>2/24</u>	<u>4.</u>	<u>22.2</u>	<u>4.03</u>	<u>161.5</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>2</u>	<u>6/23/22</u>	<u>1013</u>	<u>2202A63</u>	<u>7/23</u>	<u>10</u>	<u>23.8</u>	<u>9.99</u>	<u>184</u>	<u>(P) F</u>	<u>(L) F</u>
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;If mV are recorded: slope from 7 to 10 171.7, slope from 4 to 7 176.7 (both must be between 165 and 180 mV)Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/3/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>Plamsh</u>	<u>6/23/22</u>	<u>750</u>	<u>0.00</u>	<u>0.00</u>	<u>(P) F</u>	<u>(L) F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-06-10 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-23-47		Collected By:		Katrin Villinger, Jeremy Parrish						
Customer: WQAP		Field Report Prepared By:		Katrin Villinger						
Project ID: BLOOM_RESP		Sampling Agency:		NEROC						
Location: Dr's Lake at Camp Echokotee			Comments: Windy. Water has odor							
Field ID: NEHAB0085			Coordinates: 30.15050, -81.72513							
WIN Secondary Resource Type: Lake			Collection Equip: WaterBottle sampling pole							
WIN ID:			High Tide Time: 746		Low Tide Time: 1248					
Matrix: Marine		Grab		Tide: ☐ Rising ☒ Falling ☐ Slack ☐ N/A						
Water Quality										
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)
06/23/2022	10:15	121.3	8.94	29.6	8.28	0.3	0.6	0.3	9264	5.15
Bloom Observations										
Algal Bloom Observed? ☒ Yes ☐ No										
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque		
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:		
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:				
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☒ Glob		
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended ☐ Benthic ☐ Epiphytic		
Parameter Suite		Parameter Methods		Preservation		pH Check		#Bottles Sent to Lab		Bottle Group
Algal ID (PT-50mL)		☐ ALGAL-ID		☐ Ice				1		A
Bloom ID (PT-50mL)		☒ BLOOM-ID		☒ Ice				1		B
Chlorophyll (BP-1L)		☒ CHLSUITE-W		☒ Ice				1		B
Toxins (BG-WD250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN		☒ Ice				1		B
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP		☒ Ice ☒ H2SO4		☒ <2		1		B
		SA1308030		12/1/22						
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice		1		B
		Filter Lot # R1EB01329								
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C		☐ Ice		☐ Lugols				
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time		
Katrin Villinger		06/23/2022 16:30		1		IM		6/24/22 9:10		



Cyanobacteria Bloom Response Field Data Sheet (2022-06-10 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-23-47		Collected By:		Katrin Villinger, Jeremy Parrish						
Customer:	WQAP	Field Report Prepared By:		Katrin Villinger						
Project ID:	BLOOM_RESP	Sampling Agency:		NEROC						
Location: Dr's Lake Near Lucy Branch				Comments: Windy, slight odor						
Field ID: NEHAB0086				Coordinates: 30.13962, -81.74511						
WIN Secondary Resource Type: Lake				Collection Equip: WaterBottle sampling pole						
WIN ID:		High Tide Time:		726	Low Tide Time: 1248					
Matrix:	Marine	Grab	Tide: ☐ Rising ☒ Falling ☐ Slack ☐ N/A							
Water Quality										
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPT)
06/23/2022	10:45	156.6	11.27	31.1	8.81	0.3	0.77	0.45	9636	5.36
	10:50	153.2	11.04	31.1	8.8	0.6			9636	5.36
Bloom Observations										
Algal Bloom Observed? ☒ Yes ☐ No										
Water Clarity		☐ Clear ☒ Slightly Turbid ☐ Turbid ☐ Opaque								
Bloom Color		☒ Green ☐ Red ☐ Brown ☐ Other:								
Algae Type		☒ Planktonic ☐ Filamentous ☐ Other:								
Algae Accumulation		☐ Mat ☒ Specks ☐ Scum ☐ Glob								
Algae Distribution		☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☒ Suspended ☐ Benthic ☐ Epiphytic								
Parameter Suite		Parameter Methods		Preservation		pH Check	#Bottles Sent to Lab	Bottle Group		
Algal ID (PT-50mL)		☐ ALGAL-ID		☐ Ice						
Bloom ID (PT-50mL)		☒ BLOOM-ID		☒ Ice			1	A		
Chlorophyll (BP-1L)		☒ CHLSUITE-W		☒ Ice			1	B		
Toxins (BG-WD250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN		☒ Ice			1	B		
		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP		☒ Ice ☒ H2SO4		☒ <2	1	B		
Preserved Nutrients (500mL HDPE)		SA1308030		12/1/22						
		Acid Lot:		Exp:						
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45µm Filter		☒ Ice	1	B		
		Filter Lot # R1EB01329								
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C		☐ Ice ☐ Lugols						
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time		
Katrin Villinger		06/23/2022 16:30		1		IM		6/24/22 9:10		

Katrin E. Villinger



Cyanobacteria Bloom Response Field Data Sheet (2022-06-10 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-05-23-47		Collected By:		Katrin Villinger, Jeremy Parrish	
Customer:	WQAP	Field Report Prepared By:		Katrin Vilinger	
Project ID:	BLOOM_RESP	Sampling Agency:		NEROC	
Location: Dr's Lake end of Lawrence Rd			Comments: Windy		
Field ID: NEHAB0073			Coordinates: 30.12961, -81.75171		
WIN Secondary Resource Type: Lake			Collection Equip: WaterBottle Sampling pole		
WIN ID:			High Tide Time: 726		Low Tide Time: 1248
Matrix:	Marine	Grab	Tide: ☐ Rising ☒ Falling ☐ Slack ☐ N/A		

Water Quality										
Date	Time ■ EST □ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)
06/23/2022	11:05	93.9	6.79	31	8.31	0.3	0.65	0.35	9584	5.33

Bloom Observations	
Algal Bloom Observed?	☒ Yes ☐ No
Water Clarity	☐ Clear ☐ Slightly Turbid ☒ Turbid ☐ Opaque
Bloom Color	☒ Green ☐ Red ☐ Brown ☐ Other:
Algae Type	☒ Planktonic ☐ Filamentous ☐ Other:
Algae Accumulation	☒ Mat ☐ Specks ☐ Scum ☒ Glob
Algae Distribution	☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☒ Suspended ☐ Benthic ☐ Epiphytic

Parameter Suite	Parameter Methods	Preservation	pH Check	#Bottles Sent to Lab	Bottle Group
Algal ID (PT-50mL)	☐ ALGAL-ID	☐ Ice		1	A
Bloom ID (PT-50mL)	☒ BLOOM-ID	☒ Ice			
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice			
Toxins (BG-WD250)	☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN	☒ Ice			
Preserved Nutrients (500mL HDPE)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP SA1308030 Acid Lot: Exp: 12/1/22	☒ Ice ☒ H2SO4	☒ <2	1	B
Filtered Nutrients (P125mL)	☒ W-PO4-F ☐ Field Filtered 0.45µm Filter Filter Lot #	☒ Ice		1	R1EB01329
Quantitative Algae	☐ DTY-QN-C / PKD-QN-C	☐ Ice: ☐ Lugols			

Relinquished By:	Date/Time	Number of Coolers Shipped	Received By:	Date/Time
Katrin Villinger	06/23/2022 16:30	1	IM	6/24/22 9:10

Date of Request: 17-MAY-2022
 Created By: DAVIS_B On: 17-MAY-2022
 Modified By: JASROTIA_P On: 18-MAY-2022
 Customer: WQAP
 Project: BLOOM-RESP
 Division: Division of Environmental Assessment and Restoration
 District:
 Event Description: Algae Bloom Response-as needed

Send Coolers To:

Phone: 904-256-1568
 8800 Baymeadows Way W
 Suite 100
 Jacksonville, FL 32256
 Attn: Brian G Davis
 Cooler Ship EMail: Brian.g.davis@floridadep.gov

Priority: 1
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 18-MAY-2022
 Biology Request Reviewed By: JASROTIA_P On: 18-MAY-2022
 Sampling Kit Required: YES Ship on: 20-MAY-2022
 Sampling Kit Shipped: YES On: 19-MAY-2022
 Sampling Kit Packed By: MCCOY_I
 Preservatives Needed: NO
 Date to Receive Samples: 23-MAY-2022
 Logged In By: SHAIK_A On: 24-MAY-2022

Send Final Report To:

8800 Baymeadows Way W
 Suite 100
 Tallahassee, FL 32399-2400
 Attn: Brian G Davis
 Report Notification EMail: Brian.g.davis@floridadep.gov

Comments: RQ specifically made for addition of Saxitoxin analysis in the LSJR basin.

Bottle Group A - Priority 1
 For algal ID dominant taxon.

Bottle Group B - Priority 1
 For toxin testing Nutrients, Chlorophyll, Toxin.

Bottle Group A (Biological) With 4 Samples:

Bottle Type: PT-50ML Number of Bottles: 4 Preserved With: ICE
 BLOOM-ID Template: DEFAULT (SOP-AB05) Assessment of dominant algal taxa in bloom or mat sample

Bottle Group B (Water) With 5 Samples:

Bottle Type: BP-1L Number of Bottles: 5 Preserved With: ICE
 CHLSUITE-W Template: DEFAULT (SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: BG-250ML-WDMTH Number of Bottles: 5 Preserved With: ICE
 W-MCYST-AA Template: DEFAULT (EPA 8321B) Microcystins in water matrices by HPLC/MS/MS
 W-SAXTN-MS Template: DEFAULT (EPA 8321B) Saxitoxin in water matrices by HPLC/MS/MS

Bottle Type: P-500ML Number of Bottles: 5 Preserved With: ICE And H2SO4
 W-NH3 Template: DEFAULT (EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
 W-NO2NO3 Template: DEFAULT (EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
 W-S-A-TP Template: DEFAULT (EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
 W-TKN Template: DEFAULT (EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 5 Preserved With: FILTER-ICE
 W-PO4-F Template: DEFAULT (EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.

