

Login Checklist

RQ- 2022-08-22-52

Login Station ID: #1

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 9/13/22 @ 10:00

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
	YES	HNO ₃	Formalin		Present?		*Intact?		
					Yes	No	Yes	No	
2.1				9		✓			5225 5546 4103

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☐ No ☒

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒

Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☐ No ☐

Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☐ No ☒ Init: HK

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☐ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes ☒ No ☐ NA ☐ Init: HK

**W-1-4-DIOX (Green dot on container) preserved to pH<4.0?

Yes ☐ No ☒ NA ☐ Init: HK

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: HK

Date: 9/13/22

Event contains NCRs (Y/N)? N

Event ID: Algae Bloom Response NEROC-as needed

WQAP-2022-09-13-03 (BLOOM-RESP)

Date Received: 13-SEP-2022 10:00

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #1 Exp:10/19/2022
pH Test Strips 0 – 6	Lot# 223819AV Exp:8/30/23
pH Paper 0 – 3	
pH Paper 0 – 13	Lot# 207621 Exp:3/15/24
Chlorine Test Strips	Lot# 0070 Exp:2/23

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000 FT 1500, FD 1000-FD 4000)

Boldly "P" this box if there are qualified data on this page

Meter ID: Buce

RQ: 2022 08 22 52

Project: Blade Oysters

River

- Notes: (1) Always wait for meter to stabilize before recording any readings
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification: 8/1/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Buce</u>	<u>9/1/22</u>	<u>707</u>	<u>24.2</u>	<u>760.1</u>	<u>8.4</u>	<u>8.4</u>	<u>100.0</u>	<u>✓</u>	<u>1.03</u>	<u>P/F</u>	<u>L/F</u>
ICV		<u>↓</u>	<u>708</u>	<u>24.3</u>	<u>760.1</u>	<u>8.4</u>	<u>8.39</u>	<u>100.2</u>	<u>✓</u>	<u>1.03</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>K. Villages</u>	<u>↓</u>	<u>1139</u>	<u>24.4</u>	<u>760.1</u>	<u>8.4</u>	<u>8.39</u>	<u>100.4</u>	<u>✓</u>	<u>1.03</u>	<u>P/F</u>	<u>L/F</u>
CCV											<u>P/F</u>	<u>L/F</u>

DO Acceptance criteria from Table 1: ± 0.3 mg/L

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Buce</u>	<u>7/12/22</u>	<u>910</u>	<u>220254</u>	<u>3/22</u>	<u>1000</u>	<u>1000</u>	<u>P/F</u>	<u>L/F</u>
ICV		<u>↓</u>	<u>911</u>	<u>220254</u>	<u>9/22</u>	<u>15,000</u>	<u>15,012</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>K. Villages</u>	<u>↓</u>	<u>1141</u>	<u>220254</u>	<u>3/23</u>	<u>1000</u>	<u>998</u>	<u>P/F</u>	<u>L/F</u>
CCV								<u>P/F</u>	<u>L/F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Buce</u>	<u>9/1/22</u>	<u>919</u>	<u>M07420</u>	<u>3/24</u>	<u>7.00</u>	<u>24.0</u>	<u>7.00</u>	<u>-101</u>	<u>P/F</u>	<u>L/F</u>
Calibr.		<u>↓</u>	<u>916</u>	<u>120576</u>	<u>5/24</u>	<u>4.00</u>	<u>24.7</u>	<u>4.00</u>	<u>166.0</u>	<u>P/F</u>	<u>L/F</u>
Calibr.		<u>↓</u>	<u>917</u>	<u>120279</u>	<u>5/23</u>	<u>10.01</u>	<u>24.3</u>	<u>10.01</u>	<u>-154.0</u>	<u>P/F</u>	<u>L/F</u>
ICV		<u>↓</u>	<u>721</u>	<u>M07420</u>	<u>3/24</u>	<u>7.00</u>	<u>23.9</u>	<u>7.04</u>	<u>-11.3</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>K. Villages</u>	<u>↓</u>	<u>1143</u>	<u>220596</u>	<u>5/24</u>	<u>4.00</u>	<u>25.3</u>	<u>4.01</u>	<u>165.6</u>	<u>P/F</u>	<u>L/F</u>
CCV										<u>P/F</u>	<u>L/F</u>

pH Acceptance criteria ± 0.2 SU, mV pH 7 Range 0 ± 50 .

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 .

If mV are recorded: slope from 7 to 10 172.9, slope from 4 to 7 176.1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/1/22

If NO, what will be used? Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>Buce</u>	<u>7/12/22</u>	<u>923</u>	<u>0.00</u>	<u>0.007</u>	<u>P/F</u>	<u>L/F</u>

Report two decimal places. Round numbers < 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52		Collected By:		Brian Davis, Katrin Villinger	
Customer:	WQAP	Field Report Prepared By:		Brian Davis	
Project ID:	BLOOM_RESP	Sampling Agency:		NEROC	
Location: Ortega River @ Morven Canal			Comments: Login the DTY-QN-C/PKD-QN-C sample/s under a separate event, priority 12		
Field ID: NEHAB0090		Coordinates:		30.25259, -81.70317	
WIN Secondary Resource Type: River/Stream		Collection Equip:		Dipper	
WIN ID: NEHAB0090		High Tide Time:		07:00	Low Tide Time: 13:30
Matrix:	Fresh	Grab	Tide:	☒ Rising ☐ Falling ☐ Slack ☐ N/A	
Water Quality					
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)
09/12/2022	10:30	33.3	2.57	28	6.81
Bloom Observations					
Algal Bloom Observed? ☐ Yes ☒ No					
Water Clarity	☐ Clear	☒ Slightly Turbid	☐ Turbid	☐ Opaque	
Bloom Color	☐ Green	☐ Red	☐ Brown	☐ Other:	
Algae Type	☐ Planktonic	☐ Filamentous	☐ Other:		
Algae Accumulation	☐ Mat	☐ Specks	☐ Scum	☐ Glob	
Algae Distribution	☐ Surface-Entire	☐ Surf-Localized	☐ Surf-Windblown	☐ Suspended	☐ Benthic ☐ Epiphytic
Parameter Suite	Parameter Methods		Preservation		pH Check
Algal ID (PT-50mL)	☐ ALGAL-ID		☐ Ice		
Bloom ID (PT-50mL)	☒ BLOOM-ID		☒ Ice		1
Chlorophyll (BP-1L)	☒ CHLSUITE-W		☒ Ice		1
Toxins (P-125/BG-250)	☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN		☒ Ice		1
Preserved Nutrients (500mL HDPE)	☒ W-NH3 / ☒ W-NO2NO3 / ☒ W-TKN / ☒ W-S-A-TP		☒ Ice ☒ H2SO4		1
	Acid Lot: SA1245160 Exp: 6/2/23				
Filtered Nutrients (P125mL)	☒ W-PO4-F		☒ Field Filtered 0.45um Filter		1
	Filter Lot # 175533		☒ Ice		
Quantitative Algae	☒ DTY-QN-C / PKD-QN-C		☒ Ice ☐ Lugols		1
Relinquished By:	Date/Time	Number of Coolers Shipped		Received By:	Date/Time
Brian Davis	09/12/2022 16:30	1		HK	9.13.22 / 1000



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52		Collected By:		Brian Davis, Katrin Villinger	
Customer:	WQAP	Field Report Prepared By:		Brian Davis	
Project ID:	BLOOM_RESP	Sampling Agency:		NEROC	
Location: Ortega River @ Morven Canal			Comments: 09/12/2022 10:35 ET DISource: NED Lab DILocation: NED Lab Container ID/Type: CarNB / Carboy		
Field ID: FB_09122022			Coordinates:		
WIN Secondary Resource Type:			Collection Equip:		
WIN ID:			High Tide Time:	07:00	Low Tide Time: 13:30
Matrix:	Field Blank	Grab	Tide:	☐ Rising ☐ Falling ☐ Slack ☐ N/A	
Water Quality					
Date	Time ☐ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)
Bloom Observations					
Algal Bloom Observed? ☐ Yes ☐ No					
Water Clarity	☐ Clear	☐ Slightly Turbid	☐ Turbid	☐ Opaque	
Bloom Color	☐ Green	☐ Red	☐ Brown	☐ Other: _____	
Algae Type	☐ Planktonic	☐ Filamentous	☐ Other: _____		
Algae Accumulation	☐ Mat	☐ Specks	☐ Scum	☐ Glob	
Algae Distribution	☐ Surface-Entire	☐ Surf-Localized	☐ Surf-Windblown	☐ Suspended	☐ Benthic ☐ Epiphytic
Parameter Suite	Parameter Methods		Preservation	pH Check	#Bottles Sent to Lab
Algal ID (PT-50mL)	☐ ALGAL-ID		☐ Ice		
Bloom ID (PT-50mL)	☐ BLOOM-ID		☐ Ice		
Chlorophyll (BP-1L)	☐ CHLSUITE-W		☐ Ice		
Toxins (P-125/BG-250)	☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN		☒ Ice		1 D
Preserved Nutrients (500mL HDPE)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP		☒ Ice ☒ H2SO4	☒ <2	1 D
	Acid Lot:	SA1245160	Exp:	6/2/23	
Filtered Nutrients (P125mL)	☒ W-PO4-F	☒ Field Filtered 0.45um Filter	☒ Ice		
	Filter Lot #	175533			
Quantitative Algae	DTY-QN-C / PKD-QN-C		☐ Ice ☐ Lugols		
Relinquished By:	Date/Time	Number of Coolers Shipped		Received By:	Date/Time
Brian Davis	09/12/2022 16:30	1		HK	9.13.22 / 1000

Date of Request: 17-AUG-2022
 Created By: DAVIS_B On: 17-AUG-2022
 Modified By: JASROTIA_P On: 18-AUG-2022
 Customer: WQAP
 Project: BLOOM-RESP
 Division: Division of Environmental Assessment and Restoration
 District:
 Event Description: Algae Bloom Response NEROC-as needed

Send Coolers To:

Phone: 904-256-1568
 8800 Baymeadows Way W
 Suite 100
 Jacksonville, FL 32256
 Attn: Brian G Davis
 Cooler Ship EMail: Brian.g.davis@floridadep.gov

Priority: 1
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 18-AUG-2022
 Biology Request Reviewed By: JASROTIA_P On: 18-AUG-2022
 Sampling Kit Required: YES Ship on: 22-AUG-2022
 Sampling Kit Shipped: YES On: 22-AUG-2022
 Sampling Kit Packed By: SHAIK_A
 Preservatives Needed: YES
 Date to Receive Samples: 22-AUG-2022
 Logged In By: SHAIK_A On: 02-SEP-2022

Send Final Report To:

8800 Baymeadows Way W
 Suite 100
 Tallahassee, FL 32399-2400
 Attn: Brian G Davis
 Report Notification EMail: Brian.g.davis@floridadep.gov

Comments: -Do not send bottles in coolers. Can be shipped in Cardboard Boxes-

Bottle Group A - Priority 1
 For Bloom ID dominant taxon.

Bottle Group B - Priority 1
 For toxin testing Nutrients, Chlorophyll, Toxin.

Bottle Group C - Priority 1
 For toxin testing Nutrients, Chlorophyll, Toxin w/ SAX.

Bottle Group D - Priority 1
 Field Blank for QC on Nutrients and Toxin.

 Include 2 boxes of Sulfuric Acid.

Bottle Group A (Biological) With 12 Samples:

Bottle Type: PT-50ML Number of Bottles: 12 Preserved With: ICE
 BLOOM-ID Template: DEFAULT (SOP-AB18) Assessment of dominant algal taxa in bloom or mat sample

Bottle Group B (Water) With 6 Samples:

Bottle Type: BP-1L Number of Bottles: 6 Preserved With: ICE
 CHLSUITE-W Template: DEFAULT (SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-125ML Number of Bottles: 6 Preserved With: ICE
 W-MCYST-AA Template: DEFAULT (EPA 8321B) Microcystins in water matrices by HPLC/MS/MS
 W-SAXTN-MS Template: DEFAULT (EPA 8321B) Saxitoxin in water matrices by HPLC/MS/MS

Bottle Type: P-500ML Number of Bottles: 6 Preserved With: ICE And H2SO4
 W-NH3 Template: DEFAULT (EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
 W-NO2NO3 Template: DEFAULT (EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
 W-S-A-TP Template: DEFAULT (EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
 W-TKN Template: DEFAULT (EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 6 Preserved With: FILTER-ICE

Bottle Group B (Water) With 6 Samples:

Bottle Type: P-125ML	Number of Bottles: 6	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group C (Water) With 6 Samples:

Bottle Type: BP-1L	Number of Bottles: 6	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-125ML	Number of Bottles: 6	Preserved With: ICE
W-MCYST-AA	Template: DEFAULT	(EPA 8321B) Microcystins in water matrices by HPLC/MS/MS
W-SAXTN-MS	Template: DEFAULT	(EPA 8321B) Saxitoxin in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 6	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices

Bottle Type: P-125ML	Number of Bottles: 6	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group D (Water) With 1 Samples:

Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: ICE
W-MCYST-AA	Template: DEFAULT	(EPA 8321B) Microcystins in water matrices by HPLC/MS/MS
W-SAXTN-MS	Template: DEFAULT	(EPA 8321B) Saxitoxin in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices

Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.