

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Brude

RQ: 08-22-52

Project: Algae Response Sampson River

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 8/1/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Katrin Villinger	8/30/22	724	22.7	760.7	8.6	-	100.1	-	1.025	P/F	L/F
ICV	↓	↓	725	22.7	760.6	8.6	8.45	100.2	-	1.025	P/F	L/F
CCV	↓	↓	1649	23.7	758.8	8.5	8.48	100.1	-	1.025	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	K Villinger	8/30/22	727	220225A	3/23	1000	1000	P/F	L/F
ICV	↓	↓	729	210917A	9/22	100	100.3	P/F	L/F
CCV	↓	↓	1652	220225A	3/23	1000	1001	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K Villinger	8/30/22	731	M074-20	3/21/24	7.01	22.8	7.01	-12.6	P/F	L/F
Calibr.	↓	↓	735	2205F96	5/24	4.00	22.6	4.00	164.6	P/F	L/F
Calibr.	↓	↓	737	2202F29	9/23	10.02	22.9	10.02	-78.4	P/F	L/F
ICV	↓	↓	741	2205F96	5/24	4.00	23.0	3.94	116.2	P/F	L/F
CCV	↓	↓	1655	M074-20	3/21/24	7.01	22.8	7.01	-10.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
If mV are recorded: slope from 7 to 10 168.8, slope from 4 to 7 177.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/1/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K Villinger	8/30/22	742	0.000	0.000	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Katrin Villinger, Brian Davis					
Customer:		Wqap		Field Report Prepared By:		Katrin villinger					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Sampson River @ SW CR 225				Comments:							
Field ID: NEHAB0088				Coordinates:		29.91935, -82.21014					
WIN Secondary Resource Type: River/Stream				Collection Equip:		WaterBottle					
WIN ID: NEHAB0088				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
08/30/2022	13:00	63.8	4.84	29.9	6.82	0.3	0.9	0.7	114	0.05	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☒ Clear ☐ Slightly Turbid ☐ Turbid ☐ Opaque									
Bloom Color		☐ Green ☐ Red ☐ Brown ☐ Other:									
Algae Type		☐ Planktonic ☐ Filamentous ☐ Other:									
Algae Accumulation		☐ Mat ☐ Specks ☐ Scum ☐ Glob									
Algae Distribution		☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☐ Suspended ☐ Benthic ☐ Epiphytic									
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☐ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP Acid Lot: SA2145160 Exp: 6/2/23				☒ Ice ☒ H2SO4		☒ <2	1	B	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter Filter Lot # 175533		☒ Ice			1	B	
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped			Received By:		Date/Time		
K villinger		08/30/2022 16:30		1							