

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Bruce

RQ- 2022-08-22-52

Project: Agave Response Doctor's Lake

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 8/1/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K. Villinger	9/1/22	723	23.3	760.7	8.5	—	100.1	—	1.023	P/F	L/F
ICV			725	23.2	760.6	8.6	8.56	100.1	—	1.023	P/F	L/F
CCV			1527	24.2	760.5	8.4	8.40	100.2	—	1.023	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	K. Villinger	9/1/22	726	220225A	3/23	1000	1000	P/F	L/F
ICV			730	220225B	3/23	50000	50209	P/F	L/F
CCV			1529	220225A	3/23	1000	1000	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K. Villinger	9/1/22	736	M074-20	3/21/24	7.01	23.4	2.01	-12.9	P/F	L/F
Calibr.			738	2205F96	5/24	4.00	23.6	4.00	165.4	P/F	L/F
Calibr.			743	2202F29	8/23	10.01	23.6	10.01	-183.8	P/F	L/F
ICV			747	2205F96	5/24	4.00	23.7	3.99	165.6	P/F	L/F
CCV				2202F29	8/23	10.01	24.1	9.97	-183.7	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 170.9, slope from 4 to 7 175.3 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/1/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K. Villinger	9/1/22	748	0.000	0.000	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Kattin Villinger, Brian Davis					
Customer:		Wqap		Field Report Prepared By:		Katrin Villinger					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Doctor's Lake @ Mill Cove				Comments:							
Field ID: NEHAB0057				Coordinates:		30.12415, -81.72821					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0057				High Tide Time:		16:00		Low Tide Time:		09:20	
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
09/01/2022	12:25					0.3	1.516	0.6			
	12:30	88.4	6.6	30	8.26	1			3647	1.9	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA2145160		Exp:		6/2/23			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
		Filter Lot #		175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped				Received By:		Date/Time	
Katrin Villinger		09/01/2022 16:30		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Katrin Villinger, Brian Davis						
Customer:		WQAP		Field Report Prepared By:		Katrin Villinger						
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC						
Location: Dr's Lake End of Lawrence Ave				Comments:								
Field ID: NEHAB0073				Coordinates:		30.12947, -81.75121						
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle						
WIN ID: NEHAB0073				High Tide Time:		16:00		Low Tide Time:		09:20		
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A				
Water Quality												
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)		
09/01/2022	12:05	113.4	8.32	29.4	8.7	0.3	1.487	0.7	3551	1.85		
	12:10	109.2	8.16	29.7	8.69	0.9				3560	1.83	
Bloom Observations												
Algal Bloom Observed? ☒ Yes ☐ No												
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque				
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:				
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:						
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob				
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic		
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group		
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice						
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A		
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C		
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C		
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C		
		Acid Lot:		SA2145160		Exp:		6/2/23				
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C		
				Filter Lot # 175533								
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols						
Relinquished By:		Date/Time		Number of Coolers Shipped				Received By:		Date/Time		
Katrin Villinger		09/01/2022 16:30		2								



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52		Collected By:		Katrin Villinger, Brian Davis						
Customer:	Wqap	Field Report Prepared By:		Katrin Villinger						
Project ID:	BLOOM_RESP	Sampling Agency:		NEROC						
Location: Swimming Pen Cr @ Whitey's Fish camp			Comments:							
Field ID: NEHAB0074			Coordinates:	30.10103, -81.74644						
WIN Secondary Resource Type: River/Stream			Collection Equip:	WaterBottle						
WIN ID: NEHAB0074			High Tide Time:	16:00	Low Tide Time: 21:20					
Matrix:	Fresh	Grab	Tide:	☒ Rising ☐ Falling ☐ Slack ☐ N/A						
Water Quality										
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)
09/01/2022	12:55	57	4.3	29.8	7.19	0.3	0.939	0.7	3118	1.61
Bloom Observations										
Algal Bloom Observed? ☒ Yes ☐ No										
Water Clarity ☐ Clear ☒ Slightly Turbid ☐ Turbid ☐ Opaque										
Bloom Color ☒ Green ☐ Red ☐ Brown ☐ Other:										
Algae Type ☒ Planktonic ☐ Filamentous ☐ Other:										
Algae Accumulation ☐ Mat ☒ Specks ☐ Scum ☐ Glob										
Algae Distribution ☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☒ Suspended ☐ Benthic ☐ Epiphytic										
Parameter Suite		Parameter Methods		Preservation		pH Check	#Bottles Sent to Lab	Bottle Group		
Algal ID (PT-50mL)		☐ ALGAL-ID		☐ Ice						
Bloom ID (PT-50mL)		☒ BLOOM-ID		☒ Ice			1	A		
Chlorophyll (BP-1L)		☒ CHLSUITE-W		☒ Ice			1	C		
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN		☒ Ice			1	C		
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP		☒ Ice ☒ H2SO4		☒ <2	1	C		
		Acid Lot: SA2145160		Exp: 6/2/23						
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice	1	C		
		Filter Lot # 175533								
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C		☐ Ice ☐ Lugols						
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time		
Katrin Villinger		09/01/2022 16:30		2						



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Katrin Villinger, Brian Davis					
Customer:		Wqap		Field Report Prepared By:		Katrin Villinger					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: LSJR @ 2930 SR 13				Comments: Cloudy							
Field ID: NEHAB0077				Coordinates:		30.05038, -81.66973					
WIN Secondary Resource Type: River/Stream				Collection Equip:		WaterBottle					
WIN ID: NEHAB0077				High Tide Time:		05:00		Low Tide Time:		11:45	
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ■ EST □ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
09/01/2022	09:55	77.5	5.87	29.7	7.47	0.3	4.72	0.9	1160	0.57	
	10:05	75.7	5.73	29.6	7.46	4.22			1217	0.6	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☒ Clear ☐ Slightly Turbid ☐ Turbid ☐ Opaque ☐ Bloom Color ☒ Green ☐ Red ☐ Brown ☐ Other:									
Algae Type		☒ Planktonic ☐ Filamentous ☐ Other: ☐ Algae Accumulation ☐ Mat ☒ Specks ☐ Scum ☐ Glob									
Algae Distribution		☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☒ Suspended ☐ Benthic ☐ Epiphytic									
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP Acid Lot: SA2145160 Exp: 6/2/23				☒ Ice ☒ H2SO4		☒ <2	1	C	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter Filter Lot # 175533		☒ Ice			1	C	
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Katrin Villinger		09/01/2022 16:30		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Katrin Villinger, Brian Davis					
Customer:		Wqap		Field Report Prepared By:		Kattin Villinger					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Dr's Lake at Camp Echocotee				Comments:							
Field ID: NEHAB0085				Coordinates:		30.15037, -81.72481					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0085				High Tide Time:		16:00		Low Tide Time:		09:20	
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ■ EST □ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
09/01/2022	11:25	110.3	8.11	31.1	8.55	0.3	1.031	0.7	3719	1.94	
	11:30	84.7	6.32	30.2	7.95	0.5				3689	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP Acid Lot: 2145160 Exp: 6/2/23				☒ Ice ☒ H2SO4		☒ <2	1	C	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
		Filter Lot # 175533									
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Katrin Villinger		09/01/2022 16:30		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Kattin Villinger, Brian Davis					
Customer:		WQAP		Field Report Prepared By:		Katrin Villinger					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Dr's Lake Near Lucy Branch				Comments:							
Field ID: NEHAB0086				Coordinates:		30.13996, -81.74393					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0086				High Tide Time:		16:00		Low Tide Time:		09:30	
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ■ EST □ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
09/01/2022	11:50	115	8.7	30.4	8.58	0.3	1.834	0.6	3542	1.84	
	11:55	1.8	0.13	29.6	7.01	1.3			3411	1.77	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☐ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☐ W-MCYST-AA ☐ W-SAXTN-MS ☐ OV-TOXN				☐ Ice					
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP Acid Lot: SA2145160 Exp: 6/3/23				☒ Ice ☒ H2SO4		☒ <2	1	B	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter Filter Lot # 175533		☒ Ice			1	B	
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped				Received By:		Date/Time	
Katrin Villinger		09/01/2022 16:30		2							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2022-08-22-52				Collected By:		Kattin Villinger, Brian Davis					
Customer:		Wqap		Field Report Prepared By:		Katrin Villinger					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: LSJR @ BB Fenders				Comments:							
Field ID: NEHAB0093				Coordinates:		30.19079, -81.66793					
WIN Secondary Resource Type: River/Stream				Collection Equip:		WaterBottle					
WIN ID: NEHAB0093				High Tide Time:		15:00		Low Tide Time:		08:30	
Matrix:		Marine		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
09/01/2022	10:55	85.5	6.35	30.1	7.65	0.3	5.567	0.8	4760	2.55	
	11:00	83.9	6.24	30	7.63	5			4971	2.64	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			C	1	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot: 2145160		Exp: 6/2/23							
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
		Filter Lot # 175533									
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Katrin Villinger		09/01/2022 16:30		2							