

Chemical Analysis Report

WQAP-2022-08-10-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Algal Bloom Response - Orange Co**
Request ID: **RQ-2022-08-08-72**
Customer: **WQAP**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
3319 Maguire Blvd
Orlando, FL 32083
Attn: Matthew Bearden

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 31-AUG-2022 10:02

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Cypress Lake- West Lobe

Collection Date/Time: 08/09/2022 09:45

Field ID: HABOC0007_CypressLake1

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2347628	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P417830	
		Cylindrospermopsin	0.10	U	ug/L	P417829	
		Desmethyl microcystin LR	0.25	U	ug/L	P417829	
		Microcystin HILR	0.25	U	ug/L	P417829	
		Microcystin HtyR	0.25	U	ug/L	P417829	
		Microcystin LA	0.10	U	ug/L	P417829	
		Microcystin LF	0.10	U	ug/L	P417829	
		Microcystin LR	0.25	U	ug/L	P417829	
		Microcystin LW	0.25	U	ug/L	P417829	
		Microcystin LY	0.10	U	ug/L	P417829	
		Microcystin RR	0.10	U	ug/L	P417829	
		Microcystin WR	0.50	U	ug/L	P417829	
		Microcystin YR	0.25	U	ug/L	P417829	
		Nodularin-R	0.10	U	ug/L	P417829	
2347662	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	I	mg N/L	P418157	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.89		mg N/L	P418138	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P418107	
	EPA 365.1 Rev. 2.0	Total-P	0.016		mg P/L	P417982	
2347664	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P417910	

Sample Location: Lake Speer- NW Lobe

Collection Date/Time: 08/09/2022 10:50

Field ID: HABOC0009_LakeSpeer1

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2347629	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P417830	
		Cylindrospermopsin	0.10	U	ug/L	P417829	
		Desmethyl microcystin LR	0.25	U	ug/L	P417829	
		Microcystin HIR	0.25	U	ug/L	P417829	
		Microcystin HtyR	0.25	U	ug/L	P417829	
		Microcystin LA	0.10	U	ug/L	P417829	
		Microcystin LF	0.10	U	ug/L	P417829	
		Microcystin LR	1.2		ug/L	P417829	
		Microcystin LW	0.25	U	ug/L	P417829	
		Microcystin LY	0.10	U	ug/L	P417829	
		Microcystin RR	0.73		ug/L	P417829	
		Microcystin WR	0.50	U	ug/L	P417829	
		Microcystin YR	0.28	I	ug/L	P417829	
		Nodularin-R	0.10	U	ug/L	P417829	
2347663	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P418157	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P418197	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P418107	
	EPA 365.1 Rev. 2.0	Total-P	0.051		mg P/L	P417982	
2347665	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P417910	

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418157

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418138

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418197

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P418107

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417982

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P417910

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P417829

Component	Result	Code	Units
Cylindrospermopsin	0.10	U	ug/L
Desmethyl microcystin LR	0.25	U	ug/L
Microcystin H1LR	0.25	U	ug/L
Microcystin HtyR	0.25	U	ug/L
Microcystin LA	0.10	U	ug/L
Microcystin LF	0.10	U	ug/L
Microcystin LR	0.25	U	ug/L
Microcystin LW	0.25	U	ug/L
Microcystin LY	0.10	U	ug/L
Microcystin RR	0.10	U	ug/L
Microcystin WR	0.50	U	ug/L
Microcystin YR	0.25	U	ug/L
Nodularin-R	0.10	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B
Batch ID: P417830

Component	Result	Code	Units
Anatoxin-a	0.25	U	ug/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418157

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418138

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	106		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418197

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P418107

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	97.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417982

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P417910

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.3		P	90 - 110

Reference Method: EPA 8321B
Batch ID: P417829

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Cylindrospermopsin	87.4		P	40 - 150
Desmethyl microcystin LR	123		P	40 - 150
Microcystin H1R	116		P	40 - 150
Microcystin HtyR	116		P	40 - 150
Microcystin LA	141		P	40 - 150
Microcystin LF	141		P	40 - 150
Microcystin LR	122		P	40 - 150
Microcystin LW	147		P	40 - 150
Microcystin LY	138		P	40 - 150
Microcystin RR	101		P	40 - 150
Microcystin WR	137		P	40 - 150
Microcystin YR	112		P	40 - 150
Nodularin-R	85.8		P	40 - 150

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P417830

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Anatoxin-a	122		P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418157

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2347934	Ammonia-N	99.0	98.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418138

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2345343	Kjeldahl Nitrogen	101	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418197

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2347815	Kjeldahl Nitrogen	107	94.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P418107

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346527	NO2NO3-N	96.4	96.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417982

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2347449	Total-P	105	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P417910

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2347664	O-Phosphate-P	97.9	99.5	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P417829

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346831	Cylindrospermopsin	79.8	73.8	P/P	40 - 150
2346831	Desmethyl microcystin LR	121	124	P/P	40 - 150
2346831	Microcystin H1R	115	117	P/P	40 - 150
2346831	Microcystin HtyR	121	121	P/P	40 - 150
2346831	Microcystin LA	118	115	P/P	40 - 150
2346831	Microcystin LF	111	122	P/P	40 - 150
2346831	Microcystin LR	119	124	P/P	40 - 150
2346831	Microcystin LW	140	147	P/P	40 - 150
2346831	Microcystin LY	131	119	P/P	40 - 150
2346831	Microcystin RR	89.3	93.4	P/P	40 - 150
2346831	Microcystin WR	149	147	P/P	40 - 150
2346831	Microcystin YR	107	110	P/P	40 - 150
2346831	Nodularin-R	87.5	88.4	P/P	40 - 150

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P417830

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346831	Anatoxin-a	121	117	P/P	40 - 150

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418157

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2347934	Ammonia-N	0.870	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418138

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2345343	Kjeldahl Nitrogen	1.12	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418197

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2347815	Kjeldahl Nitrogen	6.95	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P418107

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346527	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417982

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2347449	Total-P	1.33	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P417910

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2347664	O-Phosphate-P	1.62	Spike	P	0 - 20

Reference Method: EPA 8321B
Batch ID: P417829

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346831	Cylindrospermopsin	7.77	Spike	P	0 - 30
2346831	Desmethyl microcystin LR	2.40	Spike	P	0 - 30
2346831	Microcystin HtIR	1.97	Spike	P	0 - 30
2346831	Microcystin HtyR	0.446	Spike	P	0 - 30
2346831	Microcystin LA	2.14	Spike	P	0 - 30
2346831	Microcystin LF	9.18	Spike	P	0 - 30
2346831	Microcystin LR	3.79	Spike	P	0 - 30
2346831	Microcystin LW	5.01	Spike	P	0 - 30
2346831	Microcystin LY	9.22	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P417829

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346831	Microcystin RR	4.57	Spike	P	0 - 30
2346831	Microcystin WR	1.32	Spike	P	0 - 30
2346831	Microcystin YR	2.87	Spike	P	0 - 30
2346831	Nodularin-R	1.02	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P417830

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346831	Anatoxin-a	2.99	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2347628

Field Sample ID: HABOC0007_CypressLake1

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	101	P	30 - 160
EPA 8321B	Neosaxitoxin-15N7	91.8	P	30 - 160

Lab Sample ID: 2347629

Field Sample ID: HABOC0009_LakeSpeer1

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	116	P	30 - 160
EPA 8321B	Neosaxitoxin-15N7	96.5	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A114010

Included Lab Sample IDs: 2347664, 2347665

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	99.9	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A114013

Included Lab Sample IDs: 2347628, 2347629

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Cylindrospermopsin	89.4	87.8	P/P	50 - 160
Desmethyl microcystin LR	117	112	P/P	50 - 160
Microcystin HILR	111	110	P/P	50 - 160
Microcystin HtyR	127	124	P/P	50 - 160
Microcystin LA	128	130	P/P	50 - 160
Microcystin LF	65.1	73.1	P/P	50 - 160
Microcystin LR	116	115	P/P	50 - 160
Microcystin LW	137	149	P/P	50 - 160
Microcystin LY	137	118	P/P	50 - 160
Microcystin RR	101	94.9	P/P	50 - 160
Microcystin WR	157	144	P/P	50 - 160
Microcystin YR	107	106	P/P	50 - 160
Nodularin-R	87.7	86.1	P/P	50 - 160

Reference Method: EPA 8321B

Run ID: A114028

Included Lab Sample IDs: 2347628, 2347629

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Anatoxin-a	117	114	P/P	50 - 160

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A114058

Included Lab Sample IDs: 2347663

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	91.5	94.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A114065

Included Lab Sample IDs: 2347662

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	96.8	97.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A114090

Included Lab Sample IDs: 2347662, 2347663

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	97.0	96.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A114127

Included Lab Sample IDs: 2347662, 2347663

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.7	99.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A114161

Included Lab Sample IDs: 2347662

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.6	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A114226

Included Lab Sample IDs: 2347663

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	97.8	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision LCS	SMP	MS
EPA 350.1 Rev. 2.0	Ammonia-N	95.8	99.0	98.0			0.870
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	106	101	103			1.12
	Kjeldahl Nitrogen	105	107	94.4			6.95
EPA 353.2 Rev. 2.0	NO2NO3-N	97.0	96.4	96.4			0.0
EPA 365.1 Rev. 2.0	Total-P	102	105	103			1.33
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.3	97.9	99.5			1.62
EPA 8321B	Anatoxin-a	122	121	117			2.99
	Cylindrospermopsin	87.4	79.8	73.8			7.77
	Desmethyl microcystin LR	123	121	124			2.40
	Microcystin HiLR	116	115	117			1.97
	Microcystin HtyR	116	121	121			0.446
	Microcystin LA	141	118	115			2.14
	Microcystin LF	141	111	122			9.18
	Microcystin LR	122	119	124			3.79
	Microcystin LW	147	140	147			5.01
	Microcystin LY	138	131	119			9.22
	Microcystin RR	101	89.3	93.4			4.57
	Microcystin WR	137	149	147			1.32
	Microcystin YR	112	107	110			2.87
	Nodularin-R	85.8	87.5	88.4			1.02

Reference Method Descriptions

Method	Description	Associated Samples
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2347662, 2347663
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2347662, 2347663
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2347662, 2347663
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2347662, 2347663
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2347664, 2347665
EPA 8321B	Microcystins in water matrices by HPLC/MS/MS	2347628, 2347629
EPA 8321B	Saxitoxin in water matrices by HPLC/MS/MS	2347628, 2347629

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	08/10/2022			08/16/2022 11:55	Ping Hua	2347662
	08/10/2022			08/16/2022 11:56	Ping Hua	2347663
EPA 351.2 Rev. 2.0	08/10/2022	08/16/2022 13:41	Samantha L Bates	08/17/2022 11:32	Alexandra J Mattheus	2347662
	08/10/2022	08/17/2022 14:23	Samantha L Bates	08/19/2022 15:15	Alexandra J Mattheus	2347663
EPA 353.2 Rev. 2.0	08/10/2022			08/15/2022 09:53	Beverly Y. Sanders	2347662
	08/10/2022			08/15/2022 09:54	Beverly Y. Sanders	2347663
EPA 365.1 Rev. 2.0	08/10/2022	08/11/2022 09:30	Simonne.V. Calhoun	08/12/2022 11:49	James L Waggerby	2347663
	08/10/2022	08/11/2022 09:30	Simonne.V. Calhoun	08/12/2022 13:58	Simonne.V. Calhoun	2347662
EPA 365.1 Rev. 2.0 dissolved	08/10/2022			08/10/2022 14:26	Elise M. Gillis	2347664
	08/10/2022			08/10/2022 14:29	Elise M. Gillis	2347665
EPA 8321B	08/10/2022	08/10/2022 13:00	Umesh Chiluwal	08/10/2022 22:03	Umesh Chiluwal	2347628
	08/10/2022	08/10/2022 13:00	Umesh Chiluwal	08/10/2022 22:20	Umesh Chiluwal	2347629
	08/10/2022	08/11/2022 09:00	Umesh Chiluwal	08/11/2022 15:29	Umesh Chiluwal	2347628
	08/10/2022	08/11/2022 09:00	Umesh Chiluwal	08/11/2022 15:43	Umesh Chiluwal	2347629