

Chemical Analysis Report

WQAP-2022-02-11-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Algal Bloom Response - CEROC**
Request ID: **RQ-2022-01-31-10**
Customer: **WQAP**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
3319 Maguire Blvd. Suite 232
Orlando, FL 32803
Attn: Matt Bearden

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 08-MAR-2022 09:47

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Copeland - SE Corner

Collection Date/Time: 02/10/2022 11:13

Field ID: HABCE0093_Lake Copeland1

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2305206	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P409801	
		Cylindrospermopsin	0.10	U	ug/L	P409801	
		Desmethyl microcystin LR	0.25	U	ug/L	P409801	
		Microcystin HilR**	0.25	U	ug/L	P409801	
		Microcystin HtyR**	0.25	U	ug/L	P409801	
		Microcystin LA	0.10	U	ug/L	P409801	
		Microcystin LF	0.10	U	ug/L	P409801	
		Microcystin LR	0.25	U	ug/L	P409801	
		Microcystin LW	0.25	U	ug/L	P409801	
		Microcystin LY	0.10	U	ug/L	P409801	
		Microcystin RR	0.10	U	ug/L	P409801	
		Microcystin WR	0.50	U	ug/L	P409801	
		Microcystin YR	0.25	U	ug/L	P409801	
		Nodularin-R**	0.10	U	ug/L	P409801	
2305210	EPA 350.1 Rev. 2.0	Ammonia-N	0.035		mg N/L	P410148	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.78		mg N/L	P409787	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.012		mg N/L	P410081	
	EPA 365.1 Rev. 2.0	Total-P	0.032		mg P/L	P409832	
2305214	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409749	

Sample Location: Lake Formosa - SW Park

Collection Date/Time: 02/10/2022 11:38

Field ID: HABCE0107_Lake Formosa1

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2305207	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P409801	
		Cylindrospermopsin	0.10	U	ug/L	P409801	
		Desmethyl microcystin LR	0.25	U	ug/L	P409801	
		Microcystin HilR**	0.25	U	ug/L	P409801	
		Microcystin HtyR**	0.25	U	ug/L	P409801	
		Microcystin LA	0.10	U	ug/L	P409801	
		Microcystin LF	0.10	U	ug/L	P409801	
		Microcystin LR	0.25	U	ug/L	P409801	
		Microcystin LW	0.25	U	ug/L	P409801	
		Microcystin LY	0.10	U	ug/L	P409801	
		Microcystin RR	0.10	U	ug/L	P409801	
		Microcystin WR	0.50	U	ug/L	P409801	
		Microcystin YR	0.25	U	ug/L	P409801	
		Nodularin-R**	0.10	U	ug/L	P409801	
2305211	EPA 350.1 Rev. 2.0	Ammonia-N	0.028		mg N/L	P410148	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.47		mg N/L	P409787	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.17		mg N/L	P409846	
	EPA 365.1 Rev. 2.0	Total-P	0.024		mg P/L	P409832	
2305215	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409749	

Sample Location: Lake Formosa - Pedestrian Bridge

Collection Date/Time: 02/10/2022 12:11

Field ID: HABCE0108_LakeFormosa2

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2305208	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P409801	
		Cylindrospermopsin	0.10	U	ug/L	P409801	
		Desmethyl microcystin LR	0.25	U	ug/L	P409801	
		Microcystin HILR**	0.25	U	ug/L	P409801	
		Microcystin HtyR**	0.25	U	ug/L	P409801	
		Microcystin LA	0.10	U	ug/L	P409801	
		Microcystin LF	0.10	U	ug/L	P409801	
		Microcystin LR	0.25	U	ug/L	P409801	
		Microcystin LW	0.25	U	ug/L	P409801	
		Microcystin LY	0.10	U	ug/L	P409801	
		Microcystin RR	0.10	U	ug/L	P409801	
		Microcystin WR	0.50	U	ug/L	P409801	
		Microcystin YR	0.25	U	ug/L	P409801	
		Nodularin-R**	0.10	U	ug/L	P409801	
2305212	EPA 350.1 Rev. 2.0	Ammonia-N	0.019		mg N/L	P410148	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.46		mg N/L	P409823	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.071		mg N/L	P409846	
	EPA 365.1 Rev. 2.0	Total-P	0.026		mg P/L	P409832	
2305216	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409749	

Sample Location: Lake Chelton - between Forrest Rd and Glencoe Rd

Collection Date/Time: 02/10/2022 13:19

Field ID: HABCE0104_LakeChelton

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2305209	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P409801	
		Cylindrospermopsin	0.10	U	ug/L	P409801	
		Desmethyl microcystin LR	0.25	U	ug/L	P409801	
		Microcystin HilR**	0.25	U	ug/L	P409801	
		Microcystin HtyR**	0.25	U	ug/L	P409801	
		Microcystin LA	0.26	I	ug/L	P409801	
		Microcystin LF	0.10	U	ug/L	P409801	
		Microcystin LR	0.25	U	ug/L	P409801	
		Microcystin LW	0.25	U	ug/L	P409801	
		Microcystin LY	0.10	U	ug/L	P409801	
		Microcystin RR	0.10	U	ug/L	P409801	
		Microcystin WR	0.50	U	ug/L	P409801	
		Microcystin YR	0.25	U	ug/L	P409801	
		Nodularin-R**	0.10	U	ug/L	P409801	
2305213	EPA 350.1 Rev. 2.0	Ammonia-N	0.014		mg N/L	P410311	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.29		mg N/L	P409823	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.021		mg N/L	P409846	
	EPA 365.1 Rev. 2.0	Total-P	0.011		mg P/L	P409832	
2305217	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409749	

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410148

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410311

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409787

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409823

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409846

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410081

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409832

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409749

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P409801

Component	Result	Code	Units
Anatoxin-a	0.25	U	ug/L
Cylindrospermopsin	0.10	U	ug/L
Desmethyl microcystin LR	0.25	U	ug/L
Microcystin HiLR	0.25	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B

Batch ID: P409801

Component	Result	Code	Units
Microcystin HtyR	0.25	U	ug/L
Microcystin LA	0.10	U	ug/L
Microcystin LF	0.10	U	ug/L
Microcystin LR	0.25	U	ug/L
Microcystin LW	0.25	U	ug/L
Microcystin LY	0.10	U	ug/L
Microcystin RR	0.10	U	ug/L
Microcystin WR	0.50	U	ug/L
Microcystin YR	0.25	U	ug/L
Nodularin-R	0.10	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410148

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410311

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409787

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	96.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409823

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.7		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409846

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410081

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409832

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409749

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.9		P	90 - 110

Reference Method: EPA 8321B
Batch ID: P409801

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Anatoxin-a	91.7		P	40 - 150
Cylindrospermopsin	90.0		P	40 - 150
Desmethyl microcystin LR	80.9		P	40 - 150
Microcystin HiLR	85.6		P	40 - 150

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P409801

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Microcystin HtyR	85.7		P	40 - 150
Microcystin LA	89.8		P	40 - 150
Microcystin LF	78.2		P	40 - 150
Microcystin LR	82.5		P	40 - 150
Microcystin LW	72.2		P	40 - 150
Microcystin LY	76.5		P	40 - 150
Microcystin RR	91.5		P	40 - 150
Microcystin WR	76.5		P	40 - 150
Microcystin YR	75.6		P	40 - 150
Nodularin-R	83.8		P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410148

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305106	Ammonia-N	98.4	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410311

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305213	Ammonia-N	97.0	99.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409787

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305358	Kjeldahl Nitrogen	103	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409823

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305190	Kjeldahl Nitrogen	98.6	96.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409846

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305211	NO2NO3-N	103	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410081

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305143	NO2NO3-N	100	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409832

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305270	Total-P	105	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409749

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305216	O-Phosphate-P	98.7	100	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P409801

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305206	Anatoxin-a	111	114	P/P	40 - 150
2305206	Cylindrospermopsin	103	102	P/P	40 - 150
2305206	Desmethyl microcystin LR	106	104	P/P	40 - 150
2305206	Microcystin HiLR	101	99.4	P/P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P409801

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2305206	Microcystin HtyR	98.5	94.8	P/P	40 - 150
2305206	Microcystin LA	105	99.1	P/P	40 - 150
2305206	Microcystin LF	95.6	87.7	P/P	40 - 150
2305206	Microcystin LR	90.7	97.5	P/P	40 - 150
2305206	Microcystin LW	87.5	85.9	P/P	40 - 150
2305206	Microcystin LY	91.6	84.8	P/P	40 - 150
2305206	Microcystin RR	101	102	P/P	40 - 150
2305206	Microcystin WR	96.1	91.1	P/P	40 - 150
2305206	Microcystin YR	89.9	88.7	P/P	40 - 150
2305206	Nodularin-R	96.5	97.5	P/P	40 - 150

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410148

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305106	Ammonia-N	1.87	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410311

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305213	Ammonia-N	1.89	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409787

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305358	Kjeldahl Nitrogen	0.275	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409823

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305190	Kjeldahl Nitrogen	1.67	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409846

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305211	NO2NO3-N	1.06	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410081

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305143	NO2NO3-N	0.990	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409832

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305270	Total-P	1.72	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409749

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305216	O-Phosphate-P	1.61	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8321B

Batch ID: P409801

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2305206	Anatoxin-a	3.29	Spike	P	0 - 30
2305206	Cylindrospermopsin	1.24	Spike	P	0 - 30
2305206	Desmethyl microcystin LR	2.01	Spike	P	0 - 30
2305206	Microcystin HiLR	1.42	Spike	P	0 - 30
2305206	Microcystin HtyR	3.82	Spike	P	0 - 30
2305206	Microcystin LA	5.91	Spike	P	0 - 30
2305206	Microcystin LF	8.66	Spike	P	0 - 30
2305206	Microcystin LR	7.23	Spike	P	0 - 30
2305206	Microcystin LW	1.86	Spike	P	0 - 30
2305206	Microcystin LY	7.76	Spike	P	0 - 30
2305206	Microcystin RR	0.526	Spike	P	0 - 30
2305206	Microcystin WR	5.34	Spike	P	0 - 30
2305206	Microcystin YR	1.35	Spike	P	0 - 30
2305206	Nodularin-R	1.06	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2305206

Field Sample ID: HABCE0093_Lake Copeland1

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	78.3	P	30 - 160

Lab Sample ID: 2305207

Field Sample ID: HABCE0107_Lake Formosa1

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	79.3	P	30 - 160

Lab Sample ID: 2305208

Field Sample ID: HABCE0108_LakeFormosa2

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	72.9	P	30 - 160

Lab Sample ID: 2305209

Field Sample ID: HABCE0104_LakeChelton

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	64.9	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A110364

Included Lab Sample IDs: 2305214, 2305215, 2305216, 2305217

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.4	99.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A110411

Included Lab Sample IDs: 2305211, 2305212, 2305213

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A110453

Included Lab Sample IDs: 2305210, 2305211

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.3	98.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A110467

Included Lab Sample IDs: 2305212, 2305213

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	97.0	99.2	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A110473

Included Lab Sample IDs: 2305206, 2305207, 2305208, 2305209

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Anatoxin-a	104	111	P/P	50 - 160
Cylindrospermopsin	103	106	P/P	60 - 160
Desmethyl microcystin LR	91.5	92.1	P/P	60 - 160
Microcystin H1R	95.4	86.0	P/P	60 - 160
Microcystin HtyR	96.2	88.1	P/P	60 - 160
Microcystin LA	106	103	P/P	50 - 160
Microcystin LF	97.8	93.5	P/P	60 - 160
Microcystin LR	88.7	84.9	P/P	60 - 160
Microcystin LW	95.9	84.9	P/P	60 - 160
Microcystin LY	89.5	85.0	P/P	60 - 160
Microcystin RR	103	99.8	P/P	60 - 160
Microcystin WR	97.6	93.4	P/P	60 - 160
Microcystin YR	91.8	81.2	P/P	60 - 160
Nodularin-R	93.4	96.9	P/P	60 - 160

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A110502

Included Lab Sample IDs: 2305210

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A110521

Included Lab Sample IDs: 2305210, 2305211, 2305212, 2305213

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.5	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A110530

Included Lab Sample IDs: 2305210, 2305211, 2305212

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.5	97.6	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A110602

Included Lab Sample IDs: 2305213

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.1	98.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 350.1 Rev. 2.0	Ammonia-N	100	98.4	101			1.87
	Ammonia-N	98.0	97.0	99.4			1.89
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	96.8	103	104			0.275
	Kjeldahl Nitrogen	97.7	98.6	96.2			1.67
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	104	103	101			1.06
	NO ₂ NO ₃ -N	98.5	100	102			0.990
EPA 365.1 Rev. 2.0	Total-P	100	105	103			1.72
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.9	98.7	100			1.61
EPA 8321B	Anatoxin-a	91.7	111	114			3.29
	Cylindrospermopsin	90.0	103	102			1.24
	Desmethyl microcystin LR	80.9	106	104			2.01
	Microcystin HiLR	85.6	101	99.4			1.42
	Microcystin HtyR	85.7	98.5	94.8			3.82
	Microcystin LA	89.8	105	99.1			5.91
	Microcystin LF	78.2	95.6	87.7			8.66
	Microcystin LR	82.5	90.7	97.5			7.23
	Microcystin LW	72.2	87.5	85.9			1.86
	Microcystin LY	76.5	91.6	84.8			7.76
	Microcystin RR	91.5	101	102			0.526
	Microcystin WR	76.5	96.1	91.1			5.34
	Microcystin YR	75.6	89.9	88.7			1.35
	Nodularin-R	83.8	96.5	97.5			1.06

Reference Method Descriptions

Method	Description	Associated Samples
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2305210, 2305211, 2305212, 2305213
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2305210, 2305211, 2305212, 2305213
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2305210, 2305211, 2305212, 2305213
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2305210, 2305211, 2305212, 2305213
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2305214, 2305215, 2305216, 2305217
EPA 8321B	Microcystins in water matrices by HPLC/MS/MS	2305206, 2305207, 2305208, 2305209

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	02/11/2022			02/21/2022 12:26	Ping Hua	2305210
	02/11/2022			02/21/2022 12:34	Ping Hua	2305211
	02/11/2022			02/21/2022 12:36	Ping Hua	2305212
	02/11/2022			02/24/2022 11:37	Ping Hua	2305213
EPA 351.2 Rev. 2.0	02/11/2022	02/15/2022 13:57	Samantha L Bates	02/16/2022 14:42	Alexandra J Mattheus	2305210
	02/11/2022	02/15/2022 13:57	Samantha L Bates	02/16/2022 14:44	Alexandra J Mattheus	2305211
	02/11/2022	02/16/2022 13:48	Samantha L Bates	02/17/2022 11:57	Alexandra J Mattheus	2305212
	02/11/2022	02/16/2022 13:48	Samantha L Bates	02/17/2022 11:58	Alexandra J Mattheus	2305213
EPA 353.2 Rev. 2.0	02/11/2022			02/15/2022 11:10	Beverly Y. Sanders	2305211
	02/11/2022			02/15/2022 11:18	Beverly Y. Sanders	2305212
	02/11/2022			02/15/2022 11:19	Beverly Y. Sanders	2305213
	02/11/2022			02/18/2022 10:53	Beverly Y. Sanders	2305210
EPA 365.1 Rev. 2.0	02/11/2022	02/16/2022 14:20	Simonne.V. Calhoun	02/21/2022 14:46	Sarah A Nixon	2305210
	02/11/2022	02/16/2022 14:20	Simonne.V. Calhoun	02/21/2022 14:47	Sarah A Nixon	2305211
	02/11/2022	02/16/2022 14:20	Simonne.V. Calhoun	02/21/2022 14:48	Sarah A Nixon	2305212
	02/11/2022	02/16/2022 14:20	Simonne.V. Calhoun	02/21/2022 14:49	Sarah A Nixon	2305213
EPA 365.1 Rev. 2.0 dissolved	02/11/2022			02/11/2022 14:31	James L Waggerby	2305216
	02/11/2022			02/11/2022 14:34	James L Waggerby	2305214, 2305215
	02/11/2022			02/11/2022 14:35	James L Waggerby	2305217
EPA 8321B	02/11/2022	02/17/2022 09:00	Manjita Shrestha	02/17/2022 12:04	Manjita Shrestha	2305206
	02/11/2022	02/17/2022 09:00	Manjita Shrestha	02/17/2022 12:20	Manjita Shrestha	2305207
	02/11/2022	02/17/2022 09:00	Manjita Shrestha	02/17/2022 12:37	Manjita Shrestha	2305208
	02/11/2022	02/17/2022 09:00	Manjita Shrestha	02/17/2022 12:54	Manjita Shrestha	2305209