

Chemical Analysis Report

WQAP-2022-08-25-03

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Algal Bloom Response -SOROC**
Request ID: **RQ-2022-02-28-44**
Customer: **WQAP**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
2295 Victoria AVE
Suite 364
Fort Myers, FL 33901
Attn: Kirby Wolfe

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 13-SEP-2022 14:39

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light pink rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients, Pesticides and Priority Organic Pollutants.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Clay

Collection Date/Time: 08/24/2022 14:58

Field ID: HAB-SD-082422-1505

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2352292	EPA 8270E	Acenaphthene	0.026	U	ug/L	P418679	
		Acenaphthylene	0.026	U	ug/L	P418679	
		Acetophenone	0.21	U	ug/L	P418679	
		2-Acetylaminofluorene	1.0	U	ug/L	P418679	
		4-Aminobiphenyl	4.2	U	ug/L	P418679	
		Aniline	1.0	U	ug/L	P418679	
		Anthracene	0.052	U	ug/L	P418679	
		Azobenzene/1,2-Diphenylhydrazine**	0.052	U	ug/L	P418679	
		Benzidine	10	U	ug/L	P418679	
		Benzo(a)anthracene	0.026	U	ug/L	P418679	
		Benzo(a)pyrene	0.026	U	ug/L	P418679	
		Benzo(b)fluoranthene	0.026	U	ug/L	P418679	
		Benzo(k)fluoranthene	0.026	U	ug/L	P418679	
		Benzo(g,h,i)perylene	0.026	U	ug/L	P418679	
		Benzyl alcohol	0.21	U	ug/L	P418679	
		Bis(2-chloroethoxy)methane	0.052	U	ug/L	P418679	
		Bis(2-chloroethyl)ether	0.052	U	ug/L	P418679	
		Bis(2-chloroisopropyl)ether	0.052	U	ug/L	P418679	
		Bis(2-ethylhexyl)phthalate	5.2	U	ug/L	P418679	
		Butyl benzyl phthalate	1.0	U	ug/L	P418679	
		4-Bromophenyl phenyl ether	0.052	U	ug/L	P418679	
		2-Chloronaphthalene	0.052	U	ug/L	P418679	
		4-Chlorophenyl phenyl ether	0.052	U	ug/L	P418679	
		Carbazole	0.052	U	ug/L	P418679	
		Chrysene	0.026	U	ug/L	P418679	
		m,p-Cresols	0.052	U	ug/L	P418679	
		o-Cresol	0.052	U	ug/L	P418679	
		Di-n-butyl phthalate	2.1	U	ug/L	P418679	
		Di-n-octyl phthalate	0.052	U	ug/L	P418679	
		Dibenzo(a,h)anthracene	0.026	U	ug/L	P418679	
		Dibenzofuran	0.052	U	ug/L	P418679	
		3,3'-Dichlorobenzidine	10	U	ug/L	P418679	
		Diethyl phthalate	2.1	U	ug/L	P418679	
		Dimethyl phthalate	0.21	U	ug/L	P418679	
		Dimethylaminoazobenzene	0.052	U	ug/L	P418679	
		7,12-Dimethylbenz(a)anthracene	0.10	U	ug/L	P418679	
		1,3-Dinitrobenzene	0.21	U	ug/L	P418679	
		2,4-Dinitrotoluene	0.052	U	ug/L	P418679	
		2,6-Dinitrotoluene	0.052	U	ug/L	P418679	
		Dinoseb**	4.2	U	ug/L	P418679	
		Ethyl methanesulfonate	1.0	U	ug/L	P418679	
		Fluoranthene	0.052	U	ug/L	P418679	
		Fluorene	0.026	U	ug/L	P418679	
		Hexachlorobenzene	0.052	U	ug/L	P418679	

Field ID: HAB-SD-082422-1505

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2352292	EPA 8270E	Hexachlorobutadiene	0.052	U	ug/L	P418679	
		Hexachlorocyclopentadiene	0.052	U	ug/L	P418679	
		Hexachloroethane	0.052	U	ug/L	P418679	
		Hexachloropropene	0.052	U	ug/L	P418679	
		Indeno(1,2,3-cd)pyrene	0.026	U	ug/L	P418679	
		Isophorone	0.052	U	ug/L	P418679	
		Isosafrole	0.052	U	ug/L	P418679	
		3-Methylcholanthrene	0.10	U	ug/L	P418679	
		2-Methylnaphthalene	0.10	U	ug/L	P418679	
		Naphthalene	0.10	U	ug/L	P418679	
		1-Naphthylamine	10	U	ug/L	P418679	
		2-Naphthylamine	10	U	ug/L	P418679	
		2-Nitroaniline	0.052	U	ug/L	P418679	
		Nitrobenzene	0.052	U	ug/L	P418679	
		5-Nitro-o-toluidine	0.10	U	ug/L	P418679	
		N-Nitrosodi-n-butylamine	0.052	U	ug/L	P418679	
		N-Nitrosodiethylamine	1.0	U	ug/L	P418679	
		N-Nitrosodimethylamine	2.1	U	ug/L	P418679	
		N-Nitrosodi-n-propylamine	0.052	U	ug/L	P418679	
		N-Nitrosomethylethylamine	2.1	U	ug/L	P418679	
		N-Nitrosomorpholine	0.052	U	ug/L	P418679	
		N-Nitrosopiperidine	0.052	U	ug/L	P418679	
		N-Nitrosopyrrolidine	0.10	U	ug/L	P418679	
		Pentachlorobenzene	0.052	U	ug/L	P418679	
		Pentachloroethane	0.052	U	ug/L	P418679	
		Pentachloronitrobenzene	0.052	U	ug/L	P418679	
		Phenacetin	1.0	U	ug/L	P418679	
		Phenanthrene	0.10	U	ug/L	P418679	
		2-Picoline	1.0	U	ug/L	P418679	
		Pyrene	0.10	U	ug/L	P418679	
		Pyridine	4.2	U	ug/L	P418679	
		Safrole	0.052	U	ug/L	P418679	
		1,2,4,5-Tetrachlorobenzene	0.052	U	ug/L	P418679	
		o-Toluidine	0.10	U	ug/L	P418679	
		1,2,4-Trichlorobenzene	0.052	U	ug/L	P418679	
		1,3,5-Trinitrobenzene	0.21	U	ug/L	P418679	
		4-Chloro-3-methylphenol	0.052	U	ug/L	P418679	
		2-Chlorophenol	0.052	U	ug/L	P418679	
		2,4-Dichlorophenol	0.052	U	ug/L	P418679	
		2,6-Dichlorophenol	0.052	U	ug/L	P418679	
		2,4-Dimethylphenol	0.052	U	ug/L	P418679	
		2,4-Dinitrophenol	10	U	ug/L	P418679	
		2-Methyl-4,6-dinitrophenol	3.1	U	ug/L	P418679	
		2-Nitrophenol	0.052	U	ug/L	P418679	
		4-Nitrophenol	10	U	ug/L	P418679	

Field ID: HAB-SD-082422-1505

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2352292	EPA 8270E	Pentachlorophenol	0.52	U	ug/L	P418679	
		Phenol	0.052	U	ug/L	P418679	
		2,3,4,6-Tetrachlorophenol	0.10	U	ug/L	P418679	
		2,4,5-Trichlorophenol	0.052	U	ug/L	P418679	
		2,4,6-Trichlorophenol	0.052	U	ug/L	P418679	
		1-Methylnaphthalene	0.10	U	ug/L	P418679	
2352293	FL-PRO 2018	TRPH	0.67	I J	mg/L	P418756	MS, RPD
2352294	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P418570	
		Cylindrospermopsin	0.10	U	ug/L	P418569	
		Desmethyl microcystin LR	0.25	U	ug/L	P418569	
		Microcystin HilR	0.25	U	ug/L	P418569	
		Microcystin HtyR	0.25	U	ug/L	P418569	
		Microcystin LA	0.10	U	ug/L	P418569	
		Microcystin LF	0.10	U	ug/L	P418569	
		Microcystin LR	0.25	U	ug/L	P418569	
		Microcystin LW	0.25	U	ug/L	P418569	
		Microcystin LY	0.10	U	ug/L	P418569	
		Microcystin RR	0.10	U	ug/L	P418569	
		Microcystin WR	0.50	U	ug/L	P418569	
		Microcystin YR	0.25	U	ug/L	P418569	
		Nodularin-R	0.10	U	ug/L	P418569	
2352296	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P418886	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.99		mg N/L	P418823	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P419050	
	EPA 365.1 Rev. 2.0	Total-P	0.018		mg P/L	P419131	
2352297	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P418697	

Ref. Method and Comment:

EPA 8270E: Precision for benzidine is not available due to low recoveries in the duplicate matrix spikes.

FL-PRO 2018: Refer to the Lab Analysis Report for an explanation of QC Codes.

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418886

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418823

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P419050

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P419131

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P418697

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270E
Batch ID: P418679

Component	Result	Code	Units
1-Methylnaphthalene	0.10	U	ug/L
1-Naphthylamine	10	U	ug/L
1,2,4-Trichlorobenzene	0.050	U	ug/L
1,2,4,5-Tetrachlorobenzene	0.050	U	ug/L
1,3-Dinitrobenzene	0.20	U	ug/L
1,3,5-Trinitrobenzene	0.20	U	ug/L
2-Acetylaminofluorene	1.0	U	ug/L
2-Chloronaphthalene	0.050	U	ug/L
2-Chlorophenol	0.050	U	ug/L
2-Methyl-4,6-dinitrophenol	3.0	U	ug/L
2-Methylnaphthalene	0.10	U	ug/L
2-Naphthylamine	10	U	ug/L
2-Nitroaniline	0.050	U	ug/L
2-Nitrophenol	0.050	U	ug/L
2-Picoline	1.0	U	ug/L
2,3,4,6-Tetrachlorophenol	0.10	U	ug/L
2,4-Dichlorophenol	0.050	U	ug/L
2,4-Dimethylphenol	0.050	U	ug/L
2,4-Dinitrophenol	10	U	ug/L
2,4-Dinitrotoluene	0.050	U	ug/L
2,4,5-Trichlorophenol	0.050	U	ug/L
2,4,6-Trichlorophenol	0.050	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P418679

Component	Result	Code	Units
2,6-Dichlorophenol	0.050	U	ug/L
2,6-Dinitrotoluene	0.050	U	ug/L
3-Methylcholanthrene	0.10	U	ug/L
3,3'-Dichlorobenzidine	10	U	ug/L
4-Aminobiphenyl	4.0	U	ug/L
4-Bromophenyl phenyl ether	0.050	U	ug/L
4-Chloro-3-methylphenol	0.050	U	ug/L
4-Chlorophenyl phenyl ether	0.050	U	ug/L
4-Nitrophenol	10	U	ug/L
5-Nitro-o-toluidine	0.10	U	ug/L
7,12-Dimethylbenz(a)anthracene	0.10	U	ug/L
Acenaphthene	0.025	U	ug/L
Acenaphthylene	0.025	U	ug/L
Acetophenone	0.20	U	ug/L
Aniline	1.0	U	ug/L
Anthracene	0.050	U	ug/L
Azobenzene/1,2-Diphenylhydrazine	0.050	U	ug/L
Benazidine	10	U	ug/L
Benzo(a)anthracene	0.025	U	ug/L
Benzo(a)pyrene	0.025	U	ug/L
Benzo(b)fluoranthene	0.025	U	ug/L
Benzo(g,h,i)perylene	0.025	U	ug/L
Benzo(k)fluoranthene	0.025	U	ug/L
Benzyl alcohol	0.20	U	ug/L
Bis(2-chloroethoxy)methane	0.050	U	ug/L
Bis(2-chloroethyl)ether	0.050	U	ug/L
Bis(2-chloroisopropyl)ether	0.050	U	ug/L
Bis(2-ethylhexyl)phthalate	5.0	U	ug/L
Butyl benzyl phthalate	1.0	U	ug/L
Carbazole	0.050	U	ug/L
Chrysene	0.025	U	ug/L
Di-n-butyl phthalate	2.0	U	ug/L
Di-n-octyl phthalate	0.050	U	ug/L
Dibenzo(a,h)anthracene	0.025	U	ug/L
Dibenzofuran	0.050	U	ug/L
Diethyl phthalate	2.0	U	ug/L
Dimethyl phthalate	0.20	U	ug/L
Dimethylaminoazobenzene	0.050	U	ug/L
Dinoseb	4.0	U	ug/L
Ethyl methanesulfonate	1.0	U	ug/L
Fluoranthene	0.050	U	ug/L
Fluorene	0.025	U	ug/L
Hexachlorobenzene	0.050	U	ug/L
Hexachlorobutadiene	0.050	U	ug/L
Hexachlorocyclopentadiene	0.050	U	ug/L
Hexachloroethane	0.050	U	ug/L
Hexachloropropene	0.050	U	ug/L
Indeno(1,2,3-cd)pyrene	0.025	U	ug/L
Isophorone	0.050	U	ug/L
Isosafrole	0.050	U	ug/L
m,p-Cresols	0.050	U	ug/L
N-Nitrosodi-n-butylamine	0.050	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P418679

Component	Result	Code	Units
N-Nitrosodi-n-propylamine	0.050	U	ug/L
N-Nitrosodiethylamine	1.0	U	ug/L
N-Nitrosodimethylamine	2.0	U	ug/L
N-Nitrosomethylethylamine	2.0	U	ug/L
N-Nitrosomorpholine	0.050	U	ug/L
N-Nitrosopiperidine	0.050	U	ug/L
N-Nitrosopyrrolidine	0.10	U	ug/L
Naphthalene	0.10	U	ug/L
Nitrobenzene	0.050	U	ug/L
o-Cresol	0.050	U	ug/L
o-Toluidine	0.10	U	ug/L
Pentachlorobenzene	0.050	U	ug/L
Pentachloroethane	0.050	U	ug/L
Pentachloronitrobenzene	0.050	U	ug/L
Pentachlorophenol	0.50	U	ug/L
Phenacetin	1.0	U	ug/L
Phenanthrene	0.10	U	ug/L
Phenol	0.050	U	ug/L
Pyrene	0.10	U	ug/L
Pyridine	4.0	U	ug/L
Safrrole	0.050	U	ug/L

Reference Method: EPA 8321B

Batch ID: P418569

Component	Result	Code	Units
Cylindrospermopsin	0.10	U	ug/L
Desmethyl microcystin LR	0.25	U	ug/L
Microcystin H1R	0.25	U	ug/L
Microcystin HtyR	0.25	U	ug/L
Microcystin LA	0.10	U	ug/L
Microcystin LF	0.10	U	ug/L
Microcystin LR	0.25	U	ug/L
Microcystin LW	0.25	U	ug/L
Microcystin LY	0.10	U	ug/L
Microcystin RR	0.10	U	ug/L
Microcystin WR	0.50	U	ug/L
Microcystin YR	0.25	U	ug/L
Nodularin-R	0.10	U	ug/L

Reference Method: EPA 8321B

Batch ID: P418570

Component	Result	Code	Units
Anatoxin-a	0.25	U	ug/L

Reference Method: FL-PRO 2018

Batch ID: P418756

Component	Result	Code	Units
TRPH	0.45	U	mg/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418886

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418823

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P419050

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P419131

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P418697

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: EPA 8270E
Batch ID: P418679

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1-Methylnaphthalene	92.7		P	50 - 130
1-Naphthylamine	65.2		P	20 - 130
1,2,4-Trichlorobenzene	84.1		P	50 - 130
1,2,4,5-Tetrachlorobenzene	96.0		P	50 - 130
1,3-Dinitrobenzene	103		P	50 - 130
1,3,5-Trinitrobenzene	124		P	50 - 150
2-Acetylaminofluorene	117		P	50 - 130
2-Chloronaphthalene	91.6		P	50 - 130
2-Chlorophenol	87.2		P	50 - 130
2-Methyl-4,6-dinitrophenol	99.3		P	50 - 150
2-Methylnaphthalene	92.3		P	50 - 130
2-Naphthylamine	66.1		P	20 - 130
2-Nitroaniline	104		P	50 - 130
2-Nitrophenol	92.5		P	50 - 130
2-Picoline	80.5		P	40 - 130
2,3,4,6-Tetrachlorophenol	107		P	50 - 130
2,4-Dichlorophenol	92.1		P	50 - 130
2,4-Dimethylphenol	88.3		P	50 - 130
2,4-Dinitrophenol	41.8		P	30 - 160
2,4-Dinitrotoluene	99.1		P	50 - 130
2,4,5-Trichlorophenol	87.8		P	50 - 130
2,4,6-Trichlorophenol	90.3		P	50 - 130

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P418679

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,6-Dichlorophenol	109		P	50 - 130
2,6-Dinitrotoluene	99.4		P	50 - 130
3-Methylcholanthrene	82.1		P	50 - 130
3,3'-Dichlorobenzidine	195		P	0.0 - 200
4-Aminobiphenyl	98.7		P	30 - 130
4-Bromophenyl phenyl ether	90.3		P	50 - 130
4-Chloro-3-methylphenol	91.1		P	50 - 130
4-Chlorophenyl phenyl ether	98.7		P	50 - 130
4-Nitrophenol	58.8		P	15 - 110
5-Nitro-o-toluidine	120		P	50 - 130
7,12-Dimethylbenz(a)anthracene	106		P	50 - 130
Acenaphthene	96.8		P	50 - 130
Acenaphthylene	95.8		P	50 - 130
Acetophenone	114		P	50 - 130
Aniline	90.5		P	30 - 130
Anthracene	94.3		P	50 - 130
Azobenzene/1,2-Diphenylhydrazine	99.1		P	50 - 130
Benidine	65.0		P	0.0 - 240
Benzo(a)anthracene	97.8		P	50 - 130
Benzo(a)pyrene	92.0		P	50 - 130
Benzo(b)fluoranthene	97.3		P	50 - 130
Benzo(g,h,i)perylene	74.9		P	50 - 130
Benzo(k)fluoranthene	91.7		P	50 - 130
Benzyl alcohol	91.6		P	50 - 130
Bis(2-chloroethoxy)methane	91.9		P	50 - 130
Bis(2-chloroethyl)ether	91.6		P	50 - 160
Bis(2-chloroisopropyl)ether	90.1		P	50 - 130
Bis(2-ethylhexyl)phthalate	116		P	50 - 160
Butyl benzyl phthalate	109		P	50 - 160
Carbazole	118		P	50 - 130
Chrysene	98.2		P	50 - 130
Di-n-butyl phthalate	102		P	50 - 160
Di-n-octyl phthalate	107		P	50 - 130
Dibenzo(a,h)anthracene	80.9		P	50 - 130
Dibenzofuran	96.8		P	50 - 130
Diethyl phthalate	110		P	50 - 130
Dimethyl phthalate	104		P	50 - 130
Dimethylaminoazobenzene	113		P	50 - 130
Dinoseb	126		P	50 - 150
Ethyl methanesulfonate	96.2		P	50 - 130
Fluoranthene	94.5		P	50 - 130
Fluorene	100		P	50 - 130
Hexachlorobenzene	90.9		P	50 - 130
Hexachlorobutadiene	84.5		P	20 - 130
Hexachlorocyclopentadiene	51.9		P	20 - 130
Hexachloroethane	84.0		P	40 - 130
Hexachloropropene	97.0		P	50 - 130
Indeno(1,2,3-cd)pyrene	78.8		P	50 - 130
Isophorone	96.2		P	50 - 130
Isosafrole	108		P	50 - 130
m,p-Cresols	82.5		P	50 - 130
N-Nitrosodi-n-butylamine	107		P	50 - 130

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P418679

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
N-Nitrosodi-n-propylamine	89.1		P	50 - 130
N-Nitrosodiethylamine	104		P	50 - 130
N-Nitrosodimethylamine	92.4		P	30 - 130
N-Nitrosomethylethylamine	104		P	50 - 130
N-Nitrosomorpholine	127		P	50 - 150
N-Nitrosopiperidine	109		P	50 - 130
N-Nitrosopyrrolidine	96.1		P	50 - 130
Naphthalene	90.7		P	50 - 130
Nitrobenzene	95.9		P	50 - 130
o-Cresol	86.7		P	50 - 130
o-Toluidine	111		P	50 - 130
Pentachlorobenzene	99.7		P	50 - 130
Pentachloroethane	95.8		P	50 - 130
Pentachloronitrobenzene	130		P	50 - 130
Pentachlorophenol	91.5		P	50 - 130
Phenacetin	130		P	50 - 130
Phenanthrene	95.2		P	50 - 130
Phenol	54.8		P	15 - 110
Pyrene	98.3		P	50 - 130
Pyridine	70.4		P	20 - 130
Safrrole	106		P	50 - 130

Reference Method: EPA 8321B
Batch ID: P418569

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Cylindrospermopsin	79.8		P	40 - 150
Desmethyl microcystin LR	73.2		P	40 - 150
Microcystin H1R	77.2		P	40 - 150
Microcystin HtyR	78.1		P	40 - 150
Microcystin LA	80.6		P	40 - 150
Microcystin LF	66.4		P	40 - 150
Microcystin LR	76.4		P	40 - 150
Microcystin LW	64.4		P	40 - 150
Microcystin LY	66.1		P	40 - 150
Microcystin RR	77.1		P	40 - 150
Microcystin WR	73.7		P	40 - 150
Microcystin YR	80.0		P	40 - 150
Nodularin-R	80.8		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P418570

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Anatoxin-a	75.5		P	40 - 150

Reference Method: FL-PRO 2018
Batch ID: P418756

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TRPH	90.9		P	66 - 116

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P418886

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352296	Ammonia-N	95.2	95.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P418823

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352296	Kjeldahl Nitrogen	100	96.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P419050

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352331	NO2NO3-N	99.5	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P419131

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352281	Total-P	104	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P418697

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352277	O-Phosphate-P	101	100	P/P	90 - 110

Reference Method: EPA 8270E
Batch ID: P418679

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352292	1-Methylnaphthalene	93.7	89.6	P/P	50 - 130
2352292	1-Naphthylamine	106	87.5	P/P	20 - 130
2352292	1,2,4-Trichlorobenzene	84.9	78.9	P/P	50 - 130
2352292	1,2,4,5-Tetrachlorobenzene	96.4	89.1	P/P	50 - 130
2352292	1,3-Dinitrobenzene	97.2	101	P/P	50 - 130
2352292	1,3,5-Trinitrobenzene	112	117	P/P	50 - 150
2352292	2-Acetylaminofluorene	115	121	P/P	50 - 130
2352292	2-Chloronaphthalene	91.8	87.4	P/P	50 - 130
2352292	2-Chlorophenol	86.9	81.1	P/P	50 - 130
2352292	2-Methyl-4,6-dinitrophenol	115	116	P/P	50 - 150
2352292	2-Methylnaphthalene	92.3	87.5	P/P	50 - 130
2352292	2-Naphthylamine	75.7	71.5	P/P	20 - 130
2352292	2-Nitroaniline	103	98.0	P/P	50 - 130
2352292	2-Nitrophenol	90.8	87.7	P/P	50 - 130
2352292	2-Picoline	99.4	92.6	P/P	40 - 130
2352292	2,3,4,6-Tetrachlorophenol	104	107	P/P	50 - 130
2352292	2,4-Dichlorophenol	92.8	88.7	P/P	50 - 130
2352292	2,4-Dimethylphenol	91.5	88.0	P/P	50 - 130
2352292	2,4-Dinitrophenol	74.4	72.6	P/P	30 - 160
2352292	2,4-Dinitrotoluene	99.8	96.2	P/P	50 - 130
2352292	2,4,5-Trichlorophenol	90.0	88.9	P/P	50 - 130
2352292	2,4,6-Trichlorophenol	90.6	88.3	P/P	50 - 130

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P418679

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352292	2,6-Dichlorophenol	109	103	P/P	50 - 130
2352292	2,6-Dinitrotoluene	100	95.8	P/P	50 - 130
2352292	3-Methylcholanthrene	92.4	97.2	P/P	50 - 130
2352292	3,3'-Dichlorobenzidine	180	169	P/P	0.0 - 200
2352292	4-Aminobiphenyl	103	98.1	P/P	30 - 130
2352292	4-Bromophenyl phenyl ether	88.2	88.1	P/P	50 - 130
2352292	4-Chloro-3-methylphenol	92.1	89.4	P/P	50 - 130
2352292	4-Chlorophenyl phenyl ether	100	94.6	P/P	50 - 130
2352292	4-Nitrophenol	53.7	53.0	P/P	15 - 110
2352292	5-Nitro-o-toluidine	114	121	P/P	50 - 130
2352292	7,12-Dimethylbenz(a)anthracene	85.6	97.4	P/P	50 - 130
2352292	Acenaphthene	97.8	93.4	P/P	50 - 130
2352292	Acenaphthylene	96.1	90.8	P/P	50 - 130
2352292	Acetophenone	114	104	P/P	50 - 130
2352292	Aniline	89.7	85.1	P/P	30 - 130
2352292	Anthracene	91.5	91.3	P/P	50 - 130
2352292	Azobenzene/1,2-Diphenylhydrazine	98.7	100	P/P	50 - 130
2352292	Benzidine	23.9	15.6	P/P	0.0 - 240
2352292	Benzo(a)anthracene	93.4	92.7	P/P	50 - 130
2352292	Benzo(a)pyrene	96.0	88.0	P/P	50 - 130
2352292	Benzo(b)fluoranthene	99.2	90.7	P/P	50 - 130
2352292	Benzo(g,h,i)perylene	81.8	79.0	P/P	50 - 130
2352292	Benzo(k)fluoranthene	94.5	87.7	P/P	50 - 130
2352292	Benzyl alcohol	95.1	83.2	P/P	50 - 130
2352292	Bis(2-chloroethoxy)methane	91.7	88.9	P/P	50 - 130
2352292	Bis(2-chloroethyl)ether	89.7	82.8	P/P	50 - 160
2352292	Bis(2-chloroisopropyl)ether	89.3	83.8	P/P	50 - 130
2352292	Bis(2-ethylhexyl)phthalate	115	114	P/P	50 - 160
2352292	Butyl benzyl phthalate	104	105	P/P	50 - 160
2352292	Carbazole	126	128	P/P	50 - 130
2352292	Chrysene	91.6	92.1	P/P	50 - 130
2352292	Di-n-butyl phthalate	97.5	98.0	P/P	50 - 160
2352292	Di-n-octyl phthalate	111	106	P/P	50 - 130
2352292	Dibenzo(a,h)anthracene	87.0	84.4	P/P	50 - 130
2352292	Dibenzofuran	97.6	92.8	P/P	50 - 130
2352292	Diethyl phthalate	107	104	P/P	50 - 130
2352292	Dimethyl phthalate	102	98.7	P/P	50 - 130
2352292	Dimethylaminoazobenzene	93.0	103	P/P	50 - 130
2352292	Dinoseb	128	128	P/P	50 - 150
2352292	Ethyl methanesulfonate	97.0	89.9	P/P	50 - 130
2352292	Fluoranthene	90.6	91.1	P/P	50 - 130
2352292	Fluorene	99.9	96.0	P/P	50 - 130
2352292	Hexachlorobenzene	87.9	89.1	P/P	50 - 130
2352292	Hexachlorobutadiene	85.1	80.6	P/P	20 - 130
2352292	Hexachlorocyclopentadiene	49.8	43.7	P/P	20 - 130
2352292	Hexachloroethane	80.7	72.8	P/P	40 - 130
2352292	Hexachloropropene	89.5	76.3	P/P	50 - 130
2352292	Indeno(1,2,3-cd)pyrene	85.3	81.7	P/P	50 - 130
2352292	Isophorone	96.1	92.3	P/P	50 - 130
2352292	Isosafrole	101	108	P/P	50 - 130
2352292	m,p-Cresols	81.4	78.5	P/P	50 - 130
2352292	N-Nitrosodi-n-butylamine	109	101	P/P	50 - 130

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P418679

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352292	N-Nitrosodi-n-propylamine	90.8	85.1	P/P	50 - 130
2352292	N-Nitrosodiethylamine	103	95.4	P/P	50 - 130
2352292	N-Nitrosodimethylamine	104	97.7	P/P	30 - 130
2352292	N-Nitrosomethylethylamine	103	95.0	P/P	50 - 130
2352292	N-Nitrosomorpholine	126	117	P/P	50 - 150
2352292	N-Nitrosopiperidine	109	101	P/P	50 - 130
2352292	N-Nitrosopyrrolidine	94.8	88.9	P/P	50 - 130
2352292	Naphthalene	91.0	85.6	P/P	50 - 130
2352292	Nitrobenzene	95.1	91.7	P/P	50 - 130
2352292	o-Cresol	85.7	82.2	P/P	50 - 130
2352292	o-Toluidine	110	102	P/P	50 - 130
2352292	Pentachlorobenzene	92.3	102	P/P	50 - 130
2352292	Pentachloroethane	93.9	85.1	P/P	50 - 130
2352292	Pentachloronitrobenzene	124	127	P/P	50 - 130
2352292	Pentachlorophenol	96.7	98.4	P/P	50 - 130
2352292	Phenacetin	122	125	P/P	50 - 130
2352292	Phenanthrene	92.5	94.1	P/P	50 - 130
2352292	Phenol	54.8	51.6	P/P	15 - 110
2352292	Pyrene	94.2	93.8	P/P	50 - 130
2352292	Pyridine	77.5	73.5	P/P	20 - 130
2352292	Safrole	108	100	P/P	50 - 130

Reference Method: EPA 8321B

Batch ID: P418569

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2351625	Cylindrospermopsin	91.4	88.2	P/P	40 - 150
2351625	Desmethyl microcystin LR	106	112	P/P	40 - 150
2351625	Microcystin HILR	129	136	P/P	40 - 150
2351625	Microcystin HtyR	108	115	P/P	40 - 150
2351625	Microcystin LA	98.8	94.2	P/P	40 - 150
2351625	Microcystin LF	77.2	82.0	P/P	40 - 150
2351625	Microcystin LR	110	115	P/P	40 - 150
2351625	Microcystin LW	94.1	98.5	P/P	40 - 150
2351625	Microcystin LY	94.5	84.8	P/P	40 - 150
2351625	Microcystin RR	102	107	P/P	40 - 150
2351625	Microcystin WR	107	115	P/P	40 - 150
2351625	Microcystin YR	110	116	P/P	40 - 150
2351625	Nodularin-R	115	118	P/P	40 - 150

Reference Method: EPA 8321B

Batch ID: P418570

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2351625	Anatoxin-a	82.8	81.5	P/P	40 - 150

Reference Method: FL-PRO 2018

Batch ID: P418756

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2352293	TRPH	64.3	83.6	F/P	65 - 123

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P418886

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352296	Ammonia-N	0.195	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P418823

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352296	Kjeldahl Nitrogen	1.70	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P419050

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352331	NO2NO3-N	1.19	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P419131

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352281	Total-P	1.83	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P418697

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352277	O-Phosphate-P	0.661	Spike	P	0 - 20

Reference Method: EPA 8270E

Batch ID: P418679

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352292	1-Methylnaphthalene	4.47	Spike	P	0 - 40
2352292	1-Naphthylamine	19.4	Spike	P	0 - 40
2352292	1,2,4-Trichlorobenzene	7.33	Spike	P	0 - 40
2352292	1,2,4,5-Tetrachlorobenzene	7.87	Spike	P	0 - 40
2352292	1,3-Dinitrobenzene	3.54	Spike	P	0 - 40
2352292	1,3,5-Trinitrobenzene	4.20	Spike	P	0 - 40
2352292	2-Acetylaminofluorene	4.67	Spike	P	0 - 40
2352292	2-Chloronaphthalene	4.91	Spike	P	0 - 40
2352292	2-Chlorophenol	6.90	Spike	P	0 - 40
2352292	2-Methyl-4,6-dinitrophenol	0.694	Spike	P	0 - 40
2352292	2-Methylnaphthalene	5.34	Spike	P	0 - 40
2352292	2-Naphthylamine	5.71	Spike	P	0 - 40
2352292	2-Nitroaniline	4.68	Spike	P	0 - 40
2352292	2-Nitrophenol	3.47	Spike	P	0 - 40
2352292	2-Picoline	7.08	Spike	P	0 - 40
2352292	2,3,4,6-Tetrachlorophenol	3.37	Spike	P	0 - 40

Quality Assurance Report Precision

Reference Method: EPA 8270E

Batch ID: P418679

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352292	2,4-Dichlorophenol	4.52	Spike	P	0 - 40
2352292	2,4-Dimethylphenol	3.90	Spike	P	0 - 40
2352292	2,4-Dinitrophenol	2.45	Spike	P	0 - 40
2352292	2,4-Dinitrotoluene	3.67	Spike	P	0 - 40
2352292	2,4,5-Trichlorophenol	1.23	Spike	P	0 - 40
2352292	2,4,6-Trichlorophenol	2.57	Spike	P	0 - 40
2352292	2,6-Dichlorophenol	5.66	Spike	P	0 - 40
2352292	2,6-Dinitrotoluene	4.59	Spike	P	0 - 40
2352292	3-Methylcholanthrene	5.06	Spike	P	0 - 40
2352292	3,3'-Dichlorobenzidine	5.85	Spike	P	0 - 40
2352292	4-Aminobiphenyl	4.97	Spike	P	0 - 40
2352292	4-Bromophenyl phenyl ether	0.113	Spike	P	0 - 40
2352292	4-Chloro-3-methylphenol	2.98	Spike	P	0 - 40
2352292	4-Chlorophenyl phenyl ether	5.75	Spike	P	0 - 40
2352292	4-Nitrophenol	1.31	Spike	P	0 - 40
2352292	5-Nitro-o-toluidine	6.12	Spike	P	0 - 40
2352292	7,12-Dimethylbenz(a)anthracene	12.9	Spike	P	0 - 40
2352292	Acenaphthene	4.60	Spike	P	0 - 40
2352292	Acenaphthylene	5.67	Spike	P	0 - 40
2352292	Acetophenone	9.21	Spike	P	0 - 40
2352292	Aniline	5.26	Spike	P	0 - 40
2352292	Anthracene	0.219	Spike	P	0 - 40
2352292	Azobenzene/1,2-Diphenylhydrazine	1.51	Spike	P	0 - 40
2352292	Benzo(a)anthracene	0.752	Spike	P	0 - 40
2352292	Benzo(a)pyrene	8.70	Spike	P	0 - 40
2352292	Benzo(b)fluoranthene	8.95	Spike	P	0 - 40
2352292	Benzo(g,h,i)perylene	3.48	Spike	P	0 - 40
2352292	Benzo(k)fluoranthene	7.46	Spike	P	0 - 40
2352292	Benzyl alcohol	13.3	Spike	P	0 - 40
2352292	Bis(2-chloroethoxy)methane	3.10	Spike	P	0 - 40
2352292	Bis(2-chloroethyl)ether	8.00	Spike	P	0 - 40
2352292	Bis(2-chloroisopropyl)ether	6.35	Spike	P	0 - 40
2352292	Bis(2-ethylhexyl)phthalate	1.05	Spike	P	0 - 40
2352292	Butyl benzyl phthalate	0.478	Spike	P	0 - 40
2352292	Carbazole	1.81	Spike	P	0 - 40
2352292	Chrysene	0.544	Spike	P	0 - 40
2352292	Di-n-butyl phthalate	0.512	Spike	P	0 - 40
2352292	Di-n-octyl phthalate	4.98	Spike	P	0 - 40
2352292	Dibenzo(a,h)anthracene	3.03	Spike	P	0 - 40
2352292	Dibenzofuran	5.04	Spike	P	0 - 40
2352292	Diethyl phthalate	2.75	Spike	P	0 - 40
2352292	Dimethyl phthalate	2.90	Spike	P	0 - 40
2352292	Dimethylaminoazobenzene	10.0	Spike	P	0 - 40
2352292	Dinoseb	0.391	Spike	P	0 - 40
2352292	Ethyl methanesulfonate	7.60	Spike	P	0 - 40
2352292	Fluoranthene	0.550	Spike	P	0 - 40
2352292	Fluorene	3.98	Spike	P	0 - 40
2352292	Hexachlorobenzene	1.36	Spike	P	0 - 40
2352292	Hexachlorobutadiene	5.43	Spike	P	0 - 40
2352292	Hexachlorocyclopentadiene	13.0	Spike	P	0 - 40
2352292	Hexachloroethane	10.3	Spike	P	0 - 40

Quality Assurance Report Precision

Reference Method: EPA 8270E

Batch ID: P418679

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352292	Hexachloropropene	15.9	Spike	P	0 - 40
2352292	Indeno(1,2,3-cd)pyrene	4.31	Spike	P	0 - 40
2352292	Isophorone	4.03	Spike	P	0 - 40
2352292	Isosafrole	6.82	Spike	P	0 - 40
2352292	m,p-Cresols	3.63	Spike	P	0 - 40
2352292	N-Nitrosodi-n-butylamine	7.46	Spike	P	0 - 40
2352292	N-Nitrosodi-n-propylamine	6.48	Spike	P	0 - 40
2352292	N-Nitrosodiethylamine	7.37	Spike	P	0 - 40
2352292	N-Nitrosodimethylamine	6.05	Spike	P	0 - 40
2352292	N-Nitrosomethylethylamine	8.27	Spike	P	0 - 40
2352292	N-Nitrosomorpholine	7.50	Spike	P	0 - 40
2352292	N-Nitrosopiperidine	7.51	Spike	P	0 - 40
2352292	N-Nitrosopyrrolidine	6.42	Spike	P	0 - 40
2352292	Naphthalene	6.12	Spike	P	0 - 40
2352292	Nitrobenzene	3.64	Spike	P	0 - 40
2352292	o-Cresol	4.17	Spike	P	0 - 40
2352292	o-Toluidine	7.74	Spike	P	0 - 40
2352292	Pentachlorobenzene	9.69	Spike	P	0 - 40
2352292	Pentachloroethane	9.83	Spike	P	0 - 40
2352292	Pentachloronitrobenzene	2.15	Spike	P	0 - 40
2352292	Pentachlorophenol	1.74	Spike	P	0 - 40
2352292	Phenacetin	2.18	Spike	P	0 - 40
2352292	Phenanthrene	1.71	Spike	P	0 - 40
2352292	Phenol	6.02	Spike	P	0 - 40
2352292	Pyrene	0.426	Spike	P	0 - 40
2352292	Pyridine	5.30	Spike	P	0 - 40
2352292	Safrole	7.11	Spike	P	0 - 40

Reference Method: EPA 8321B

Batch ID: P418569

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2351625	Cylindrospermopsin	3.49	Spike	P	0 - 30
2351625	Desmethyl microcystin LR	5.15	Spike	P	0 - 30
2351625	Microcystin H1R	5.47	Spike	P	0 - 30
2351625	Microcystin HtyR	5.96	Spike	P	0 - 30
2351625	Microcystin LA	4.76	Spike	P	0 - 30
2351625	Microcystin LF	6.09	Spike	P	0 - 30
2351625	Microcystin LR	3.94	Spike	P	0 - 30
2351625	Microcystin LW	4.58	Spike	P	0 - 30
2351625	Microcystin LY	10.8	Spike	P	0 - 30
2351625	Microcystin RR	4.42	Spike	P	0 - 30
2351625	Microcystin WR	7.17	Spike	P	0 - 30
2351625	Microcystin YR	4.96	Spike	P	0 - 30
2351625	Nodularin-R	3.05	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P418570

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2351625	Anatoxin-a	1.55	Spike	P	0 - 30

Reference Method: FL-PRO 2018
Batch ID: P418756

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2352293	TRPH	22.1	Spike	F	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2352292

Field Sample ID: HAB-SD-082422-1505

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	2-Fluorobiphenyl	107	P	30 - 150
EPA 8270E	2-Fluorophenol	65.2	P	20 - 150
EPA 8270E	2,4,6-Tribromophenol	95.1	P	30 - 150
EPA 8270E	Nitrobenzene-d5	120	P	30 - 150
EPA 8270E	Phenol-d5	58.6	P	20 - 150
EPA 8270E	Terphenyl-d14	127	P	30 - 150

Lab Sample ID: 2352293

Field Sample ID: HAB-SD-082422-1505

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
FL-PRO 2018	Nonatriacontane(C39)	70.3	P	40 - 120
FL-PRO 2018	o-Terphenyl	93.5	P	66 - 137

Lab Sample ID: 2352294

Field Sample ID: HAB-SD-082422-1505

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	104	P	30 - 160
EPA 8321B	Neosaxitoxin-15N7	75.5	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A114331
Included Lab Sample IDs: 2352294

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Anatoxin-a	88.9	86.3	P/P	50 - 160

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A114347
Included Lab Sample IDs: 2352297

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.1	97.9	P/P	90 - 110

Reference Method: EPA 8321B
Run ID: A114353
Included Lab Sample IDs: 2352294

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Cylindrospermopsin	87.8	79.2	P/P	50 - 160
Desmethyl microcystin LR	76.5	82.4	P/P	50 - 160
Microcystin H1IR	84.2	84.1	P/P	50 - 160
Microcystin HtyR	83.3	86.6	P/P	50 - 160
Microcystin LA	79.6	87.7	P/P	50 - 160
Microcystin LF	77.3	73.4	P/P	50 - 160
Microcystin LR	85.2	83.5	P/P	50 - 160
Microcystin LW	73.4	74.3	P/P	50 - 160
Microcystin LY	79.7	70.0	P/P	50 - 160
Microcystin RR	79.0	80.2	P/P	50 - 160
Microcystin WR	78.4	87.1	P/P	50 - 160
Microcystin YR	77.5	79.2	P/P	50 - 160
Nodularin-R	94.0	92.9	P/P	50 - 160

Reference Method: FL-PRO 2018
Run ID: A114416
Included Lab Sample IDs: 2352293

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
TRPH	96.5	92.5	P/P	75 - 125

Reference Method: EPA 8270E
Run ID: A114431
Included Lab Sample IDs: 2352292

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
1-Methylnaphthalene	83.6		P	70 - 130
1,2,4-Trichlorobenzene	99.5		P	70 - 130
1,2,4,5-Tetrachlorobenzene	91.9		P	70 - 130
1,3-Dinitrobenzene	75.8		P	70 - 130
1,3,5-Trinitrobenzene	74.9		P	70 - 130
2-Chloronaphthalene	104		P	70 - 130
2-Chlorophenol	92.0		P	70 - 130
2-Methylnaphthalene	114		P	70 - 130
2-Nitroaniline	90.1		P	70 - 130
2-Nitrophenol	104		P	70 - 130
2,3,4,6-Tetrachlorophenol	87.7		P	70 - 130

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
 Run ID: A114431
 Included Lab Sample IDs: 2352292

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-Dichlorophenol	88.0		P	70 - 130
2,4-Dimethylphenol	110		P	70 - 130
2,4-Dinitrotoluene	99.2		P	70 - 130
2,4,5-Trichlorophenol	124		P	70 - 130
2,4,6-Trichlorophenol	113		P	70 - 130
2,6-Dichlorophenol	99.0		P	70 - 130
2,6-Dinitrotoluene	105		P	70 - 130
3-Methylcholanthrene	99.4		P	70 - 130
4-Bromophenyl phenyl ether	102		P	70 - 130
4-Chloro-3-methylphenol	97.8		P	70 - 130
4-Chlorophenyl phenyl ether	101		P	70 - 130
5-Nitro-o-toluidine	90.1		P	70 - 130
7,12-Dimethylbenz(a)anthracene	81.9		P	70 - 130
Acenaphthene	101		P	70 - 130
Acenaphthylene	103		P	70 - 130
Acetophenone	94.9		P	70 - 130
Anthracene	104		P	70 - 130
Azobenzene/1,2-Diphenylhydrazine	111		P	70 - 130
Benzo(a)anthracene	108		P	70 - 130
Benzo(a)pyrene	106		P	70 - 130
Benzo(b)fluoranthene	96.3		P	70 - 130
Benzo(g,h,i)perylene	105		P	70 - 130
Benzo(k)fluoranthene	99.6		P	70 - 130
Benzyl alcohol	81.7		P	70 - 130
Bis(2-chloroethoxy)methane	97.3		P	70 - 130
Bis(2-chloroethyl)ether	92.7		P	70 - 130
Bis(2-chloroisopropyl)ether	101		P	70 - 130
Carbazole	98.3		P	70 - 130
Chrysene	102		P	70 - 130
Di-n-octyl phthalate	121		P	70 - 130
Dibenzo(a,h)anthracene	117		P	70 - 130
Dibenzofuran	101		P	70 - 130
Dimethyl phthalate	100		P	70 - 130
Dimethylaminoazobenzene	88.6		P	70 - 130
Fluoranthene	103		P	70 - 130
Fluorene	102		P	70 - 130
Hexachlorobenzene	96.4		P	70 - 130
Hexachlorobutadiene	96.6		P	70 - 130
Hexachlorocyclopentadiene	101		P	70 - 130
Hexachloroethane	96.0		P	70 - 130
Hexachloropropene	93.4		P	70 - 130
Indeno(1,2,3-cd)pyrene	114		P	70 - 130
Isophorone	111		P	70 - 130
Isosafrole	95.4		P	70 - 130
m,p-Cresols	86.6		P	70 - 130
N-Nitrosodi-n-butylamine	87.4		P	70 - 130
N-Nitrosodi-n-propylamine	94.9		P	70 - 130
N-Nitrosomorpholine	88.1		P	70 - 130
N-Nitrosopiperidine	93.4		P	70 - 130
N-Nitrosopyrrolidine	76.4		P	70 - 130
Naphthalene	101		P	70 - 130

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A114431
Included Lab Sample IDs: 2352292

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Nitrobenzene	78.4		P	70 - 130
o-Cresol	95.8		P	70 - 130
o-Toluidine	105		P	70 - 130
Pentachlorobenzene	91.7		P	70 - 130
Pentachloroethane	88.9		P	70 - 130
Pentachloronitrobenzene	91.2		P	70 - 130
Phenanthrene	96.7		P	70 - 130
Phenol	82.4		P	70 - 130
Pyrene	104		P	70 - 130
Safrole	89.9		P	70 - 130

Reference Method: EPA 8270E
Run ID: A114433
Included Lab Sample IDs: 2352292

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
1-Naphthylamine	124		P	60 - 130
2-Acetylaminofluorene	98.8		P	70 - 150
2-Methyl-4,6-dinitrophenol	107		P	70 - 130
2-Naphthylamine	88.7		P	60 - 130
2-Picoline	111		P	70 - 130
2,4-Dinitrophenol	74.6		P	70 - 130
3,3'-Dichlorobenzidine	128		P	50 - 130
4-Aminobiphenyl	91.9		P	70 - 130
4-Nitrophenol	98.3		P	70 - 130
Aniline	106		P	70 - 130
Benzidine	77.1		P	50 - 130
Bis(2-ethylhexyl)phthalate	114		P	70 - 130
Butyl benzyl phthalate	113		P	70 - 130
Di-n-butyl phthalate	105		P	70 - 130
Diethyl phthalate	108		P	70 - 130
Dinoseb	91.8		P	70 - 130
Ethyl methanesulfonate	109		P	70 - 130
N-Nitrosodiethylamine	114		P	70 - 130
N-Nitrosodimethylamine	128		P	70 - 130
N-Nitrosomethylethylamine	115		P	70 - 130
Pentachlorophenol	93.4		P	70 - 130
Phenacetin	108		P	70 - 130
Pyridine	113		P	70 - 130

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A114437
Included Lab Sample IDs: 2352296

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.5	96.1	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A114482

Included Lab Sample IDs: 2352296

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.9	99.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A114510

Included Lab Sample IDs: 2352296

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	103	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A114650

Included Lab Sample IDs: 2352296

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 350.1 Rev. 2.0	Ammonia-N	96.8	95.2	95.4			0.195
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	100	96.9			1.70
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	99.5	102			1.19
EPA 365.1 Rev. 2.0	Total-P	104	104	106			1.83
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	101	100			0.661
EPA 8270E	1-Methylnaphthalene	92.7	89.6	93.7			4.47
	1-Naphthylamine	65.2	87.5	106			19.4
	1,2,4-Trichlorobenzene	84.1	78.9	84.9			7.33
	1,2,4,5-Tetrachlorobenzene	96.0	89.1	96.4			7.87
	1,3-Dinitrobenzene	103	101	97.2			3.54
	1,3,5-Trinitrobenzene	124	117	112			4.20
	2-Acetylaminofluorene	117	121	115			4.67
	2-Chloronaphthalene	91.6	87.4	91.8			4.91
	2-Chlorophenol	87.2	81.1	86.9			6.90
	2-Methyl-4,6-dinitrophenol	99.3	116	115			0.694
	2-Methylnaphthalene	92.3	87.5	92.3			5.34
	2-Naphthylamine	66.1	71.5	75.7			5.71
	2-Nitroaniline	104	98.0	103			4.68
	2-Nitrophenol	92.5	87.7	90.8			3.47
	2-Picoline	80.5	92.6	99.4			7.08
	2,3,4,6-Tetrachlorophenol	107	107	104			3.37
	2,4-Dichlorophenol	92.1	88.7	92.8			4.52
	2,4-Dimethylphenol	88.3	88.0	91.5			3.90
	2,4-Dinitrophenol	41.8	72.6	74.4			2.45
	2,4-Dinitrotoluene	99.1	96.2	99.8			3.67
	2,4,5-Trichlorophenol	87.8	88.9	90.0			1.23
	2,4,6-Trichlorophenol	90.3	88.3	90.6			2.57
	2,6-Dichlorophenol	109	103	109			5.66
	2,6-Dinitrotoluene	99.4	95.8	100			4.59
	3-Methylcholanthrene	82.1	97.2	92.4			5.06
	3,3'-Dichlorobenzidine	195	169	180			5.85
	4-Aminobiphenyl	98.7	98.1	103			4.97
	4-Bromophenyl phenyl ether	90.3	88.1	88.2			0.113
	4-Chloro-3-methylphenol	91.1	89.4	92.1			2.98
	4-Chlorophenyl phenyl ether	98.7	94.6	100			5.75
	4-Nitrophenol	58.8	53.0	53.7			1.31
	5-Nitro-o-toluidine	120	121	114			6.12
	7,12-Dimethylbenz(a)anthracene	106	97.4	85.6			12.9
	Acenaphthene	96.8	93.4	97.8			4.60
	Acenaphthylene	95.8	90.8	96.1			5.67
	Acetophenone	114	104	114			9.21
	Aniline	90.5	85.1	89.7			5.26
	Anthracene	94.3	91.3	91.5			0.219
	Azobenzene/1,2-Diphenylhydrazine	99.1	100	98.7			1.51
	Benzidine	65.0	15.6	23.9			
	Benzo(a)anthracene	97.8	92.7	93.4			0.752
	Benzo(a)pyrene	92.0	88.0	96.0			8.70
	Benzo(b)fluoranthene	97.3	90.7	99.2			8.95
	Benzo(g,h,i)perylene	74.9	79.0	81.8			3.48
	Benzo(k)fluoranthene	91.7	87.7	94.5			7.46
	Benzyl alcohol	91.6	83.2	95.1			13.3
	Bis(2-chloroethoxy)methane	91.9	88.9	91.7			3.10
	Bis(2-chloroethyl)ether	91.6	82.8	89.7			8.00
	Bis(2-chloroisopropyl)ether	90.1	83.8	89.3			6.35

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 8270E	Bis(2-ethylhexyl)phthalate	116	114	115			1.05
	Butyl benzyl phthalate	109	105	104			0.478
	Carbazole	118	128	126			1.81
	Chrysene	98.2	92.1	91.6			0.544
	Di-n-butyl phthalate	102	98.0	97.5			0.512
	Di-n-octyl phthalate	107	106	111			4.98
	Dibenzo(a,h)anthracene	80.9	84.4	87.0			3.03
	Dibenzofuran	96.8	92.8	97.6			5.04
	Diethyl phthalate	110	104	107			2.75
	Dimethyl phthalate	104	98.7	102			2.90
	Dimethylaminoazobenzene	113	103	93.0			10.0
	Dinoseb	126	128	128			0.391
	Ethyl methanesulfonate	96.2	89.9	97.0			7.60
	Fluoranthene	94.5	91.1	90.6			0.550
	Fluorene	100	96.0	99.9			3.98
	Hexachlorobenzene	90.9	89.1	87.9			1.36
	Hexachlorobutadiene	84.5	80.6	85.1			5.43
	Hexachlorocyclopentadiene	51.9	43.7	49.8			13.0
	Hexachloroethane	84.0	72.8	80.7			10.3
	Hexachloropropene	97.0	76.3	89.5			15.9
	Indeno(1,2,3-cd)pyrene	78.8	81.7	85.3			4.31
	Isophorone	96.2	92.3	96.1			4.03
	Isosafrole	108	108	101			6.82
	m,p-Cresols	82.5	78.5	81.4			3.63
	N-Nitrosodi-n-butylamine	107	101	109			7.46
	N-Nitrosodi-n-propylamine	89.1	85.1	90.8			6.48
	N-Nitrosodiethylamine	104	95.4	103			7.37
	N-Nitrosodimethylamine	92.4	97.7	104			6.05
	N-Nitrosomethylethylamine	104	95.0	103			8.27
	N-Nitrosomorpholine	127	117	126			7.50
	N-Nitrosopiperidine	109	101	109			7.51
	N-Nitrosopyrrolidine	96.1	88.9	94.8			6.42
	Naphthalene	90.7	85.6	91.0			6.12
	Nitrobenzene	95.9	91.7	95.1			3.64
	o-Cresol	86.7	82.2	85.7			4.17
	o-Toluidine	111	102	110			7.74
	Pentachlorobenzene	99.7	102	92.3			9.69
	Pentachloroethane	95.8	85.1	93.9			9.83
	Pentachloronitrobenzene	130	127	124			2.15
	Pentachlorophenol	91.5	98.4	96.7			1.74
	Phenacetin	130	125	122			2.18
	Phenanthrene	95.2	94.1	92.5			1.71
	Phenol	54.8	51.6	54.8			6.02
	Pyrene	98.3	93.8	94.2			0.426
	Pyridine	70.4	73.5	77.5			5.30
	Safrole	106	100	108			7.11
EPA 8321B	Anatoxin-a	75.5	82.8	81.5			1.55
	Cylindrospermopsin	79.8	91.4	88.2			3.49
	Desmethyl microcystin LR	73.2	106	112			5.15
	Microcystin HiIR	77.2	129	136			5.47
	Microcystin HtyR	78.1	108	115			5.96
	Microcystin LA	80.6	98.8	94.2			4.76
	Microcystin LF	66.4	77.2	82.0			6.09
	Microcystin LR	76.4	110	115			3.94

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 8321B	Microcystin LW	64.4	94.1	98.5		4.58
	Microcystin LY	66.1	94.5	84.8		10.8
	Microcystin RR	77.1	102	107		4.42
	Microcystin WR	73.7	107	115		7.17
	Microcystin YR	80.0	110	116		4.96
	Nodularin-R	80.8	115	118		3.05
FL-PRO 2018	TRPH	90.9	64.3	83.6		22.1

Reference Method Descriptions

Method	Description	Associated Samples
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2352296
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2352296
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2352296
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2352296
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2352297
EPA 8270E	EPA Method 8270, Semi-volatile organic pollutants including PAHs, excluding PCBs and Toxaphene, in water matrices by GC/MS.	2352292
EPA 8321B	Microcystins in water matrices by HPLC/MS/MS	2352294
EPA 8321B	Saxitoxin in water matrices by HPLC/MS/MS	2352294
FL-PRO 2018	Total recoverable petroleum hydrocarbons in water samples by GC-FID.	2352293

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	08/25/2022			08/30/2022 14:42	Judith K. Clark	2352296
EPA 351.2 Rev. 2.0	08/25/2022	08/30/2022 14:31	Samantha L Bates	09/01/2022 10:35	Alexandra J Mattheus	2352296
EPA 353.2 Rev. 2.0	08/25/2022			09/02/2022 09:09	Judith K. Clark	2352296
EPA 365.1 Rev. 2.0	08/25/2022	09/07/2022 16:10	Simonne.V. Calhoun	09/09/2022 13:04	James L Waggeberby	2352296
EPA 365.1 Rev. 2.0 dissolved	08/25/2022			08/25/2022 14:41	Elise M. Gillis	2352297
EPA 8270E	08/25/2022	08/29/2022 09:00	Hoor Shaik	08/30/2022 11:51	Mohammad Ghaffari	2352292
	08/25/2022	08/29/2022 09:00	Hoor Shaik	08/30/2022 12:37	Mohammad Ghaffari	2352292
EPA 8321B	08/25/2022	08/25/2022 13:00	Manjita Shrestha	08/25/2022 20:10	Manjita Shrestha	2352294
	08/25/2022	08/25/2022 13:00	Manjita Shrestha	08/25/2022 23:34	Manjita Shrestha	2352294
FL-PRO 2018	08/25/2022	08/26/2022 08:00	Lipi Saha	08/29/2022 13:42	Lipi Saha	2352293