

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: TH60MG

RQ- 2022-04-25-47

Project: BLOOM-RGSP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4/14/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	BARRINGTON	5/3/22	0700	23.9	762.4	8.5	8.50	100.3	-	1.10	P / F	L / F
ICV	↓	↓	0703	23.8	762.4	8.5	8.51	100.7	-	-	P / F	L / F
CCV	William Muller	5/03/22	1141	24.9	763.2	8.5	8.01	102.3	-	-	P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	BARRINGTON	5/3/22	0706	220400	3/24	2500	2500	P / F	L / F
ICV	↓	↓	0709	2112628	12/23	100	100.8	P / F	L / F
CCV	William Muller	5/03/22	1150	22.0309	03/2023	50,000	49,310	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	BARRINGTON	5/3/22	0712	1112468	11/23	7.00	24.0	7.0	-	P / F	L / F
Calibr.	↓	↓	0715	2203624	3/24	4.01	24.1	4.00	-	P / F	L / F
Calibr.	↓	↓	0718	4202048	6/23	10.02	24.3	10.01	-	P / F	L / F
ICV	↓	↓	0721	2203624	3/24	9.01	24.2	4.07	-	P / F	L / F
CCV	William Muller	5/03/22	1156	4202048	6/23	10.02	24.5	10.03	-	P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 -, slope from 4 to 7 - (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 4/24/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: DEEKE BLUER; Date of last verification: 4/24/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: