

Chemical Analysis Report

SO-DIST-2015-07-29-03

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Algal Bloom Response - As needed**
Request ID: **RQ-2015-06-01-67**
Customer: **SO-DIST**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Kirby Wolfe

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Timothy Fitzpatrick, Program Administrator

Date Certified: 14-AUG-2015 10:01

A handwritten signature in black ink, appearing to read "Timothy Fitzpatrick", with a stylized flourish at the end.

Abbreviations and data remark codes

CERT # - NELAP (National Environmental Laboratory Accreditation Program) Certification Number of the laboratory that performed the analysis.

LCS – Laboratory Control Sample; in the QC Failures column, this notation indicates a batch recovery failure.

MS – Matrix Spike; in the QC Failures column, this notation indicates a batch recovery failure.

RPD – Relative Percent Difference; in the QC Failures column, this notation indicates a batch failure for precision.

CCV – Continuing Calibration Verification; in the QC Failures column, this notation indicates a failure of the calibration verification check sample.

SUR – Surrogate; in the QC Failures column, this notation indicates a recovery failure for the associated surrogate.

RSD – Relative Standard Deviation expressed as a percentage.

SMP – Sample.

** - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.

A - Value reported is the mean of two or more determinations

B - Results based on colony counts outside the acceptable range.

I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.

J - Estimated value and/or the analysis did not meet established quality control criteria

K - Actual value is known to be less than value given

L - Actual value is known to be greater than value given

N - Presumptive evidence of presence of material.

O - Sampled, but analysis lost or not performed.

Q - Sample held beyond normal holding time.

T - Value reported is less than the criterion of detection.

U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.

V - Analyte was detected in both sample and method blank.

X - Too few individuals to calculate SCI value.

Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate

Z - Colonies were too numerous to count (TNTC).

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example:
1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Precision is reported as relative percent difference unless otherwise noted.

Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published test performance acceptance criteria.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

All sample collection performed by Bureau of Laboratories staff followed the field Standard Operating Procedures adopted by reference in Rule 62-160 FAC unless otherwise noted.

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Results for NELAP accredited tests contained in this report meet the requirements specified by the National Environmental Laboratory Accreditation Conference (NELAC). All samples received were in acceptable condition and met NELAC requirements unless otherwise noted. Results generated in this report pertain to the samples collected and submitted for analysis.

All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Sample Location: Marco Island

Collection Date/Time: 07/28/2015 11:35

Field ID: 200 Westbrook Ct.

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Failures
1723812	EPA 350.1 Rev. 2.0	Ammonia-N	0.037		mg N/L	P288979	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.63		mg N/L	P289048	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.024		mg N/L	P288926	
	EPA 365.1 Rev. 2.0	Total-P	0.018		mg P/L	P289090	
1723813	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.009	I	mg P/L	P288869	

Ref. Method and Comment:

EPA 350.1 Rev. 2.0: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P288979

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P289048

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P288926

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P289090

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P288869

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P288979

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.7		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P289048

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P288926

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	94.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P289090

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P288869

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	105		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P289048

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1723770	Kjeldahl Nitrogen	109	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P288926

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1723765	NO2NO3-N	95.0	97.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P289090

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1723771	Total-P	96.2	99.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P288869

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1723773	O-Phosphate-P	91.7	92.2	P/P	90 - 110

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P288979

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1723769	Ammonia-N	3.09	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P289048

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1723770	Kjeldahl Nitrogen	6.27	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P288926

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1723765	NO2NO3-N	2.08	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P289090

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1723771	Total-P	3.37	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P288869

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1723773	O-Phosphate-P	0.544	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A64034
Included Lab Sample IDs: 1723813

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	96.8	95.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A64055
Included Lab Sample IDs: 1723812

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A64078
Included Lab Sample IDs: 1723812

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A64142
Included Lab Sample IDs: 1723812

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	100	96.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A64170
Included Lab Sample IDs: 1723812

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	103	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 350.1 Rev. 2.0	Ammonia-N	98.7				3.09
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	109 100			6.27
EPA 353.2 Rev. 2.0	NO2NO3-N	94.5	95.0 97.0			2.08
EPA 365.1 Rev. 2.0	Total-P	104	96.2 99.6			3.37
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	105	91.7 92.2			0.544

Reference Method Descriptions

Method / NELAC Cert. #	Description	Associated Samples
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1723812
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1723812
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1723812
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1723812
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1723813

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	07/29/2015			07/30/2015	Diana M. Turner	1723812
EPA 351.2 Rev. 2.0	07/29/2015	08/03/2015	Christopher Armour	08/04/2015	Bryan A. Werth	1723812
EPA 353.2 Rev. 2.0	07/29/2015			07/30/2015	Peter D. Kaus	1723812
EPA 365.1 Rev. 2.0	07/29/2015	08/04/2015	Christopher Armour	08/05/2015	Lipi Saha	1723812
EPA 365.1 Rev. 2.0 dissolved	07/29/2015			07/29/2015 13:48	Bryan A. Werth	1723813