

CALIBRATION AND VERIFICATION LOG (DEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Stitch

RQ: 2023-01-02-48

Project: Algae Resample

George's Lake

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 11/14/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K. Halbert	01/18/23	747	20.1	763.2	9.1	—	100.4	1.069	1.069	(P) F	(L) F
ICV	K. Halbert	01/18/23	754	20.2	763.2	9.1	9.12	100.6	—	1.069	(P) F	(L) F
CCV	Brian Davis	↓	1507	21.0	761.2	8.9	8.93	100.1	✓	1.069	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	K. Halbert	01/18/23	756	221028A	11/20/23	1000	1000	(P) F	(L) F
ICV	K. Halbert	01/18/23	759	220719B	07/20/23	100	99.3	(P) F	(L) F
CCV	Brian Davis	↓	1508	221028A	11/23	1000	995	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K. Halbert	01/18/23	802	M118-15	5/2/24	7.02	21.1	7.02	-36.4	(P) F	(L) F
Calibr.	K. Halbert	01/18/23	810	M188-01	7/11/24	4.00	20.9	4.00	137.6	(P) F	(L) F
Calibr.	K. Halbert	01/18/23	812	M163-01	4/19/24	10.04	21.3	10.04	-209.2	(P) F	(L) F
ICV	K. Halbert	01/18/23	816	M188-01	7/11/24	4.00	21.0	4.03	137.8	(P) F	(L) F
CCV	Brian Davis	11/18/23	1510	M118-15	5/2/24	7.01	21.6	7.08	-38.8	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.8, slope from 4 to 7 174.0 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 11/14/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K. Halbert	01/18/23	820	0.000	0.002	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-01-02-48				Collected By:		Brian Davis, Katherine Halbert					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: George's Lake @ Center HAB				Comments:							
Field ID: George's Lake @ Center HAB				Coordinates:		29.79628, -81.85102					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID:NEHAB0098				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
01/18/2023	10:50	92.8	9.41	14.7	6.2	0.3	4.3	0.5	52.9	0.02	
	10:57	89.2	9.18	14.2	6.2	2			53	0.02	
	10:55	85.9	8.86	14	6.21	3.8			53.1	0.02	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☒ Scum		☒ Glob			
Algae Distribution		☒ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B	
		Acid Lot:		SA2145160		Exp:		6/2/23			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
		Filter Lot #		175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Brian Davis		01/18/2023 16:30		1							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-01-02-48				Collected By:		Brian Davis, Katherine Halbert					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Sampson River @ SW CR 225				Comments:							
Field ID: Sampson River @ SW CR 225				Coordinates:		29.9191, -82.21071					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID:NEHAB0088				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
01/18/2023	13:15	20.3	2.22	10.2	6.41	0.3	1	1	124.1	0.06	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☐ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☐ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☐ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☐ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B	
		Acid Lot:		SA2145160		Exp:		6/2/2023			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
				Filter Lot # 175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Brian Davis		01/18/2023 16:30		1							